



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

Public Sector Undertaking of Government of Jammu & Kashmir

Jammu Office: Plot No. 9, Transport Nagar, Jammu – 180004 (J&K) India

Srinagar Office: 121-Green Avenue, Hyderpora, Srinagar - 190014 (J&K) India

CORRIGENDUM

Sub.: Corrigendum in respect of RFP for Operationalization of 102 Ambulance Service across the State – reg.

In light of the representations received from the prospective bidders and discussions held during pre-bid meeting held on 09th Dec., 2015 in respect of “**RFP for Operationalization of 102 Ambulance Service across the State**”, following modifications have been made in the RFP:

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
1	Notice Inviting e-Tender		
1.1	Clause (5)	Bidders shall have to submit Rs.10,000.00 (Rupees Six Thousand only)	Bidders shall have to submit Rs.10,000.00 (Rupees Ten Thousand only)
2	Scope of Work		
2.1	Clause (3)(II)(d)	Comprehensive Maintenance Contract (CMC) for a period of five years from the end of guarantee period	Comprehensive Maintenance Contract (CMC) for a period of five years from the end of guarantee period including CMC of all retrofitting and of all the Machinery & Equipments to be fitted in the vehicle as a result of retrofitting of ambulance as well as healthcare equipments to be fitted in the ambulances
3	Terms & Conditions of Tender and Contract		
3.1	Clause (2)	Only Vehicle/ Ambulance manufacturing units (four or more than four wheels)/ Fabricators/ Consortium [up to three partners in which at least one has to be vehicle	Only Vehicle/ Ambulance manufacturing units (four or more than four wheels)/ Fabricators/ Ambulance Service Provider/ Consortium [up to three

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
		<p>manufacturer, one fabricator & one having experience in computer telephony integration with the ability to log calls with GPS (Global Positioning System) incorporated in GIS (Geographical Information System) with GSM/ GPRS (Global System for Mobile Communication/ General Packet Radio Service) integrated Ambulance monitoring and tracking system] who are holding valid Manufacturing/ fabricating license with average annual turnover of not less than Rs.10.00 Crores, during the preceding three year are eligible to participate in the Tender</p>	<p>partners in which at least one has to be vehicle manufacturer, one fabricator & one having experience in computer telephony integration with the ability to log calls with GPS (Global Positioning System) incorporated in GIS (Geographical Information System) with GSM/ GPRS (Global System for Mobile Communication/ General Packet Radio Service) integrated Ambulance monitoring and tracking system] who are holding valid Manufacturing/ fabricating license with average annual turnover of not less than Rs.10.00 Crores, during the preceding three year are eligible to participate in the Tender</p>
3.2	Guarantee Clause - (11)(ii)	<p>In case of machinery and equipment guarantee shall have to be given that the Tenderer shall during the guarantee period replace parts if any and remove the manufacturing defect if found during the above period so as to make the machinery and equipment operative</p>	<p>Guarantee includes all the Machinery & Equipments to be fitted in the vehicle including as a result of retrofitting of ambulance as well as healthcare equipments to be fitted in the ambulances</p>
3.3	Agreement - Clause 16(i)	<p>All firms whose offers are accepted will have to deposit a security equal to five per cent (5%) of the total value of approximate quantity as</p>	<p>Performance Bank Guarantee @ 5% of the Total Contract Value for 1st Year (Capex + Opex) which shall remain valid up to</p>

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
		per tender catalogue in favour of CAO, JKMSCL at the time of agreement. The security amount shall in no case be less than the earnest money.	Defect Liability Period for due performance of the Contract
3.4	Terms of Payment - Clause 19(e)	Rest 10% payment will be made against the 5% bank guarantee amount of the total value of the supply order which will be released after guarantee period.	10% Payment shall be withheld in case the bidder is failed to furnish the Performance Bank Guarantee and the same shall be released after Guarantee Period. In case, bidder has furnished Bank Guarantee, as required, 100% payment shall be released.
4	Specifications of Equipments (General for both)		
4.1	List of Healthcare Equipments (for BLS/ ALS Ambulances and Critical Care Ambulances)	Oxygen - System including Oxygen cylinder portable and Oxygen Key	Oxygen - System including Oxygen Cylinder D Type, Oxygen cylinder portable and Oxygen Key
4.2	Standards and Safety	FDA/ CE	US FDA/ European CE
4.3	Warranty & Maintenance	3 Years	5 Years including CMC of all retrofitting and of all the Machinery & Equipments to be fitted in the vehicle as a result of retrofitting of ambulance as well as healthcare equipments to be fitted in the ambulances
5	Specifications of Equipments for BLS Ambulances		
5.1	Auto Loader – Collapsible stretcher	Automatic Loading, made of aluminium alloy	Automatic Loading, made of Aluminium Alloy/ Stainless Steel
5.2	AC/ DC/ Foot Operated - Suction Pump	Collection bottle 600 ml capacity and e-Sterilizable	Collection bottle 600 ml or more capacity and e-Sterilizable
6	Specifications of Equipments for ALS Ambulances		

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
6.1	(1) Monitor Defibrillator - Technical Characteristics (1.1)(3)	Should be able to deliver shock from 50-200 joules in biphasic mode via metal chest pads	Should be able to deliver shock from 2-200 joules in biphasic mode via metal chest paddles
6.2	(1) Monitor Defibrillator - Technical Characteristics (1.1)(5)	The whole system should have an inbuilt recorder; Telemetry Not Recommended	The whole system should have an inbuilt recorder
6.3	(1) Monitor Defibrillator - Energy Source - Battery operated (3.2)	Rechargeable battery backup of approximately 5 hours	Rechargeable battery backup of at least 2.5 hours/ 100 shocks
6.4	(1) Monitor Defibrillator - Energy Source - Battery operated (3.5)	Should not be more than 160 W	Should be more than 160 W
6.5	(1) Monitor Defibrillator - (4) Accessories, Spare Parts, Consumables - (4.2) Consumables/ Reagents (Open, Closed System)	ECG Cable; Recording Paper Rolls; Disposable Pads	ECG Cable - 1 Inbuilt plus 2 Spare; Recording Paper Rolls for 100 Patients and Disposable Pads – 50
6.5	(2) Syringe Pump - Technical Characteristics - (1.1)	Must work on commonly available 20, 50 and 60 ml syringes	Must work on 10-50 ml syringes
6.6	(2) Syringe Pump - Technical Characteristics - (1.1)	Maximum pressure generated ≤ 20 psi	Deleted
6.7	(2) Syringe Pump - Technical Characteristics (1.1)	Comprehensive alarm package required including: occlusion alarm, near end of infusion pre-alarm and alarm, volume limit pre-alarm and alarm, low battery pre-alarm and alarm, AC power failure, drive disengaged, syringe loading	Comprehensive alarm package required including: occlusion alarm, near end of infusion alarm, volume limit alarm, low battery alarm, AC power failure, drive disengaged, syringe loading error, maintenance required

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
		error, maintenance required	
6.8	(2) Syringe Pump - Technical Characteristics - Settings (1.2)	Double loadable with one syringe of minimum 20ml	Single loadable with one syringe of minimum 20ml
6.9	(2) Syringe Pump - Physical Characteristics - Weight (lbs, kg) (2.2)	<500gm	Weight (lbs, kg) = up to 5 kg
6.10	(3) Transport Ventilator - Technical Characteristics - Settings (1.2)	Tidal volume up to 100 ml	Tidal volume up to 50-2000 ml
6.11	(3) Transport Ventilator - Technical Characteristics - (1.1)(8)	Visual and audible alarms Accessories and tubing should be supplied for adult, pediatric & neo-natal size requirements	Visual and audible alarms Accessories and tubing should be supplied for adult & paediatric size requirements
6.12	(3) Transport Ventilator - Settings - (1.2)(c)	Volume (inspiratory) up to 120 l/min	Flow (inspiratory) up to 120 l/min
6.13	(3) Transport Ventilator - Settings - (1.2)(f)	PEEP up to 20 cm H ₂ O	PEEP up to 20 cm H ₂ O or more
6.14	(3) Transport Ventilator - Settings - (1.2)(h)	FiO ₂ between 21 to 100 %	FiO ₂ between 40 to 100 %
7	Qualification of Para-Medics for Ambulances		
7.1	(3.) Driver	Matric pass holding a valid driving license and able to read and write. In addition, he shall be familiar with the local language	Matric pass holding a valid Heavy Goods Carrier driving license with minimum 2 years' experience and able to read and write. In addition, he shall be familiar with the local language
8	General Clarifications		
8.1	Only Vehicle/ Ambulance manufacturing units (four or more than four wheels)/ Fabricators/ Ambulance Service Provider/ Consortium [up to three partners in which at least one has to be vehicle manufacturer, one fabricator & one having experience in computer telephony integration with the ability to log calls with GPS (Global		

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
	Positioning System) incorporated in GIS (Geographical Information System) with GSM/ GPRS (Global System for Mobile Communication/ General Packet Radio Service) integrated Ambulance monitoring and tracking system] who are holding valid Manufacturing/ Fabricating/ Service license with average annual turnover of not less that Rs.10.00 Crores, during the preceding three year are eligible to participate in the Tender. It includes consequent amendments regarding submission of necessary documents of the Ambulance Service Provider including amongst other its Registration Certificates/ Licenses with the concerned Authorities.		
8.2	In Consortium, experience of associated partners/ members shall also be considered subject to production of relevant and valid documentary evidences duly self-attested and notarized.		
8.3	CMC, wherever mentioned, shall be for a period of 5 years and includes CMC of all retrofitting and of all the Machinery & Equipments to be fitted in the vehicle as a result of retrofitting of ambulance as well as healthcare equipments to be fitted in the ambulances.		
8.4	The successful bidder shall conduct Training programs for Para-medics and Doctors as per the National/ State Guidelines for ALS/ BLS Trainings and to be certified by the HoD, Anaesthesia, Govt. Medical College, Jammu/ Srinagar		
8.5	Equipments shall be US FDA/ European CE/ BIS/ equivalent certified, wherever applicable. In addition, bidders can submit higher quality/ accreditation certificates, duly self-attested and notarized.		
8.6	JKMSCL shall issue 'C' form in respect of CST charged against valid Invoice submitted by the successful bidder for Capex portion. Successful bidder has to ensure applicability of various taxes including VAT/ CST/ Service Tax leviable at the time of supply of ambulances, duly retrofitted alongwith Healthcare Equipments or during supply of services (Operationalization of Call Centre/ Supply of manpower).		
8.7	Any of the intended bidder can inspect the existing GPS/ VTMS call-centres/ make & model of existing ambulances to be fabricated after seeking written permission from MD, JKMSCL		
9	Instructions/ Clarifications regarding BoQ		
9.1	Rates for Healthcare Equipments (reqd. for ALS/ BLS as per Annexures) shall have to be quoted collectively on per Unit basis i.e., One combined rate for all the equipments/ supplies to be made available in one ambulance, keeping in consideration the requisite provision of all Consumables/ Tools/ Other Supplies required in the ambulances. Bidder shall have to submit Quantity-wise details in respect of same (without mentioning the Rates) alongwith Technical Bid.		
9.2	Rates for Hardware (A.3.1.), Software (A.3.2) and Other Cost (A.3.3) reqd. for establishment of each Call Centre shall have to be quoted collectively on per Unit basis i.e., One combined rate for all the respective items required for establishment		

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
	of each Call Centres Bidder shall have to submit Quantity-wise details of items in respect of same (without mentioning the Rates) alongwith Technical Bid.		
9.3	Rates for Manpower and Running & Maintenance Cost of Ambulances shall have to be quoted on per Month per Ambulance basis i.e., One combined monthly rate for all the manpower and running & maintenance cost of ambulance to be made available/ required in one ambulance (ALS/ BLS) to ensure its operationalization on 24x7 basis. Bidder shall have to submit details in respect of manpower and running & maintenance (without mentioning the Rates) alongwith Technical Bid.		
9.4	Rates for Manpower and Running & Maintenance Cost of Call Centres have to be quoted on per Month per Call Centre basis i.e., One combined monthly rate for all the manpower and running & maintenance cost of call centre to be made available/ required in one call centre to ensure its functioning on 24x7 basis. Bidder shall have to submit details in respect of manpower and running & maintenance (without mentioning the Rates) alongwith Technical Bid.		
9.5	Rates for various Capital Items (e.g., Healthcare Equipments, Hardware, Furniture & Fixtures, etc.) as well as Operational Items (e.g., Manpower, Running & Maintenance Cost) shall have to be quoted on composite basis (as above). However, JKMSCL may ask for break-up of Rates from bidders at any point of time.		
9.6	All the scrap and usable equipments coming out of ambulance as a result of retrofitting shall be the property of the bidder, the value of which should be adjusted in BoQ.		

**Sd/-
General Managing (Admn.)**

Copy to the:

- 1 Managing Director, JKMSCL
- 2 FA/ CAO, JKMSCL

: for information
: for information

General Managing (Admn.)