



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

Public Sector Undertaking of Government of Jammu & Kashmir

Jammu Office: Plot No. 9, Transport Nagar, Jammu – 180004 (J&K) India

Srinagar Office: 121-Green Avenue, Hyderpora, Srinagar - 190014 (J&K) India

CORRIGENDUM

Sub.: Corrigendum in respect of RFP for Purchase of Critical Care Ambulances in the State – reg.

In light of the representations received from the prospective bidders and discussions held during pre-bid meeting held on 10th Dec., 2015 in respect of “**Purchase of Critical Care Ambulances in the State**”, following modifications have been made in the RFP:

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
1	Scope of Work		
1.1	Clause (3)(II)(d)	Comprehensive Maintenance Contract (CMC) for a period of five years from the end of guarantee period	Comprehensive Maintenance Contract (CMC) for a period of five years from the end of guarantee period including ambulances, all the Machinery & Equipments to be fitted in the ambulances as well as healthcare equipments to be fitted in the ambulances
2	Terms & Conditions of Tender and Contract		
2.1	Guarantee Clause - (11)(ii)	In case of machinery and equipment guarantee shall have to be given that the Tenderer shall during the guarantee period replace parts if any and remove the manufacturing defect if found during the above period so as to make the machinery and equipment operative	Guarantee includes all the Machinery & Equipments to be fitted in the vehicle including as a result of retrofitting of ambulance as well as healthcare equipments to be fitted in the ambulances

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
2.2	Terms of Payment - Clause 19(e)	Rest 10% payment will be made against the 5% bank guarantee amount of the total value of the supply order which will be released after guarantee period.	10% Payment shall be withheld in case the bidder is failed to furnish the Performance Bank Guarantee and the same shall be released after Guarantee Period. In case, bidder has furnished Bank Guarantee, as required, 100% payment shall be released.
3	Specification for Retrofitting of Ambulances		
3.1	Diesel Engine and Power Train	Alternator capacity output not less than 90 amp	Alternator capacity output shall be 135 amp or more
3.2	Overall Dimensions	The overall width of the ambulance should be 2000 to 2100 mm	The overall width of the ambulance should be 1900 to 2100 mm
4	Specifications of Equipments		
4.1	List of Healthcare Equipments (for BLS/ ALS Ambulances and Critical Care Ambulances)	Oxygen - System including Oxygen cylinder portable and Oxygen Key	Oxygen - System including Oxygen Cylinder D Type, Oxygen cylinder portable and Oxygen Key
4.2	Standards and Safety	FDA/ CE	US FDA/ European CE
4.3	Warranty & Maintenance	3 Years	5 Years including CMC of all retrofitting and of all the Machinery & Equipments to be fitted in the vehicle as a result of retrofitting of ambulance as well as healthcare equipments to be fitted in the ambulances
4.4	Auto Loader – Collapsible stretcher	Automatic Loading, made of aluminium alloy	Automatic Loading, made of Aluminium Alloy/ Stainless Steel
4.5	AC/ DC/ Foot Operated - Suction	Collection bottle 600 ml	Collection bottle 600 ml or more capacity and e-

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
	Pump	capacity and e-Sterilizable	Sterilizable
4.6	(1) Monitor Defibrillator - Technical Characteristics - (1.1)(3)	Should be able to deliver shock from 50-200 joules in biphasic mode via metal chest pads	Should be able to deliver shock from 2-200 joules in biphasic mode via metal chest paddles
4.7	(1) Monitor Defibrillator - Technical Characteristics - (1.1)(5)	The whole system should have an inbuilt recorder; Telemetry Not Recommended	The whole system should have an inbuilt recorder
4.8	(1) Monitor Defibrillator - Energy Source - Battery operated (3.2)	Rechargeable battery backup of approximately 5 hours	Rechargeable battery backup of at least 2.5 hours/ 100 shocks
4.9	(1) Monitor Defibrillator - Energy Source - Battery operated (3.5)	Should not be more than 160 W	Should be more than 160 W
4.10	(1) Monitor Defibrillator - (4) Accessories, Spare Parts, Consumables - (4.2) Consumables/ Reagents (Open, Closed System)	ECG Cable; Recording Paper Rolls; Disposable Pads	ECG Cable - 1 Inbuilt plus 2 Spare; Recording Paper Rolls for 100 Patients and Disposable Pads – 50
4.11	(2) Syringe Pump - Technical Characteristics - (1.1)	Must work on commonly available 20, 50 and 60 ml syringes	Must work on 10-50 ml syringes
4.12	(2) Syringe Pump - Technical Characteristics - (1.1)	Maximum pressure generated \leq 20 psi	Deleted
4.13	(2) Syringe Pump - Technical Characteristics (1.1)	Comprehensive alarm package required including: occlusion alarm, near end of infusion pre-alarm and alarm, volume limit pre-alarm and alarm, low battery	Comprehensive alarm package required including: occlusion alarm, near end of infusion alarm, volume limit alarm, low battery alarm, AC power failure,

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
		pre-alarm and alarm, AC power failure, drive disengaged, syringe loading error, maintenance required	drive disengaged, syringe loading error, maintenance required
4.14	(2) Syringe Pump - Technical Characteristics - Settings (1.2)	Double loadable with one syringe of minimum 20m	Single loadable with one syringe of minimum 20ml
4.15	(2) Syringe Pump - Physical Characteristics - Weight (lbs, kg) (2.2)	<500gm	Weight (lbs, kg) = up to 5 kg
4.16	(3) Transport Ventilator - Technical Characteristics - Settings (1.2)	Tidal volume up to 100 ml	Tidal volume up to 50-2000 ml
4.17	(3) Transport Ventilator - Technical Characteristics - (1.1)(8)	Visual and audible alarms Accessories and tubing should be supplied for adult, pediatric & neo-natal size requirements	Visual and audible alarms Accessories and tubing should be supplied for adult & paediatric size requirements
4.18	(3) Transport Ventilator - Settings - (1.2)(b)	Pressure (inspiratory) up to 80 cm H ₂ O	
4.19	(3) Transport Ventilator - Settings - (1.2)(c)	Volume (inspiratory) up to 120 l/min	Flow (inspiratory) up to 120 l/min
4.20	(3) Transport Ventilator - Settings - (1.2)(f)	PEEP up to 20 cm H ₂ O	PEEP up to 20 cm H ₂ O or more
4.21	(3) Transport Ventilator - Settings - (1.2)(h)	FiO ₂ between 21 to 100 %	FiO ₂ between 40 to 100 %
5	General Clarifications		
5.1	In Consortium, experience of associated partners/ members shall also be considered subject to production of relevant and valid documentary evidences duly self-attested and notarized.		

S. No.	Particulars/ Issue(s) pertaining to	Earlier Specification	Revised Specification/ Clarification
5.2	CMC, wherever mentioned, shall be for a period of 5 years and includes CMC of all retrofitting and of all the Machinery & Equipments to be fitted in the vehicle as a result of retrofitting of ambulance as well as healthcare equipments to be fitted in the ambulances.		
5.3	Equipments shall be US FDA/ European CE/ BIS/ equivalent certified, wherever applicable. In addition, bidders can submit higher quality/ accreditation certificates, duly self-attested and notarized.		
5.4	JKMSCL shall issue 'C' form in respect of CST charged against valid Invoice submitted by the successful bidder. Successful bidder has to ensure applicability of various taxes including VAT/ CST Tax leviable at the time of supply of ambulances, duly fitted alongwith Healthcare Equipments.		
6	Instructions/ Clarifications regarding BoQ		
6.1	Rates for Healthcare Equipments shall have to be quoted collectively on per Unit basis i.e., One combined rate for all the equipments/ supplies to be made available in one ambulance, keeping in consideration the requisite provision of all Consumables/ Tools/ Other Supplies required in the ambulances. Bidder shall have to submit Quantity-wise details in respect of same (without mentioning the Rates) alongwith Technical Bid.		
6.2	Rates for Healthcare Equipments shall have to be quoted on composite basis (as above). However, JKMSCL may ask for break-up of Rates from bidders at any point of time.		

**Sd/-
General Managing (Admn.)**

Copy to the:

- 1 Managing Director, JKMSCL
- 2 FA/ CAO, JKMSCL

: for information
: for information

General Managing (Admn.)