



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of Govt of Jammu & Kashmir)

Corporate Head Office: Plot No: 58, Friends Colony, Satyam Road, Trikuta Nagar Jammu

Corporate Office: Opp State Motor Garages Near Haj House, Bemina, Srinagar

Telephone: 0191-2580842, Fax: 0191-2581845 (Jammu); Telefax: 0194-2432008 (Srinagar)

NOTICE


Sub: Shortcomings / minor infirmities of tenders invited for the procurement of **"Medical N-95 Masks for COVID-19 without valve with Head Strings"** related to COVID-19

Ref: Tenders invited for procurement of COVID Logistics" under reference no: JKMSCL/ COVID/ 2020/ 506 DATED: 08-01-2022

The evaluation of technical documents against aforesaid NIT was done by the Sub-Committee. Some documents from the participating bidders, are found deficit as per the Notice Inviting Bid. With reference to clause 2.1.9- Chapter II of Standard Procurement Procedure approved during 2nd Board Meeting, the under mentioned shortcomings comes under minor infirmities. It is therefore, impressed upon the under mentioned firms to upload the shortcomings/Clarifications asked for, against each, before **2:00 p.m. of 28-01-2022**, failing which the e-bids of the respective firms shall be liable to be rejected. The desired documents may be submitted on official mail ids gmjkscl.mi@gmail.com.

S. No	Name of the bidder	Shortcomings / Infirmities	Recommendations
1.	Paramount Surgimed Limited	Copies of Audited Profit Loss Account for last three financial years from Chartered Accountant not submitted	To be submitted.
		Quality compliant certificate indicating filtration efficiency of 0.3 micron, certified by BIS or any Govt testing laboratories , of the same batch /model as the sample submitted /supplies executed	To submit fresh and valid BIS certificate
		Quality certificate of BIS /NIOSH /EN149 FFP-2.	to be submitted
		The manufacturer of Medical N-95Masks should obtain the requisite certification from Bureau of Indian Standards	To submit fresh and valid BIS certificate
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
2.	Megha International	Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 test alongwith

			supporting approval of Govt. Organization.
3.	Control Print Limited	Bid security in the shape of declaration not submitted	to be submitted
		Latest GST Return of Bidder not submitted.	Latest GST Return of Bidder to be submitted
		Copies of Audited Profit Loss Account for the financial year 2019-20 from CA not submitted	to be submitted
		Product permission by the licensing authority for the products to be quoted, where ever applicable	to be submitted
		Quality compliant certificate indicating filtration efficiency of 0.3 micron, certified by BIS or any Govt testing laboratories , of the same batch /model as the sample submitted /supplies executed	to be submitted
		The manufacturer of Medical N-95Masks should obtain the requisite certification from Bureau of Indian Standards (Reference: Ministry of Chemical & Fertilizers, GOI. Department of Pharmaceutical, National Pharmaceutical Pricing authority dated 23-08-2020).	to be submitted
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on b) ISO 22609.	To be submitted
4.	Indiattitude Marketing Pvt Ltd	Sample not submitted hence not evaluated.	
5.	Jai Durga Medical and Surgical Agencies (Bidder) Manufacturer: M/s Alliance Mediequip Pvt. Ltd. New Delhi.	Non Conviction certificate to be submitted by M/s Jai Durga Medical and Surgical Agencies (only blacklisting submitted)	to be submitted
		Product permission by the licensing authority for the products to be quoted, where ever applicable.	to be submitted
		Quality compliant certificate indicating filtration efficiency of 0.3 micron, certified by BIS or any Govt testing laboratories ,	to be submitted



		of the same batch /model as the sample submitted /supplies executed	
		Quality certificate of BIS /NIOSH /EN149 FFP-2.	to be submitted
		The manufacturer of Medical N-95Masks should obtain the requisite certification from Bureau of Indian Standards (Reference: Ministry of Chemical & Fertilizers, GOI. Department of Pharmaceutical, National Pharmaceutical Pricing authority dated 23-08-2020).	to be submitted
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on b) ISO 22609.	To be submitted
6.	Vijay Sabre Safety Pvt Ltd.	Sample not submitted hence not evaluated.	
7.	Rajshree Fabrics	Sample not submitted hence not evaluated.	
8.	Precitex Equipments Private Limited (Bidder) Manufacturer Livsafe Products Pvt. Ltd.	Average Annual Turnover Statement not less than 1.00 (ONE) crores of Original Manufacturer/ Importer for 3 financial Years from Chartered Accountant on letter Head of CA with UDIN. (submitted is of bidder)	To be submitted of M/s Livsafe Products Pvt. Ltd.
		Copies of Audited Balance Sheet & Profit Loss Account for last three financial years from Chartered Accountant. (submitted is of bidder)	To be submitted of Livsafe Products Pvt. Ltd.
		Latest Non Conviction Certificate issued by the Licensing authority of the respective state / Self certified on letter Head duly sealed and signed of manufacturer.	to be submitted
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum	To be submitted

9.	Swami Textiles Pvt Ltd	80mmHg pressure based on b) ISO 22609.	
		Latest GST return of Bidder	to be submitted
		Copies of Audited Balance Sheet & Profit Loss Account for the financial year 2020-21 from Chartered Accountant not submitted.	Copies of Audited Balance Sheet & Profit Loss Account for the financial year 2020-21 from Chartered Accountant to be submitted.
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
10.	Dhawan Communication Pvt. Ltd (Bidder & Manufacturer)	Fluid Resistance: Minimum 80mmHg pressure based on b) ISO 22609.	To be submitted
		Average Annual Turnover Statement not less than 1.00 (ONE) crores of Original Manufacturer/ Importer for 3 financial Years from Chartered Accountant on letter Head of CA with UDIN. Not submitted	to be submitted
		Latest GST Return of Bidder	to be submitted
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on b) ISO 22609.	To be submitted
11.	Novex Healthcare (Bidder) Manufacturer Associates Packaging Delhi	Average Annual Turnover Statement not less than 1.00 (ONE) crores of Original Manufacturer/ Importer for 3 financial Years from Chartered Accountant on letter Head of CA with UDIN. (submitted is of bidder)	To be submitted of M/s Associates Packaging Delhi.
		Copies of Audited Balance Sheet & Profit Loss Account for last three financial years from Chartered Accountant. (submitted is of bidder)	To be submitted of M/s Associates Packaging Delhi
		Copy of the PAN Card of the bidder	To be submitted
		Fluid Resistance: Minimum 80mmHg pressure based on	To submit test certificate from Govt. Institution or from a

		a) ASTM F 1862	Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on b) ISO 22609.	To be submitted
12.	Bansal Trading Company (Bidder) Manufacturer Kamtronics Technology Pvt. Ltd.	Product permission by the licensing authority for the products to be quoted, where ever applicable.	to be submitted
		Quality compliant certificate indicating filtration efficiency of 0.3 micron, certified by BIS or any Govt testing laboratories , of the same batch /model as the sample submitted /supplies executed. BIS stand expired on 02-11-2021	To submit fresh and valid BIS certificate
		The manufacturer of Medical N-95Masks should obtain the requisite certification from Bureau of Indian Standards	To submit fresh and valid BIS certificate
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on ISO 22609.	To be submitted
13.	VA Technologies (Bidder) Manufacturer M/s Sanjay Trading Co.	Product permission by the licensing authority for the products to be quoted, where ever applicable.	to be submitted
		Quality compliant certificate indicating filtration efficiency of 0.3 micron, certified by BIS or any Govt testing laboratories , of the same batch /model as the sample submitted /supplies executed.	To submit fresh and valid BIS certificate
		Quality certificate of BIS /NIOSH /EN149 FFP-2.	to be submitted
		The manufacturer of Medical N-95Masks should obtain the requisite certification from Bureau of Indian Standards	To submit fresh and valid BIS certificate
		Fluid Resistance: Minimum 80mmHg pressure based on	To submit test certificate from Govt. Institution or from a

		a) ASTM F 1862	Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on ISO 22609.	To be submitted
14.	Mann Electronics (Bidder) Manufacturer M/s Manohar Filaments Pvt. Ltd	As per the documents submitted by the Bidder, the manufacturer of TLM is M/s Manohar Filaments Pvt. Ltd. which reflects that the bidders is not an OEM for TLM whereas the bidder has availed the benefit of MEME by not submitting the tender fee.	Cannot be considered for further evaluation as the bidder has not submitted the tender fee as required.
15.	3V Shoppe (Bidder) Manufacturer Sanson Promo Inovoation LLP.	Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on ISO 22609.	To be submitted
16.	Jay Pee Knit Fab Manufacturer M/s Neel Kanth Packaging	Latest Non Conviction Certificate issued by the Licensing authority of the respective state / Self certified on letter Head duly sealed and signed of manufacturer not submitted :	to be submitted
		Product permission by the licensing authority for N-95 Masks not submitted	to be submitted
		Quality compliant certificate indicating filtration efficiency of 0.3 micron, certified by BIS or any Govt testing laboratories not submitted	to be submitted
		Quality certificate of BIS /NIOSH /EN149 FFP-2	to be submitted
		BIS attached with Technical Bid was valid till 11-10-2021.	To be submitted fresh and valid certificate
		Fluid Resistance: Minimum 80mmHg pressure based on ASTMF 1862... submitted from Shree Ram Testing Laboratory .	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Copies of Audited Balance Sheet & Profit Loss Account for	to be submitted

		financial year 2019-20 from Chartered Accountant of Original Manufacturer not submitted	
17.	M/S Sanmati Packaging	Sample not submitted hence not evaluated.	
18.	Medisign Pharmaceuticals Distributors (Bidder) 'A' Manufacturer Venus Safety & Health Pvt. Ltd.	Latest Non Conviction Certificate issued by the Licensing authority of the respective state / Self certified on letter Head duly sealed and signed of manufacturer not submitted	To be submitted
		Declaration of Non Blacklisting and Non Conviction of bidder as per Format to be submitted (only Non Blacklisting submitted with technical bid)	to be submitted
		Product permission by the licensing authority for the products to be quoted, where ever applicable.	to be submitted
		Latest GST Return of Bidder	to be submitted
		Quality compliant certificate indicating filtration efficiency of 0.3 micron, certified by BIS or any Govt testing laboratories , of the same batch /model as the sample submitted /supplies executed.	To submit fresh and valid BIS certificate
		Quality certificate of BIS /NIOSH /EN149 FFP-2. Not submitted	To be submitted
		The manufacturer of Medical N-95Masks should obtain the requisite certification from Bureau of Indian Standards	To submit fresh and valid BIS certificate
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on ISO 22609 .	To be submitted
19.	M/S L M Agencies (Bidder)	Bid security in the shape of declaration not submitted	to be submitted
		Copies of Audited Balance Sheet & Profit Loss Account	to be submitted

	Manufacturer: D.K. Power Engineering Pvt. Ltd.	for last three financial years from Chartered Accountant of Original Mfr. not submitted	
		Latest Non Conviction Certificate issued by the Licensing authority of the respective state / Self certified on letter Head duly sealed and signed of manufacturer not submitted .	to be submitted
		Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTMF 1862 alongwith supporting approval of Govt. Organization.
		Fluid Resistance: Minimum 80mmHg pressure based on ISO 22609.	To be submitted
20.	Adish Healthcare Services	Sample not submitted hence not evaluated.	
21.	M/S New Alpine Traders (Bidder) Manufacturer Vizin India Pvt. Ltd.	Bid security in the shape of declaration not submitted	to be submitted
		Copies of Audited Balance Sheet & Profit Loss Account for last three financial years from Chartered Accountant of M/s Vizin India Pvt. Ltd. not submitted	to be submitted
22.	Aastha Medical Devices (Bidder) Manufacturer Dress well Surgical India	Bid security in the shape of declaration not submitted	to be submitted
		Average Annual Turnover Statement not less than 1.00 (ONE) crores of Original Manufacturer/ Importer for 3 financial Years from Chartered Accountant on letter Head of CA with UDIN not submitted.	to be submitted
		Copies of Audited Balance Sheet & Profit Loss Account for last three financial years from Chartered Accountant of Original Manufacturer not submitted	to be submitted
		Latest Non Conviction Certificate issued by the Licensing authority of the respective state / Self certified on letter Head duly sealed and signed of manufacturer not submitted .	to be submitted
		Permission to manufacture	to be submitted

	N-95 Mask from competent authority not submitted	
	Latest GST Return of Bidder not submitted.	to be submitted
	Quality compliant certificate indicating filtration efficiency of 0.3 micron, certified by BIS or any Govt testing laboratories, of the same batch /model as the sample submitted /supplies executed. BIS stand expired on 02-11-2021	To submit fresh and valid BIS certificate
	Quality certificate of BIS /NIOSH /EN149 FFP-2. Not submitted	To be submitted
	The manufacturer of Medical N-95 Masks should obtain the requisite certification from Bureau of Indian Standards	To submit fresh and valid BIS certificate
	Fluid Resistance: Minimum 80mmHg pressure based on a) ASTM F 1862	To submit test certificate from Govt. Institution or from a Laboratory duly authorized by the Govt. for conducting ASTM F 1862 alongwith supporting approval of Govt. Organization.
	Fluid Resistance: Minimum 80mmHg pressure based on ISO 22609.	To be submitted


General Manager (P&S)
JKMSCL

No: JKMSCL/2022/506/TLM/ 10216-20

Dated:- 25 /01-2022

Copy for information to the:-

1. Managing Director JKMSCL
2. FA&CAO, JKMSCL
3. GM (Adm) JKMSCL
4. M/s Say Technologies for uploading of Notice on www.jkmsclbusiness.com
5. Office Record File.