



# JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House, Bemina (190018), Srinagar

email: gmjjkmscl@gmail.com, jkmsclepm@gmail.com; website: www.jkmscl.nic.in

## MINUTES OF MEETING

**Subject: Meeting of the Technical Evaluation/Advisory Committee held on 26.08.2025 for evaluation of technical bid(s) for the finalization of Rate Contract for the procurement of "Machinery & Equipment" under NIT No. JKMSCL/Mach/2024/645 dated 02.12.2024.**

1. E- bid for the finalization of **Rate Contract for the procurement of Machinery & Equipments** was uploaded vide no NIT No: NIT/JKMSCL/Mach/2024/645 Dated 02.12.2024 and is opened on 17<sup>th</sup> March 2025 at Corporate Head Office, Jammu and Srinagar at 11.00 A.M.
2. With reference to clause 2.1.9 Chapter II of the Standard Procurement Procedure approved during 2<sup>nd</sup> Board Meeting, it was decided by the technical advisory/evaluation committee that the Chairman, Technical Advisory/Evaluation Committee shall get the preliminary evaluation done by the sub-committee and then shall get the minor infirmities asked from the respective bidders, if any. After downloading the minor infirmities submitted by the bidders, the same shall be re-evaluated by the Sub Committees followed by technical evaluation by the technical experts and the final report shall be placed before the Technical Advisory Committee for further decision/recommendation.
3. Further, sub committees constituted to assist the Technical Evaluation Committee of JKMSCL completed the compilation and scrutiny/evaluation process and thereafter the observations made by the sub- committee were placed before Technical Advisory Committee constituted for the purpose by Board of Director, JKMSCL for final scrutiny/evaluation of bid documents so as to arrive at qualifying bidders for further evaluation by means of power point presentation/ demonstration and catalogue evaluation by the technical experts

### Item Name: Stand Alone Microwave Mobile Waste Disinfectant system.

S. No.	Name of the bidder	Shortcomings	Status	Remarks
1/3	M/s Precision Electronics Instruments & Components, Maharashtra <b>(Bidder/Manufacturer/MSME)</b>  <b>Model No. VIROSURE 400.</b>	<b>To Submit:</b> 1. Latest Satisfactory performance certificate of Purchase order placed for NTPC Limited or Another other Purchase order from any Govt. institution with latest satisfactory performance of year 2022-23. 2. The product should have validation Test report for efficacy of all <b>8 cycles mentioned in point 20 (a) to 20 (h) above</b> , from any one of the national Laboratories	Satisfactory performance certificate of purchase order placed for NTPC Limited submitted wherein stamp on the certificate is seems to be forged ( <b>copy enclosed</b> ). However verification of Stamp needs to be Verified from NTPC before proceeding in	<b>Recommended to TEC subject to the verification of digital stamp on satisfactory performance report dated 15.10.2024 from NTPC Limited.</b>

Handwritten signatures and initials at the bottom of the page.

		<p><b>CSIR/ICMR/NCDC.</b> <b>OR</b> The product must have EU CE product quality certificate with <b>4 digit notified body number.</b> Self attested <b>notarized</b> copies of all such test reports/certificates must be submitted with technical bid.</p> <p>3. The manufacturer should have ISO 9001: 2015, ISO 13485:2016 QMS for Medical Devices certificate from <b>NABCB/IAF Accredited agency.</b> (Submitted certification is not accredited by NABCB/IAF)</p> <p>4. The product should comply with GSR Notification 343(E) and must have dedicated cycles for disinfection operated through one touch operations, viz. Blood bags, Microbiology Waste, Bactec / Culture Bottles etc so that compliance of records/ cycle can be recorded with the time and date of treatment in compliance to the aforesaid GSR notification.</p> <p>5. Copy of CDSCO Registration Certification for the quoted item and model as the item is included in <b>Class-B of Medical devices.</b></p>	<p>the matter.</p> <p>Submitted</p> <p>Submitted</p> <p>Statement to Comply submitted.</p> <p>Submitted</p>	
2/3	<p>M/s S.R Technomed, Jammu. <b>(Bidder)</b></p> <p>M/s SS Medical System (I) Pvt. Ltd., Uttarakhand. <b>(Manufacturer)</b></p> <p><b>Model: OptiMaser® 30.</b></p>	<p><b>To Submit:</b></p> <p>1. The product should have validation Test report for efficacy of all <b>8 cycles mentioned in point 20 (a) to 20 (h) above,</b> from any one of the national Laboratories CSIR/ICMR/NCDC. <b>OR</b> The product must have EU CE product quality certificate with 4 digit notified body number. Self attested notarized copies of all such test reports/certificates must be submitted with technical bid.</p>	<p>Submitted.</p> <p><b>Submitted certificate issued after the opening of bid and for the quoted product</b></p>	<p><b>Not Recommend ed to TEC.</b></p>

Handwritten signatures and initials at the bottom of the page.



		<p>2. The manufacturer should have ISO 9001: 2015, ISO 13485:2016 QMS for Medical Devices certificate from NABCB/IAF Accredited agency. <b>(Submitted certification is not accredited by NABCB/IAF)</b> Submitted ISO-13485: 2016 is of IAF/NABCB accredited body. However the quoted item has not been listed in the said certificate.)</p> <p>3. The product should comply with GSR Notification 343(E) and must have dedicated cycles for disinfection operated through one touch operations, viz. Blood bags, Microbiology Waste, Bactec / Culture Bottles etc so that compliance of records/ cycle can be recorded with the time and date of treatment in compliance to the aforesaid GSR notification.</p> <p>4. Copy of CDSCO Registration Certification for the quoted item and model as the item is included in Class-B of Medical devices. (Submitted Screenshot does not confirms the registration of the quoted product by CDSCO)</p>	<p>and model.</p> <p>Statement to Comply submitted.</p> <p>Submitted.</p>	
3/3	<p>M/s Shiv Shakti Enterprises, Mumbai. <b>(Bidder)</b></p> <p>M/s Azonic International, Mumbai. <b>(Manufacturer)</b></p> <p><b>Model: BWM-45.</b></p>	<p><b>To Submit:</b></p> <p>1. Since the payment has been made in the form of Demand Draft, Proof of Submission of Tender Processing charges and Bid Security to be submitted as mode of payment through DD for cost of tender/tender processing fees/Bid Security shall only be entertained if the same is deposited physically against proper receipt in the Corporate Office of JKMSCL, before the closing due date of e-bid.</p> <p>2. Compliance Sheet for each equipment (self-attested)</p> <p>3. Client Base on Letter Head of the Bidder / Manufacture/Indian Subsidiary of Principal Manufacturer with references of the supply orders, for any of the <b>three financial</b></p>	<p><b>Not Submitted all the Infirmities</b></p>	<p><b>Not Recommend ed to TEC.</b></p>

70 8 12

years in last five financial years along with satisfactory performance certificate of minimum one Purchase Order and installation report of Govt. Institution (Copies of reference supply orders and satisfactory performance certificate need to be attached)

4. Average Annual Turnover Statement for Last 3 financial Years of Principal Manufacturer i.e. M/s Azonic International issued by Chartered Accountant/competent authority with UDIN (2020-21, 2021-22 and 2022-23). **(Submitted is not of OEM as per NIT)**
5. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the manufacturer i.e. M/s Azonic International for the year 2020-21, 2021-22 & 2022-23 with UDIN. **(Submitted is not of OEM)**
5. The product should have validation Test report for efficacy of all **8 cycles mentioned in point 20 (a) to 20 (h) above**, from any one of the national Laboratories CSIR/ICMR/NCDC.  
**OR**  
The product must have EU CE product quality certificate with 4 digit notified body number. Self attested notarized copies of all such test reports/certificates must be submitted with technical bid.
6. The manufacturer should have ISO 9001: 2015, ISO 13485:2016 QMS for Medical Devices certificate from NABCB/IAF Accredited agency. (Submitted certification is not accredited by NABCB/IAF) Submitted ISO-13485: 2016 is of IAF/NABCB accredited body.
7. The equipment should be type approved with safety Test reports from **NABL /ILAC accredited** test Laboratory.
8. The product should comply with GSR Notification 343(E) and





		<p>must have dedicated cycles for disinfection operated through one touch operations, viz. Blood bags, Microbiology Waste, Bactec / Culture Bottles etc so that compliance of records/ cycle can be recorded with the time and date of treatment in compliance to the aforesaid GSR notification.</p> <p>9. Name, photograph &amp; specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.</p> <p>10. Specify point of supply with full Address.</p> <p>11. Copy of CDSCO Registration Certification for the quoted item and model as the item is included in <b>Class-B of Medical devices.</b></p> <p><b>Note: The Submitted all of the quality certifications in the tender document are not verifiable. Submit the certifications as per NIT which can be verified by this office as well)</b></p>		
--	--	---	--	--

## 2. Item Name: Full body UV Therapy for vitiligo and Psoriasis. (ME2357)

1/1	<p>M/s Merino International, New Delhi. <b>(Bidder/Manufacturer/MSME)</b></p> <p><b>Model No. GRUVC-44B.</b></p>	<p><b>To Submit:</b></p> <p>1. Client Base on Letter Head of the Bidder / Manufacture of Principal Manufacturer with references of the supply orders, for any of the three financial years in last five financial years along with satisfactory performance certificate of minimum one Purchase Order and installation report of Govt. Institution (Copies of reference supply orders and satisfactory performance need to be attached) <b>(Submitted only for one year i.e. 2023-24 of RML Hospital. Submit for two more years other than 2023-24 as per NIT)</b></p> <p>2. Name, photograph &amp; specimen signature of the designated officer/</p>	<p>Submitted.</p> <p>Submitted.</p>	<p><b>Recommended to TEC.</b></p>
-----	--	---	-------------------------------------	-----------------------------------

		<p>representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.</p> <p>3. Specify point of supply with full Address.</p> <p>4. Copy of CDSCO License for the quoted product and model as the product comes under Medical devices Class-II.</p>	<p>Submitted.</p> <p>Submitted.</p>	
--	--	--	-------------------------------------	--

### 3. Item Name: Cold Surface for Blood component labeling and sorting.

1/1	<p>M/s Merino International, New Delhi. <b>(Bidder/Manufacturer/MSME)</b></p> <p><b>ModelNo. QCC-350.</b></p>	<p><b>To Submit:</b></p> <ol style="list-style-type: none"> <li>1. Client Base on Letter Head of the Bidder/ Manufacture with references of the supply orders, for any of the three financial years in last five financial years (Copies of reference supply orders need to be attached) (Submitted only for one financial year 2022-23 only. Other submitted Purchase Orders are not for the quoted products)</li> <li>2. Average Annual Turnover Statement for the financial Year 2020-21 of the Principal Manufacturer issued by Chartered Accountant with UDIN.</li> <li>3. Copies of Audited Balance sheet &amp; profit loss account for the financial year 2020-21 certified by Chartered Accountant of principal manufacturer with UDIN.</li> <li>4. Quality Certifications on the quoted product and model viz. Indian CE &amp; ISO. (Submitted ISO is expired and CE is not for the quoted product and model)</li> <li>5. Certificate of Compliant to ADR/RID/IMDG/ ICAOTI/IATA GDR mandatory to be enclosed.</li> <li>6. Product should be in use from more than 5 years, successful user certificate and order copies to be attached as proof.</li> </ol>	<p>Submitted</p> <p>Submitted.</p> <p>Submitted</p> <p>Submitted.</p> <p>Clarified and submitted.</p> <p>Submitted.</p> <p>Submitted.</p>	<p><b>Not Recommended to TEC.</b></p>
-----	---	--	---	---------------------------------------

*M 7 e 8*



		<p>7. Company should be involved in manufacturing this product for more than 5 years. (Submit documentation proof)</p> <p>8. Name, photograph &amp; specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.</p> <p>9. Specify point of supply with full Address.</p> <p>10. <b>Copy of CDSCO Registration Certificate for the quoted product and model.</b></p>	<p>Submitted.</p> <p>Submitted</p> <p><b>Not Submitted</b></p>	
--	--	--	--	--

#### 4. Item Name: Blood Transport Box (GMCU03)

1/2	<p>M/s Tropicool Car Gadgets P Ltd., Maharashtra. <b>(Bidder/Manufacturer/MSME)</b></p> <p><b>Model No. Tropicool TFF-50 with Battery Backup.</b></p>	<p><b>To Submit:</b></p> <ol style="list-style-type: none"> <li>1. Tender Processing charges not Submitted. Clarify and submit the proof of submission of tender processing charges, if already submitted.</li> <li>2. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Principal Manufacturer as well as by bidder as per proforma duly notarized.</li> <li>3. Copies of Audited Balance sheet &amp; profit loss account for last three financial years certified by Chartered Accountant of the Manufacturer for 2020-21, 2021-22 &amp; 2022-23 with UDIN along with turnover. <b>(Submitted Balance sheets are without UDIN and also the figures mismatch with turnover submitted).</b></li> <li>4. Latest GST Returns of the Bidder</li> <li>5. Name, photograph &amp; specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.</li> <li>6. Specify point of supply with full Address.</li> <li>7. Declaration of bidder regarding acceptance Bid for terms &amp; conditions (Annexure-A1).</li> </ol>	<p><b>Tender processing charges not submitted. Hence not considered</b></p>	<p><b>Not Recommended to TEC.</b></p>
-----	---	--	---	---------------------------------------

*Handwritten signature/initials*

		<p>8. Copy of CDSCO License (MD-5) for the quoted product and model.</p> <p>9. CE certificate for the quoted product as the submitted certificate expired on 13.04.2025.</p> <p>10. ISO 13485 certificate for the quoted product as the submitted certificate is expired.</p>		
2/2	<p>M/s Mark Enterprises, Maharashtra. <b>(Bidder/Manufacturer/MSME)</b></p> <p><b>Model No. MACB-06.</b></p>	<p><b>To Submit:</b></p> <p>1. Tender Processing charges not Submitted. Clarify and submit the proof of submission of tender processing charges, if already submitted.</p> <p>2. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs100 furnished by the Principal Manufacturer as well as by bidder as per proforma duly notarized. <b>(Submit on non-judicial stamp paper)</b></p> <p>3. Client Base on Letter Head of the Bidder / Manufacture/Indian Subsidiary of Principal Manufacturer with references of the supply orders, for any of the three financial years in last five financial years along with satisfactory performance certificate of minimum one Purchase Order and installation report of Govt. Institution <b>(Copies of reference supply orders and satisfactory performance certificate need to be attached)</b> <b>(Submitted Purchase orders are not for quoted product and model)</b></p> <p>4. Nature of the Firm/Public Company/ Private Company/ Partnership/</p>	<p><b>Not Submitted as the submitted document could not be traced and was also not mentioned in jktender.</b></p> <p><b>Submitted Declaration is not as per Tendered Annexure format.</b></p> <p><b>Not submitted.</b></p> <p>Submitted.</p>	<p><b>Not Recommended to TEC.</b></p>

*Handwritten signature/initials*









		Proprietorship/any other with Documentary proof.		
		5. Copy of GST Registration of the Bidder	Submitted.	
		6. Latest GST Returns of the Bidder	Submitted.	
		7. Copy of the PAN Card of the Bidder	Submitted.	
		8. CE certificate for the quoted product and model.	Submitted.	
		9. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.	Submitted.	
		10. Specify point of supply with full Address.	Submitted.	

**Decision:** Accordingly, the Technical Advisory /Evaluation Committee accepted the report of the Sub Committee and recommended Power Point Presentation/Demonstration of the equipments before the Technical Experts. It was further recommended that the quality certifications are to be checked and verified at the time of demonstration and the authenticity and applicability of quality certification shall be decided by the technical experts strictly as per the terms & conditions envisaged in the technical specifications of the NIT.

Regarding the items which are not recommended by the experts, the committee decided to retender the said items.

The following officers attended the meeting:

S. No.	Name of the Officer	Designation	Signatures
1.	Mohd. Ashraf Choudhary,	General Manager (Adm)	
2.	Dr. Upinder Vasni	MO QC	
3.	Karan Singh	AAO	
4.	Shamندر Singh	HEO	
5.	Vishal Mahajan	Asst.	
6.	Ankur Kalota	FM	


**(Mohd. Ashraf Choudhary ) JKAS  
General Manager (Adm.)JKMSCL.**

Dated: 29 /08/2025

No.: JKMSCL/GM/2025/ 5303-12

Copy for information to the:-

1. Principal, Govt. Medical College, Srinagar (Kashmir).
2. Principal, Government Medical College, Jammu.
3. Principal, Government Dental College, Srinagar.
4. Principal, Government Dental College, Jammu.
5. Director General, Indian System of Medicines, J&K.
6. Managing Director, Jammu and Kashmir Medical Supplies Corporation Limited, Jammu
7. Director, Health Services, Jammu.
8. Director, Health Services, Kashmir.
9. Controller, Drug & Food Control Organization, J&K.
10. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited





# JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House, Bemina (190010), Srinagar

email: gm@jkm scl@gmail.com, jkm scl@pm@gmail.com; website: www.jkm scl.nic.in

## MINUTES OF MEETING

Subject: Meeting of the Technical Evaluation/Advisory Committee held on 26.08.2025 for evaluation of technical bid(s) for the finalization of Rate Contract for the procurement of "Machinery & Equipment" under NIT No. JKMSCL/Mach/2024/645 dated 02.12.2024.

1. E- bid for the finalization of Rate Contract for the procurement of Machinery & Equipments was uploaded vide no NIT No: NIT/JKMSCL/Mach/2024/645 Dated 02.12.2024 and is opened on 17<sup>th</sup> March 2025 at Corporate Head Office, Jammu and Srinagar at 11.00 A.M.
2. With reference to clause 2.1.9 Chapter II of the Standard Procurement Procedure approved during 2<sup>nd</sup> Board Meeting, it was decided by the technical advisory/evaluation committee that the Chairman, Technical Advisory/Evaluation Committee shall get the preliminary evaluation done by the sub-committee and then shall get the minor infirmities asked from the respective bidders, if any. After downloading the minor infirmities submitted by the bidders, the same shall be re-evaluated by the Sub Committees followed by technical evaluation by the technical experts and the final report shall be placed before the Technical Advisory Committee for further decision/recommendation.
3. Further, sub committees constituted to assist the Technical Evaluation Committee of JKMSCL completed the compilation and scrutiny/evaluation process and thereafter the observations made by the sub- committee were placed before Technical Advisory Committee constituted for the purpose by Board of Director, JKMSCL for final scrutiny/evaluation of bid documents so as to arrive at qualifying bidders for further evaluation by means of power point presentation/ demonstration and catalogue evaluation by the technical experts

### Item Name: Stand Alone Microwave Mobile Waste Disinfectant system.

S. No.	Name of the bidder	Shortcomings	Status	Remarks
1/3	M/s Precision Electronics Instruments & Components, Maharashtra (Bidder/Manufacturer/MSME)  Model No. VIOSURE 400.	<b>To Submit:</b> 1. Latest Satisfactory performance certificate of Purchase order placed for NTPC Limited or Another other Purchase order from any Govt. institution with latest satisfactory performance of year 2022-23.  2. The product should have validation Test report for efficacy of all 8 cycles mentioned in point 20 (a) to 20 (h) above, from any one of the national Laboratories	Satisfactory performance certificate of purchase order placed for NTPC Limited submitted wherein stamp on the certificate is seems to be forged (copy enclosed). However verification of Stamp needs to be Verified from NTPC before proceeding in	Recommend ed to TEC subject to the verification of digital stamp on satisfactory performance report dated 15.10.2024 from NTPC Limited.

		<p>CSIR/ICMR/NCDC. OR</p> <p>The product must have EU CE product quality certificate with <b>4 digit notified body number</b>. Self attested <b>notarized</b> copies of all such test reports/certificates must be submitted with technical bid.</p> <p>3. The manufacturer should have ISO 9001: 2015, ISO 13485:2016 QMS for Medical Devices certificate from <b>NABCB/IAF Accredited agency</b>. (Submitted certification is not accredited by NABCB/IAF)</p> <p>4. The product should comply with GSR Notification 343(E) and must have dedicated cycles for disinfection operated through one touch operations, viz. Blood bags, Microbiology Waste, Bactec / Culture Bottles etc so that compliance of records/ cycle can be recorded with the time and date of treatment in compliance to the aforesaid GSR notification.</p> <p>5. Copy of CDSCO Registration Certification for the quoted item and model as the item is included in <b>Class-B of Medical devices</b>.</p>	<p>the matter.</p> <p>Submitted</p> <p>Submitted</p> <p>Statement to Comply submitted.</p> <p>Submitted</p>	
2/3	<p>M/s S.R Technomed, Jammu. (Bidder)</p> <p>M/s SS Medical System (I) Pvt. Ltd., Uttarakhand. (Manufacturer)</p> <p>Model: OptiMaser® 30.</p>	<p><b>To Submit:</b></p> <p>1. The product should have validation Test report for efficacy of all <b>8 cycles mentioned in point 20 (a) to 20 (h) above</b>, from any one of the national Laboratories CSIR/ICMR/NCDC. OR</p> <p>The product must have EU CE product quality certificate with <b>4 digit notified body number</b>. Self attested notarized copies of all such test reports/certificates must be submitted with technical bid.</p>	<p>Submitted.</p> <p>Submitted certificate issued after the opening of bid and for the quoted product</p>	Not Recommended to TEC.



		<p>2. The manufacturer should have ISO 9001: 2015, ISO 13485:2016 QMS for Medical Devices certificate from NABCB/IAF Accredited agency. <b>(Submitted certification is not accredited by NABCB/IAF)</b> Submitted ISO-13485: 2016 is of IAF/NABCB accredited body. However the quoted item has not been listed in the said certificate.)</p> <p>3. The product should comply with GSR Notification 343(E) and must have dedicated cycles for disinfection operated through one touch operations, viz. Blood bags, Microbiology Waste, Bactec / Culture Bottles etc so that compliance of records/ cycle can be recorded with the time and date of treatment in compliance to the aforesaid GSR notification.</p> <p>4. Copy of CDSCO Registration Certification for the quoted item and model as the item is included in Class-B of Medical devices. (Submitted Screenshot does not confirms the registration of the quoted product by CDSCO)</p>	<p>and model.</p> <p>Statement to Comply submitted.</p> <p>Submitted.</p>	
3/3	<p>M/s Shiv Shakti Enterprises, Mumbai. <b>(Bidder)</b></p> <p>M/s Azonic International, Mumbai. <b>(Manufacturer)</b></p> <p><b>Model: BWM-45.</b></p>	<p><b>To Submit:</b></p> <p>1. Since the payment has been made in the form of Demand Draft, Proof of Submission of Tender Processing charges and Bid Security to be submitted as mode of payment through DD for cost of tender/tender processing fees/Bid Security shall only be entertained if the same is deposited physically against proper receipt in the Corporate Office of JKMSCL, before the closing due date of e-bid.</p> <p>2. Compliance Sheet for each equipment (self-attested)</p> <p>3. Client Base on Letter Head of the Bidder / Manufacture/Indian Subsidiary of Principal Manufacturer with references of the supply orders, for any of the three financial</p>	<p><b>Not Submitted all the Infirmities</b></p>	<p><b>Not Recommended to TEC.</b></p>

years in last five financial years along with satisfactory performance certificate of minimum one Purchase Order and Installation report of Govt. Institution (Copies of reference supply orders and satisfactory performance certificate need to be attached)

4. Average Annual Turnover Statement for Last 3 financial Years of Principal Manufacturer i.e. M/s Azonic International issued by Chartered Accountant/competent authority with UDIN (2020-21, 2021-22 and 2022-23). (Submitted is not of OEM as per NIT)

5. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the manufacturer i.e. M/s Azonic International for the year 2020-21, 2021-22 & 2022-23 with UDIN. (Submitted is not of OEM)

5. The product should have validation Test report for efficacy of all 8 cycles mentioned in point 20 (a) to 20 (h) above, from any one of the national Laboratories CSIR/ICMR/NCDC.

OR

The product must have EU CE product quality certificate with 4 digit notified body number. Self attested notarized copies of all such test reports/certificates must be submitted with technical bid.

6. The manufacturer should have ISO 9001: 2015, ISO 13485:2016 QMS for Medical Devices certificate from NABCB/IAF Accredited agency. (Submitted certification is not accredited by NABCB/IAF) Submitted ISO-13485: 2016 is of IAF/NABCB accredited body.

7. The equipment should be type approved with safety Test reports from NABL /ILAC accredited test Laboratory.

8. The product should comply with GSR Notification 343(E) and



		<p>must have dedicated cycles for disinfection operated through one touch operations, viz. Blood bags, Microbiology Waste, Bactec / Culture Bottles etc so that compliance of records/ cycle can be recorded with the time and date of treatment in compliance to the aforesaid GSR notification.</p> <p>9. Name, photograph &amp; specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.</p> <p>10. Specify point of supply with full Address.</p> <p>11. Copy of CDSCO Registration Certification for the quoted item and model as the item is included in Class-B of Medical devices.</p> <p><b>Note:</b> The Submitted all of the quality certifications in the tender document are not verifiable. Submit the certifications as per NIT which can be verified by this office as well)</p>		
--	--	--	--	--

**2. Item Name: Full body UV Therapy for vitiligo and Psoriasis. (ME2357)**

1/1	<p>M/s Merino International, New Delhi. (Bidder/Manufacturer/MSME)</p> <p><u>Model No.</u> <u>GRUVC-44B.</u></p>	<p><b>To Submit:</b></p> <p>1. Client Base on Letter Head of the Bidder / Manufacture of Principal Manufacturer with references of the supply orders, for any of the three financial years in last five financial years along with satisfactory performance certificate of minimum one Purchase Order and installation report of Govt. Institution (Copies of reference supply orders and satisfactory performance need to be attached) (Submitted only for one year i.e. 2023-24 of RML Hospital. Submit for two more years other than 2023-24 as per NIT)</p> <p>2. Name, photograph &amp; specimen signature of the designated officer/</p>	<p>Submitted.</p> <p>Submitted.</p>	<p>Recommended to TEC.</p>
-----	--	--	-------------------------------------	----------------------------

		<p>representative of the Bidder who is authorized to make correspondence with the JKMBCL, if any.</p> <p>3. Specify point of supply with full Address.</p> <p>4. Copy of CDSCO License for the quoted product and model as the product comes under Medical devices Class-II.</p>	<p>Submitted.</p> <p>Submitted.</p>	
<b>3. Item Name: Cold Surface for Blood component labeling and sorting.</b>				
1/1	<p>M/s Merino International, New Delhi. (Bidder/Manufacturer/MSME)</p> <p>ModelNo. QCC-350.</p>	<p><b>To Submit:</b></p> <p>1. Client Base on Letter Head of the Bidder/ Manufacture with references of the supply orders, for any of the three financial years in last five financial years (Copies of reference supply orders need to be attached) (Submitted only for one financial year 2022-23 only. Other submitted Purchase Orders are not for the quoted products)</p> <p>2. Average Annual Turnover Statement for the financial Year 2020-21 of the Principal Manufacturer issued by Chartered Accountant with UDIN.</p> <p>3. Copies of Audited Balance sheet &amp; profit loss account for the financial year 2020-21 certified by Chartered Accountant of principal manufacturer with UDIN.</p> <p>4. Quality Certifications on the quoted product and model viz. Indian CE &amp; ISO. (Submitted ISO is expired and CE is not for the quoted product and model)</p> <p>5. Certificate of Compliant to ADR/RID/IMDG/ ICAOTI/IATA GDR mandatory to be enclosed.</p> <p>6. Product should be in use from more than 5 years, successful user certificate and order copies to be attached as proof.</p>	<p>Submitted</p> <p>Submitted.</p> <p>Submitted</p> <p>Submitted.</p> <p>Clarified and submitted.</p> <p>Submitted.</p> <p>Submitted.</p>	<p>Not Recommended to TEC.</p>



	<p>7. Company should be involved in manufacturing this product for more than 5 years. (Submit documentation proof)</p> <p>8. Name, photograph &amp; specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.</p> <p>9. Specify point of supply with full Address.</p> <p>10. <b>Copy of CDSCO Registration Certificate for the quoted product and model.</b></p>	Submitted.	
		Submitted	
		Not Submitted	

**4. Item Name: Blood Transport Box (GMCU03)**

1/2	<p>M/s Tropicool Car Gadgets P Ltd., Maharashtra. (Bidder/Manufacturer/MSME)</p> <p>Model No. Tropicool TFF-50 with Battery Backup.</p>	<p><b>To Submit:</b></p> <ol style="list-style-type: none"> <li>1. Tender Processing charges not Submitted. Clarify and submit the proof of submission of tender processing charges, if already submitted.</li> <li>2. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Principal Manufacturer as well as by bidder as per proforma duly notarized.</li> <li>3. Copies of Audited Balance sheet &amp; profit loss account for last three financial years certified by Chartered Accountant of the Manufacturer for 2020-21, 2021-22 &amp; 2022-23 with UDIN along with turnover. (Submitted Balance sheets are without UDIN and also the figures mismatch with turnover submitted).</li> <li>4. Latest GST Returns of the Bidder</li> <li>5. Name, photograph &amp; specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.</li> <li>6. Specify point of supply with full Address.</li> <li>7. Declaration of bidder regarding acceptance Bid for terms &amp; conditions (Annexure-A1).</li> </ol>	<p><b>Tender processing charges not submitted. Hence not considered</b></p>	<p>Not Recommended to TEC.</p>
-----	---	---	---	--------------------------------

		<p>8. Copy of CDSCO License (MD-5) for the quoted product and model.</p> <p>9. CE certificate for the quoted product as the submitted certificate expired on 13.04.2025.</p> <p>10. ISO 13485 certificate for the quoted product as the submitted certificate is expired.</p>		
2/2	<p>M/s Mark Enterprises, Maharashtra. (Bidder/Manufacturer/MSME)</p> <p>Model No. MACB-06.</p>	<p><b>To Submit:</b></p> <p>1. Tender Processing charges not Submitted. Clarify and submit the proof of submission of tender processing charges, if already submitted.</p> <p>2. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs100 furnished by the Principal Manufacturer as well as by bidder as per proforma duly notarized. (Submit on non-judicial stamp paper)</p> <p>3. Client Base on Letter Head of the Bidder / Manufacture/Indian Subsidiary of Principal Manufacturer with references of the supply orders, for any of the three financial years in last five financial years along with satisfactory performance certificate of minimum one Purchase Order and installation report of Govt. Institution (Copies of reference supply orders and satisfactory performance certificate need to be attached) (Submitted Purchase orders are not for quoted product and model)</p> <p>4. Nature of the Firm/Public Company/ Private Company/ Partnership/</p>	<p>Not Submitted as the submitted document could not be traced and was also not mentioned in jktender.</p> <p>Submitted Declaration is not as per Tendered Annexure format.</p> <p>Not submitted.</p> <p>Submitted.</p>	<p>Not Recommended to TEC.</p>



	Proprietorship/any other with Documentary proof.		
	5. Copy of GST Registration of the Bidder	Submitted.	
	6. Latest GST Returns of the Bidder	Submitted.	
	7. Copy of the PAN Card of the Bidder	Submitted.	
	8. CE certificate for the quoted product and model.	Submitted.	
	9. Name, photograph & specimen signature of the designated officer/representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.	Submitted.	
	10. Specify point of supply with full Address.	Submitted.	

**Decision:** Accordingly, the Technical Advisory /Evaluation Committee accepted the report of the Sub Committee and recommended Power Point Presentation/Demonstration of the equipments before the Technical Experts. It was further recommended that the quality certifications are to be checked and verified at the time of demonstration and the authenticity and applicability of quality certification shall be decided by the technical experts strictly as per the terms & conditions envisaged in the technical specifications of the NIT.

Regarding the items which are not recommended by the experts, the committee decided to retender the said items.

The following officers attended the meeting:

S. No.	Name of the Officer	Designation	Signatures
1.	Mohd. Ashraf Choudhary.	General Manager (Adm)	
2.	Dr. Ummer Yaseen	A.P Dermatology Technical expert	Ummer Yaseen


(Mohd. Ashraf Choudhary ) JKAS  
General Manager (Adm.) JKMSCL.

Dated: 10/02/2025

No.: JKMSCL/GM/2025/

Copy for information to the:-

1. Principal, Govt. Medical College, Srinagar (Kashmir).
2. Principal, Government Medical College, Jammu.
3. Principal, Government Dental College, Srinagar.
4. Principal, Government Dental College, Jammu.
5. Director General, Indian System of Medicines, J&K.
6. Managing Director, Jammu and Kashmir Medical Supplies Corporation Limited, Jammu
7. Director, Health Services, Jammu.
8. Director, Health Services, Kashmir.
9. Controller, Drug & Food Control Organization, J&K.
10. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited