JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House "Bemina(190018), Srinagar

email: mdjkmscl2@gmail.com, website: www.jkmsclbussiness.com

M/s Hans Raj and Sons Agencies Jammu (Bidder)

No: JKMSCL/GM(Adm.)/2025/8166-79

Dated: 05 -12-2025.

Sub: Finalization and issuance of Rate Contract for the item "Mid Range CR System" under procurement of "Machinery & Equipments".

Ref. No.

1 NIT No. JKMSCL/Mach/2023/568 dated 14-01-2023.

- Minutes of the Purchase Committee meeting issued vide No. JKMSCL/PS/MD/2025-26/492-510 dated 30.08.2025.
- Letter of Intent issued vide No. JKMSCL/GM(Adm)/2025/5468-73 dated 30.08.2025.
- Agreement executed on 30th October 2025.

Sir,

IKMSCL

In line with the decision taken by the Purchase Committee of JKMSCL in its meeting held on 25.08.2025 and notified vide its minutes referred to above, the following item/items as detailed below has/have been approved in your favour on the quoted rates in the offered

documents submitted by you in response to the above referred to NIT:

S. No.	Name of the item	Currency	Basic Rate without Tax with Conversion (1)	Taxes GST/ other applicable taxes (2)	Total Amount (in Rs) with taxes (1+2)
1	Mid Range CR System Manufacturer: M/s Konica Minolta Inc. Tokyo Japan Importer: M/s Konica Minolta Health care India Pvt Ltd Mumbai Model: Regius Sigma II, Dry Pro Sigma II	INR	625000.00	75000.00	7,00,000.00 (Rupees seven lakhs only)

The approved rates for Comprehensive Maintenance Contract after the expiry of warranty

period of 05 (Five) years are as follows:

S. No	Name of item	CMC for five years with taxes (3)
1	Mid Range CR System	6th yr: 1,23,200.00 7th yr: 1,29,360.00 8th yr: 1,35,828.00 9th yr: 1,42,620.00 10 yr: 1,49,751.00 Total: 6,80,759.00

Terms and Conditions:

Delivery Site: Rates are FOR Drug Warehouse of JKMSCL/any Government Institution of J&K State,

wherever installation is required.

2. Validity: The rate contract shall remain valid for a period of two years from the date of issuance of the

Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.

Taxes: As mentioned above and as applicable.
 Warranty: Five Years from the date of installation.

5. Performance

Security You are further required to deposit the performance security equivalent to 5% of the total

invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven

days from the date of issuance of Purchase Order(s) from time to time.

6. Supplies - Within 90 days from the date of issuance of Supply order.

7. Penalty In case of extension in the delivery period with liquidated damages, recovery

of liquidated damages shall be made at the rate of 0.25% per day for every day of delay

"subject to maximum of 10%".

 Installation: Installation / commissioning and necessary training for upkeep of the equipment to be provided free of cost by the firm.

- Availability of Spares: Firm shall certify that the parts of the equipment shall remain available at least for a period of five years after the year of expiry of warranty period of five years.
- 10. The firm has to certify that the rates mentioned above as quoted are the lowest rates applicable to all Govt/private Institution/PSUs etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.
- 11. The supplier/manufacturer/importer shall ensure to supply the item(s) Brand new, unused and of latest manufacturing batch strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/importer/contractor.
- 12. JKMSCL is at liberty to get the feedback from the end-user department/intending department regarding the functioning/Performance of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms and conditions of the NIT/guidelines issued by the Govt, from time to time.
- 13. Fall Clause: The prices under contract shall be subject to price fall clause. The prices charged for the goods supplied under the contract by successful bidder shall in no event exceed the lowest price at which the successful bidder sells the goods or offers to sell the goods of identical description to any other person(s)/Organization(s) including procurement agencies or any other department of Central Govt. or State Govt. as the case may be during the period till the performance of the contract. If at any time, during the period of the contract, the bidder reduces the sales price chargeable under the contract, the same shall be forth with notified and made difference to the JKMSCL, J&K and the price payable under the contract for the stores supplied after the date of coming into force of such reduction or sale shall stand reduced correspondingly. It implies that if the contract holder quotes/ reduces its price "to render similar goods at a price lower" than the contract price to anyone in the Central/State Govt. at any time during the currency of contract including extension period, the contract price shall be automatically reduced with effect from the date of reducing or quoting lower price for all delivery of subject matter of procurement under contract and the contract shall be amended accordingly.
- 14. Risk Purchase: If the bidder is unable to complete the supply within the specified or extended period, the corporation shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the bidder on his (i.e., bidders) account at his cost and risk, with the prior approval of Managing Director JKMSCL, J&K. The bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the bidder.

The bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the bidder under this or any other contract with the corporation/government. If recovery is not possible from the bill and the bidder fails to pay the loss or damage within one month of the demand, the recovery of such amount or sum due from the bidder shall be made from the bidder. In case supplier fails to deliver ordered goods, the risk purchases may be made at market rate from any other firm. It is mandatory for the approved supplier to acknowledge receipt of orders within seven days from the date of dispatch of order, failing which the procuring entity will be at liberty to initiate action to purchase the items on risk purchase provision at the expiry of the prescribed supply period. All additional expenditure incurred shall be recoverable from the bidder.

- Orders shall only be placed on the basis of actual requirement and subject to the availability of funds under the relevant Head of Accounts.
- 16. Payment shall be released after successful installation and commissioning of the equipment duly verified by the intending department.
- 17. Approved technical specifications shall remain same as per NIT (Annexure I)

18. All the terms and conditions of the NIT, Standard Procurement Procedure (SPP) of JKMSCL and LOI shall constitute an Integral part of this Rate Contract.

(M A Choudhary) JKAS General Manager (Adm.)

Yours sincerely,

J&K Medical Supplies Corporation Limited

Copy for information to the:

- 1. Secretary to the Government, Health & Medical Education Department, J&K.
- Mission Director, National Health Mission, J&K.
- Principal, Govt. Medical College, Srinagar /Jammu/ Anantnag/ Rajouri/ Kathua/ Doda/ Handwara/ Udhampur/Baramulla.
- 4. Principal, Government Dental College, Srinagar/Jammu.
- 5. Director Finance, Health & Medical Education Department, J&K
- 6. Director, Health Services, Jammu/Kashmir.
- Controller, Drug & Food Control Organization, J&K.
- 8. Director, Family Welfare, MCH & Immunization, J&K.
- 9. Financial Advisor/CAO, J&K Medical Supplies Corporation Limited.
- 10. General Manager (K), J&K Medical Supplies Corporation Limited.
- 11. Dy. General Manager (P&S)/(Tendering), J&K Medical Supplies Corporation Limited
- 12. Assistant Programmer (M) with the directions to upload the Rate Contract on official website of JKMSCL.
- Medical Officer (I.T), JKMSCL with the directions to upload the Rate Contract on DVDMS Portal.
- 14. File

Annexure to Rate Contract No.	Dated	
Annexure to Nate Continues 1vo.		

MID Range CR Specifications (CR SYSTEM)

Specifications of Computerized Radiology (CR) System for the department of Radio diagnosis.

Image Recording System (Cassettes & Image Plates) Cassettes with Image Plate

- a) 14" x 17" 1 No.
- b) 10" x 12" 1 No.

Image Reader (CR Reader)/Digitizer:

- Digitizer/ Image Reader suitable for Hospitals which is capable to handle high workload is required, having the following features
- Processing Capacity 60 Plates/hour for 14" x 17" size
- Minimum image preview time should be less than 50 sec.
- Depth acquisition resolution should be of 12 bits.
- Unit offered should have resolution of 100 microns or less.
- Should have provision for highest resolution scanning for any size.
- Should have various post- processing features like window level adjustment, Rotation, Zooming, Annotation, Noise reduction, cropping etc.
- · It should be DICOM compatible.
- It should have be able to process cassettes of size 8 x 10", 10" x 12", 11" x 14", 14" x 14", 14" x 17.
- It should be able to pan & zoom the image directly when required, to check the quality of the image.

Dedicated Advanced Work Station

The CR System should have advanced workstation provided with 17" high-resolution monitor keyboard & mouse, with the following features

- It should be capable of processing raw image data of the CR Reader/Digitizer.
- It should have post processing facilities like window level adjustments, Image flipping, rotating, image zooming, edge enhancement, image noise reduction, grey Scale adjustment, Soft tissue adjustment, Bone density adjustment, stepwise image rotation etc.
- It should also have Facility for Multi image printing on Single Film.
- Should have facility of customizable presets.
- The workstation should also have an option to connect to additional viewing workstation for doctor's room.

Dry Laser Camera/Imager:

A dry printer of latest technology should be provided with the system, with the following features:

- Print technology should be dry/Laser.
- Printing resolution should be not less than 500 DPI ± 100 for standard resolution printing.
- Processing capacity: It should be able to print at least 70 ± 10 and above sheets/hour of 14" x 17" size
- Pixels depth architecture/gray scale resolution should be 14 bits of more.
- It should have compatibility of networking & connectivity -there should be provision of direct connectivity to any DICOM modality & on installation there should be available essential provision of connection to at least 4 DICOM modalities simultaneously, for high volume centralized US, CT, MRI, NM, CR & DR, C- ARM printing applications.
- The systems should be capable to connect/link non-DICOM modalities also by connecting it to DICOM.
- It should be able to support film sizes 14 x 17, 14 x 14, 11 x 14, 10 x 12 & 8 x 10.
- It should have two or more online tray or more, capable of changing any film size at any given point of time.
 Should support all films sizes such as 14 x 17, 11 x 14, 10 x 12 & 8 x 10. The changing of film size should be such that only film packet needs to be changed and not the whole tray to avoid wears and tear. In case tray needs to be changed then the company has to supply at least 2/3 tray printer to print multiple film sizes.
- The system should be supplied with suitable UPS with a backup of 15 mins.
- The system should be CE/FDA approved.