JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)
Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu
Corporate Office: Kashmir, Near Haj House, Bemina 190018, Srinagar
email: jkmscljepm@gmail.com; website: www.jkmsclbuisness.com

CORRIGENDUM

In light of the decisions taken by the Technical Expert(s) with reference to the pre-bid meeting for the representation(s) submitted by the prospective bidder(s), the amendments/incorporations have been made in the tender document for the procurement of Adult Flexible Fibre Optic Video Bronchoscope under Group "Machinery & Equipments" uploaded on GeM portal. The amendments in the technical specifications as recommended by the technical experts are appended herewith as Annexure I.

The Critical dates are as under:

- 1. Last date and time for submission of online bids: 08.04.2022 upto 1100 hrs
- 2. Date and time for online opening of technical bids: 08.04.2022 at 1600 hrs Rest of the specifications & conditions of bid other than mentioned in Annexure I shall remain unchanged.

Please Note:

JKMSCL.

- 1. Those firms/bidders who have already uploaded their bids are required to re-upload their bids as per amendments and corrigendum issued thereof.
- 2. All the bidders are requested to keep themselves updated & submit their e-portal bids through as per specifications BOOs. The amendments/modifications shall be available on e.portal www.jktenders.gov.in and www.jkmsclbusiness.com.

Encl: 01 lv

General Manager (Adm)

Dated: 01.04.2022

No.: JKMSCL/Corg/2022/5377-79

Copy for information to the:-

1. Managing Director, JKMSCL.

2. General Manager-K (P&S, I.T), JKMSCL.

3. I/C website to upload the amendments on www.jkmsclbusiness.com and GeM portal.

"Annexure"

S.No.	Existing Technical Specification	Amendment Approved	Remarks	
VIDEO BRONCHOSCOPE				
1.	three or more no. of remote control	Specification retained		
	switches on control body			
2.	Distal end outer diameter: 4-6mm	Distal end outer diameter :	meter : 4- 6.2	
	i i	mm		
VIDEO PROCESSOR MODULE				
3.	must have high resolution HDTV,	must have high resolution HDTV,		
	Should be compatible with optical Should be compatible with optical			
	image enhancement technology -	Image enhancement technology - NBI/BLI/I Scan OE/FICE		
	NBI/BLI/I Scan OE			
4.	should have output : HD-SDI; rgb , dv,	Specification retained		
	dvi, svideo and composite.			
5.	Medical grade monitor-at least 19	19 Inches to be removed.		
	inches (01 No)	1		
LED/300 WATT XENON LIGHT SOURCE				
6.	powerful LED/300 Watt Xenon lamp	Powerful 300 Watt Xenon lamp with		
	LED backup lamp of halogen (at	backup lamp of halogen (at	: least 35	
least 35 watt) watt)				
HIGH RESOLUTION MONITOR				
7.	26" High resolution LCD medical grade	26" or more High resolution	LCD	
	colormonitor . HD Compatible	medical grade color monito	r. HD	
		Compatible		
8.	Compatible trolley, software for	Compatible trolley, Comput	er,	
	report generation.	printer and		
		HD software for report gen	eration.	

Note:-Working-demonstration of Bronchoscope should be given in the Hospital.

Dr. Sanjay Kumar Kalsotra HOD & Professor Department of Anaesthesia GMC, Kathua. Dr. Manik Mahajan Assistant Professor & I/C Department of CD & T.B GMC, Kathua.