



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Opposite J&K Motor Garage Deptt near Hajj House Bemina Srinagar

Telephone: 0191-2478842; 191-3510489 (Jammu), 0194-2490662 (Srinagar)

email: mdjkmscl2@gmail.com; ismjkmscl2018@gmail.com website: www.jkmsclbusiness.com

CORRIGENDUM

As recommended by the Technical experts from end user Department, following amendments are hereby made in the tendered specifications, of items floated under e-NIT No: JKMSCL/ 2025/661 dated 03-05-2025 for procurement of "Dressing Material and Disposable Items",

Item Code	Name of item as per NIT	Amendments made
DRM-032	Adhesive Paper Tape: Porous, Non-Woven, Viscose Rayon Backing with Acrylate Hypoallergenic Adhesive. Should Have Long Self Life With Good Adhesive Capability. Size 1.25 Cm X 9 Mtrs.	Adhesive Paper Tape: Porous, Non-Woven, Viscose Rayon Backing with Acrylate Hypoallergenic Adhesive. Should Have Long Self Life With Good Adhesive Capability. Size 1.25 Cm X 9 Mtrs. Surgical Tape should be safe for use on skin with no skin irritation and it should be a Hypoallergenic tape with low dermatitis potential. Surgical Tape should be nonskin irritation and nonskin sensitizing alongwith being non cytotoxic Test report as per ISO 10993, Bio Compatibility testing for Medical devices as well as Humen Repeat Insult Patch Test (HRIPT) and Human Cumulative irritation patch test (HCIPT), issued by a Registered MDT Laboratory are required to be produced with the supplies.
DRM-033	Adhesive Paper Tape : Porous, Non-Woven, Viscose Rayon Backing With Acrylate Hypoallergenic Adhesive. Should Have Long Self Life With Good Adhesive Capability. Size 2.5 Cm X 9 Mtrs.	Adhesive Paper Tape : Porous, Non-Woven, Viscose Rayon Backing With Acrylate Hypoallergenic Adhesive. Should Have Long Self Life With Good Adhesive Capability. Size 2.5 Cm X 9 Mtrs. Surgical Tape should be safe for use on skin with no skin irritation and it should be a Hypoallergenic tape with low dermatitis potential. Surgical Tape should be non skin irritation and nonskin sensitizing alongwith being non cytotoxic Test report as per ISO 10993, Bio Compatibility testing for Medical devices as well as Humen Repeat Insult Patch Test (HRIPT) and Human Cumulative irritation patch test (HCIPT), issued by a Registered MDT Laboratory are required to be produced with the supplies.
DRM-034	Adhesive Paper Tape : Porous, Non-Woven, Viscose Rayon Backing With Acrylate Hypoallergenic Adhesive. Should Have Long Self Life With Good Adhesive Capability. Size 5 Cm X 9 Mtrs.	Adhesive Paper Tape: Porous, Non-Woven, Viscose Rayon Backing with Acrylate Hypoallergenic Adhesive. Should Have Long Self Life with Good Adhesive Capability. Size 5 Cm X 9 Mtrs. Surgical Tape should be safe for use on skin with no skin irritation and it should be a Hypoallergenic tape with low dermatitis

		<p>potential. Surgical Tape should be nonskin irritation and nonskin sensitizing alongwith being non cytotoxic.</p> <p>Test report as per ISO 10993, Bio Compatibility testing for Medical devices as well as Humen Repeat Insult Patch Test (HRIPT) and Human Cumulative irritation patch test (HCIPT), issued by a Registered MDT Laboratory, are required to be produced with the supplies.</p>
DRM-035	<p>Adhesive Paper Tape: Porous, Non-Woven, Viscose Rayon Backing with Acrylate Hypoallergenic Adhesive. Should Have Long Self Life with Good Adhesive Capability. Size 7.5 Cm X 9 Mtrs.</p>	<p>Adhesive Paper Tape : Porous, Non-Woven, Viscose Rayon Backing with Acrylate Hypoallergenic Adhesive. Should Have Long Self Life with Good Adhesive Capability. Size 7.5 Cm X 9 Mtrs.</p> <p>Surgical Tape should be safe for use on skin with no skin irritation and it should be a Hypoallergenic tape with low dermatitis potential. Surgical Tape should be nonskin irritation and nonskin sensitizing along with being non cytotoxic</p> <p>Test report as per ISO 10993, Bio Compatibility testing for Medical devices as well as Humen Repeat Insult Patch Test (HRIPT) and Human Cumulative irritation patch test (HCIPT), issued by a Registered MDT Laboratory are required to be produced with the supplies.</p>

- Financial Bids of only those bidders shall be considered for opening after being recommended by the Technical Experts, from the concerned Department, out of the bidders recommended by the subcommittee after evaluation of Technical Bids and acceptance of Technical Evaluation/Advisory Committee.
- In case of any default by the bidder, at any stage of tender or subsequent approval by JKMSCL, for a particular items/s, the Disciplinary Committee/ any other committee constituted for the purpose shall be at liberty to take appropriate action as per provisions of Standard Procurement Procedures (SPP) and / or Policy for Blacklisting of JKMSCL.

In view of above, scheduled dates for receipt and opening of Bids are also extended as under:

1	Last Date for submission of Technical Bids	26-06-2025 upto 4:00 p.m.
2.	Date of Opening of Technical Bids	28-06-2025 at 11:00 a.m.

Only technical/financial bids uploaded as per revised corrigendum shall be considered for evaluation.

All the bidders are requested to keep themselves updated and submit their e- bids through e-portal www.jktenders.gov.in as per specifications & BOQs. The amendments/ modifications shall be available on www.jktenders.gov.in, and www.jkmsclbusiness.com.

Rest all shall remain unchanged.

General Manager (K)
JKMSCL

Dated: 14-06-2025

No. JKMSCL/GM/Drugs/Corg/661/2025/1702-04

Copy for information to the:-

1. Managing Director, JKMSCL
2. Assistant programmer to uploaded on www.jktenders.gov.in and on www.jkmsclbusiness.com
3. Office Record.