



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Jammu Office: Plot No. 9, Transport Nagar, Narwal (J&K)-180006; Tele: 0191-2490001; Telefax: 0191-2490902

Srinagar Office: 121-Green Avenue, Hyderpora (J&K)-190014; Telefax: 0194-2432008

email: enquiryjkmscl@gmail.com; website: www.jkmscl.nic.in

Subject: Corrigendum- E-bid for the procurement of “Surgical Instruments & Implants” uploaded vide No. JKMSCL/INSTRUMENTS /2015/111 dated 02.12.2015.

In continuation to e-bid uploaded for the finalization of Annual Rate Contract for the procurement of Surgical Instruments & Implants vide No. JKMSCL/ Instruments/ 2015/111 dated 02.12.2015 & pre-bid meeting held thereof on 17.12.2015, following amendments in light of recommendations/suggestions made by the various vendors/firms/bidders are hereby made as under :-

S. No.	Item No.	Name of the item	Specifications in tender document	Amendments : Please read the specification column (4) amended as
	1	2	3	4
1	GYNI 049	Delivery Set	Delivery Set (High quality SS, corrosion resistant. Long durability, reusable	Delivery Set (High quality SS, corrosion resistant. Long durability, reusable containing : Scissors (Cord cutting 01, Epsitomy -01, Mayo 17.5 cm -01) Artery forceps 15 cm- 02 Spong holding forceps 01 Speculum -01 Bowl for antiseptic lotion 01, Kidney tray -01 Cover bag -01
	GYNI 51	Plastic apron	Plastic apron	To be read as “Deleted”
	GSII 70	Laparotomy Set	Laparotomy Set	To be read as “Deleted”
	AOTIO 24	Tracheostomy tube plain all sizes	Tracheostomy tube plain all sizes	To be read as “Deleted”
	AOTIO 25	Tracheostomy tube Cuff all sizes	Tracheostomy tube Cuff all sizes	To be read as “Deleted”
	AOTIO 37	Laryngeal Mask Airway	Laryngeal Mask Airway	To be read as “Deleted”
	AOTIO 47	Pediatric Spinal	Pediatric Spinal Needle nos	To be read as “Deleted”

		Needle nos 25 disposable	25 disposable	
	AOTIO 48	Pediatric Spinal Needle nos 27 disposable	Pediatric Spinal Needle nos 27 disposable	To be read as “Deleted”
	Terms & Conditions			
		<p>Mandatory Quality parameters for surgical instruments (Annexure AVIII)</p> <ul style="list-style-type: none"> • All scissors, needle holders and dissecting forceps (toothed/non toothed) should be made of Tungston carbide. The tungsten carbide inlay should be nickel based. The Tungston Carbide should be welded to the instrument and not pasted. All Other instruments should be made of high grade stainless steel meeting National/International Standards. • It should be of high quality and precision. • It should be light weight, strong and durable. • It should be non-magnetic. • All instruments should be autoclavable at 134 °C. • All Instruments should meet the National/International 	<p>Please Read as :</p> <p>Mandatory Quality parameters for surgical instruments (Annexure AVIII)</p> <ul style="list-style-type: none"> • All scissors, needle holders and dissecting forceps (toothed/non toothed) should be made of Tungston carbide. The tungsten carbide inlay should be nickel based. The Tungston Carbide should be welded to the instrument and not pasted. All Other instruments should be made of high grade stainless steel meeting National/International standards. The non corrosion resistance should be further enhanced by a series of processes such as hardening, tempering and polishing. • It should be of high quality and precision. • It should be light weight, strong and durable. • It should be non-magnetic. • All instruments should be autoclavable at 134 °C. • All Instruments should meet the National/International 	

		<p>Certification like CE/USFDA/ DIN/ISO or any other equivalent standardization.</p> <ul style="list-style-type: none"> • It Should be highly heat resistance/non corrosive. • The catalogues/literature of the instruments, if any shall be submitted in original to the office of JKMSCL against proper receipt. • The manufacturer/firm shall certify that all instruments are brand new. • The manufacturer should have a direct repair facility available in India. 	<p>Certification like CE/USFDA/DIN/ ISO or any other equivalent standardization.</p> <ul style="list-style-type: none"> • It Should be highly heat resistance/non corrosive. • The catalogues/literature of the instruments, if any shall be submitted in original to the office of JKMSCL against proper receipt. • The manufacturer/firm shall certify that all instruments are brand new. • The manufacturer should have a direct repair facility available in India.
		<p>Clause 6 (General Conditions of the contract) : All instruments supplied should bear marking "JKMSCL SUPPLY 2015-16, NOT FOR SALE, " or as mentioned in supply order in English, without which the supply shall not be entertained.</p>	<p>Please read as : All instruments supplied should bear marking & also be engraved as "JKMSCL SUPPLY 2015-16, Country of origin/ manufacturing, item code" or as mentioned in supply order, without which the supply shall not be entertained.</p>

Sd/-

Dr. Rajinder Kr. Khajuria, KAS
General Manager (Adm)

No.: JKMSCL/Instruments/Corg/2015/336

Dated:30.12.2015

Copy for information to the:-

1. Managing Director, JKMSCL, Jammu.
2. Financial Advisor / Chief Accounts Officer, JKMSCL, Jammu.