

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of Govt of Jammu & Kashmir)

Corporate Head Office: 1ST Floor, Drug Store Building, Govt. Medical College, Bakshi Nagar, Jammu

Corporate Office: 121, Green Avenue, Hyderpora, Opp. Al-Farooq Masjid, Srinagar Telephone: 0191-2580842, Fax: 0191-2581845 (Jammu); Telefax: 0194-2432008 (Srinagar)

Subject: Minutes of the Meeting of Technical Advisory/Evaluation Committee of Power Point Presentation/Demonstration & Evaluation for Dental Machinery Equipment held on 07th and 08th September 2018 starting from 11.00 AM at Government Dental College Srinagar with reference to NIT No.NIT/JKMSCL/DEN/2018-19/292 Dated 23.04.2018 for Finalization of Rate Contract for supply of Dental Machinery & Equipment.

A Power Point Presentation/Physical Demonstration of Sample based and Demonstration of Catalogue based items of below mentioned firms/manufacturers was given on07th and 08th September 2018 startingfrom 11.00 AM

S.No.	Name of the firms.	**	
1	M/s Confident Dental Equipment's.		
2	M/s Suz Dent India Pvt Ltd.	, i	
3	M/s V.A Technologies.	•	

Decision:

After detailed deliberations with regard to various queries and after Evaluation of Sample based and Catalogue based items held on 07th and 08th September 2018 starting from 11.00 A.M, Technical experts are of the opinion that

Hayleman 8/9/18

leave my months

12 rain Dans

American Constitution of the Constitution of t

The oblinou man

A/

following firms/bidders fulfilling all the requisite eligible conditions and specifications as per tender document who are declared Qualified, their Financial Bid recommended for opening. :

S.no	Item code	Name of the item	Name of the firm	Remarks.
1.	DE0001	LED Light Cure (Catalogue based)	M/s Confident Dental Equipment's	Disqualified It's not as per Specifications of the NIT because intensity not 1600 Mw/cm2
2.	DE0002	OPG MACHINE (Catalogue Based)	M/s Confident Dental Equipment's	Qualified But printer should be provided which should be compatible to all X-ray films (of any make) not that printer which will be compatible with planmeca films only.
			M/s V.A Technologies	Disqualified (it is not as per the specifications of the NIT)
3.	DE0003	Dental X-Ray Unit (Catalogue based)	M/s Confident Dental Equipment's	Disqualified (it is not as per the specifications of the NIT)
4.	DE0004	RVG with Laptop (Digital Radiography machine) (Catalogue based)	M/s Confident Dental Equipment's	Not as pr the specifications of the NIT as CMOS sensor not mentioned on Catalogue
5.	DE0005	Glass Bead sterilizer (Catalogue based)	M/s Confident Dental Equipment's	Qualified It is as per specifications of th NIT

D4/

Ortife 3 tears

Hayley 8/8/18

Sersi Males outream

Janai Don 8/9/18

only for the mose

2008 2016 By 2018

ON Alad.

6.	DE0006	Suction apparatus (Sample based)	M/s Confident Dental Equipment's	The sample was not provided.
7.	DE0007	Air compressor : (Sample based)	M/s Confident Dental Equipment's	Qualified It is as per specifications of the NIT
			M/s Suz Dent India PvtLtd	Disqualified The Air compressor has pasted ordinary stickers on the compressor and even European CE is pasted as a ordinary sticker but they are not as embedded stickers. So the quality and the specifications of the compressor cannot be ascertained and plastic make of compressor is of low grade.
8.	DE0008	Piezo electric ultrasonic scalar with minimum of 3 tips. (Catalogue based)	M/s Confident Dental Equipment's	Qualified Specifications on catalogue as per mentioned on NIT against item no. DE0008.
9.	DE0009	Apex locator (Catalogue based)	M/s Confident Dental Equipment's	Qualified It is as per specifications of the NIT
10	DE0010	Endo motor within built apex locator programmed setting for rotary endo dentist files, digital display, attached hand piece, auto stop / reverse / forward rotary mode.	M/s Confident Dental Equipment's	Disqualified Catalogue not provided.

beeg soll

GR100 819/18

who have been of a state of the state of the

ויין די סוכט

		(Catalogue based)		
11	DE0011	Dental chair Fully Loaded	15/ 5 5	
		(Sample Based)	M/s Confident Dental	Qualified.
		(Lasea)	Equipment's	As per specifications
			M/s Suz Dent India PvtLtd	Disqualified.
				Micromotor power chord gets
				detached from the motor on
				slight digital pressure (it should
	Year			be tightly screwable). There is
				poor insulation on the chord
				which shows its poor quality.
				Digital display of the micro
				motor control fitted in the chair
				is of poor quality. The overall make and finish of dental chair
	"555040		CALLER CALLERY	seems to be of poor quality. Oualified.
12.	DE0012	Dental chair (Basic) Sample Ba		As per specifications
			Equipment's	Disqualified.
	1		M/s Suz Dent India PvtLtd	Micromotor power chord get
				detached from the motor of
				slight digital pressure (it should
			*	be tightly screwable). There is
				poor insulation on the chor
				which shows its poor quality
		18		Digital display of the micr
				motor control fitted in the cha
				is of poor quality. The overa
	ľ			make and finish of dental cha
				seems to be of poor quality.
			M/s Confident Dental	Qualified
	= 2010	Dental chair (Pedo) (Catalogue	Wils Confident Bental	Many Many

8/9/18

SETUP SINIT

g2, win 8/9/18

MD 1012 2019 2 2

Alah

			to be the
	based).	Equipment's	Accessories provided as per the specifications of the NIT.
No.		M/s Suz Dent India PvtLtd	Disqualified Has provided the Catalogue of
4.	-		basic dental chair)

The Power Point Presentation was made in the presence of following members:

S.No.	Name & Designation of Officer	Signature
1	Dr. Mohammad Iqbal General Manager (P&S, IT/Adm)- JKMSCL. (Chairman EPM)	Brone'
2	Dr. Altal Chalkoo Professor (Oral Medicine and Radiology) GDC Srinagar	only the items
3	Dr. Aijaz Shah Professor (Oral and Maxillo-facial Surgery) GDC Srinagar	Marin 10 19 1 1 1 20 19
4	Dr. Sandeep Kour Professor (Prosthodontics) GDC Srinagar	Reall DE GOLL
5	Dr. Suhail Majid Jan Professor (Periodontia) GDC Srinagar.	(De door de control d

6	Dr. Nazia Lone	The glatte
	Assistant Professor (Pedodontia) GDC Srinagar	Y
7	Dr. Ajay Kumar Gupta . Associate Professor (Conservative and Endodontics) IGGDC Jammu.	Ajor leman
8	Dr. Aamir Rashid Associate Professor (Conservative and Endodontics) GDC Srinagar	Barg 1948.
9	Dr. Syed Zamir Khurshaid Associate Professor (Orthodontia) GDC Srinagar	00 b 00007
10 md Max	Dr. Roopav Nargotra Assistant Professor (Oral and Maxillofacial Surgery) IGGDC Jammu.	& 8/1/18
11.	Dr. Satvinder Singh Assistant Professor (Oral Medicine and Radiology) IGGDC Jammu.	85ing 18
. 12.	Dr. Sanjeet Singh Risam (Consultant Dentistry) DHS Kashmir	grain 8/9/18.
13.	Mr. Gafas-din A.E.E MECH.	

Don

Otessor (Mrai:

14.	Dr. Qazi Qamar	Q V
14.	MO-JKMSCL	Okg - 10/19
15.	Dr. Aijaz Naik	In In
	(Dental Surgeon) DHS Kashmir.	71.
16.	Mr. Javed Ahmed Wani	1
	AD (QC)-JKMSCL	08/19/18
17.	Mr. Pawan Sharma	\a
	AAO-JKMSCL	
18		The state of the s
		<u></u>

Dr. Mohammad Iqbal General Manager (P&S, IT/Adm), JKMSCL Chairman (EPM)

Dated: 12 /09/2018.

No: - JKMSCL/GM/PS/2018/ 2649-52

Copy for the information to:

1. Principal Government Dental College, Srinagar.

2. Principal Government Dental College, Jammu.

3. FA/CAO, JKMSCL

4. Assistant Director (QC), JKMSCL.