

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

Corporate Head Office: Plot No: 58, Friends Colony, Satyam Road, Trikuta Nagar Jammu

Corporate Office: Opp State Motor Garages Near Haj House, Bemina, Srinagar

Telephone: 0191-2580842, Fax: 0191-2581845 (Jammu); Telefax: 0194-2432008 (Srinagar)

OFFER TO MATCH THE RATES

In view of sudden upsurge in COVID cases, Purchase orders for the supply of Oxygen Concentrators 10 LPM capacity stands issued with the approved supplier of JKMSCL. The approved supplier is executing the supplies in phased manner due to shortage of item in the market. In the interest of patient care and the urgency expressed by the end users, the Oxygen Concentrators are required to be installed in various Health Care Institutions on war footing basis. Accordingly, firms, importers/manufacturers dealing in the line are requested to give their counter offers to match the rates approved by JKMSCL, if they are willing to provide minimum 2000 Oxygen Concentrators of 10 LPM capacity in one week.

S. No	Detailed specifications		Approved Rates including taxes	Minimum qty required
1.	Technical Specifications (specific to this type of device)	<ol style="list-style-type: none"> Flow rate: 0~10 LPM, purity > 93%. O₂ delivery pressure: 0.03 to 0.07 MPA (4.35- 10.15 PSI). Atomizing pellet (ml/min.) > 0.5, uninterrupted flow of oxygen. Oxygen monitoring system (optional). Low pressure alarm, high pressure alarm and power failure alarm. Unit capable for supplying oxygen to two outlets simultaneously using two independent flow meters shall be preferred. 	Rs. 1,05,840 PER PC. FOR Stores	2000 Nos. per week
	Settings	Should be capable of providing minimum 12 hours of continuous operation.		
	User's interface	Front panel access to reset switch.		
	Dimensions (metric)	Max spec: 640 mm (H) x 410 mm (W) x 410 mm (D).		
	Weight (lbs, kg)	Max 30 kg.		
	Configuration	NA		
	Noise (in dBA)	<50 db		
	Heat dissipation	Heat desipated using an internal exhaust, so that a maximum of 36.5 degree C is maintained.		
	Mobility, Portability	Yes		
	Power requirements	230 +/- 10% VAC, 50 Hz, 2 amps.		
	Battery operated	NA		
	Tolerance (to variations, shutdowns)	Fuse controlled variation, automatic switch over from AC to DC and vice versa.		
	Protection	OVP, earth leakage protection		
	Power consumption	<500 Watts		
	Accessories (mandatory, standard, optional)	Humidifier Bottles-4nos, power cord-1 no.		

Terms & Conditions:

- The Offers should be submitted through official e-mail id: pbmjkmscl@gmail.com latest by **17-05-2021 upto 11:00 am. (quotation submitted on any other e-mail of JKMSCL shall not be entertained).**
- No quotation shall be entertained after the above mentioned date and time.

- iii. Rates must be submitted FOR RDWH Bemina Srinagar basis.
- iv. Payments shall be released after receipt and verification of supplies and No advance payment shall be made at all.
- v. Catalogue of the product offered must be attached.
- vi. GST Registration of the bidder must be attached.
- vii. Authorization from the original Manufacturer, if the bidder happens to be a authorized dealers of the original manufacturer must be attached.
- viii. Quotations offered by the defaulter suppliers of JKMSCL for the above items shall not be entertained.
- ix. The JKMSCL shall not be bound to accept the lowest offer. In view of Pandemic, the delivery schedule shall be considered for accepting the counter offer. However, if the supplier fails to execute the supplies, he shall be blacklisted for 5 years from participating in JKMSCL.

Issued with approval of Managing Director


General Manager (Adm)
JKMSCL

No: GM/JKMSCL/ADM/2021/264-268

Dated: 14/05/2021

Copy to the:-

1. Financial Commissioner, H&ME, Civil Secretariat, Jammu.
2. MD JKMSCL, FA& CAO, JKMSCL.
3. FA & CAO, JKMSCL
4. I/C Drug Regional Ware House Jammu/ Srinagar.
5. Office copy.