

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: GMC Complex, Bakshi Nagar, Jammu (J&K)-180006, Tele : 0191-2580842
Corporate Office: 121-Green Avenue, Hyderpora (J&K)-190014: Telefax: 0194-2432008
email: jkmscljepm@gmail.com; website: www.jkmscl.nic.in

CORRIGENDUM

SUBJECT: E Bid For the Finalization Of Rate Contract For The Procurement Of "Machinery & Equipments"

In light of the decision taken during the pre bid meeting held on 8th, 9th and 10th May 2018 & recommendation/suggestion made by the technical experts from controller, Drugs, food & control organization, j&k on the basis of representation made by the firms/bidders, the amendments/incorporations have been made for the following items in the tender documents for the finalization of rate contract for the procurement of "Machinery & Equipments was uploaded vide no. NIT/JKMSCL/MACH/2018/288. Dated 17-04-2018.

1. Technical Specification of **Anaesthesia work station**

Sr No	Item Code	Item Description	Specification As NIT	Modification /Addendum /Clarification During Pre Bid Meeting.
04	MC1756	Anaesthesia work station	<p>The machine should have the following accessories: (Read As Deleted)</p> <ol style="list-style-type: none">Reusable Spo2 sensors 10 each for adult 5 for pediatric 2 for neonate.ECG cable 10 No.NIBP cuff 7 each for adult 3 each for pediatric and neonateIBP transducers 10 No.IBP cable 10 in No.ETCO2 sample 100Water trap 100 No.Reusable autoclave breathing circuit 5 each for adult / paed.Temperature probe 5 for adult 2 for pediatric.Should have provision for des fluorine vaporizerDisposable breathing circuit with water trap 100 No. each for adult and PediatricNMT and BIS Entropy cables and electrodesShould have accessories for suction and active AGSS	<p>The machine should have the following accessories: (Read As Deleted)</p> <ul style="list-style-type: none">Magill Breathing System 2Patient Circuit(Circle System) Adult 4 Sets.Silicon & Autoclavable Paediatric 4 SetsReservoir bags (Adult 2 liters & 5 for each Paediatric 1 liter respectively).Brain Breathing Systems 4 No.sPediatric Circuit (Jackson Rees) 4 NO.sFace masks 0-5 4 setsFace Masks 0-000 paed 2 sets5 lead ECG Cables 4 setsSpo2 Cable & Sensor (Adult & Paediatric) 5 No.s .Skin temperature probes 5 No.s.Nasopharyngeal (Adult And Paediatric) 5 No.sETCo2 sample lines 20No.sNIBP Cuff with tubing<ul style="list-style-type: none">✓ Adult 5NO.s✓ Pediatric 5 NO.s✓ Infant 2 No.

Handwritten signature/initials

				<ul style="list-style-type: none"> • Airways pressure and flow measurement lines 10 No.s • Ventilator Bellows (Adult and Paediatrics) 2 For Each. • Ventilator filters /flow Sensors 10No.s
--	--	--	--	--

Please Read Also:

- **Last date and time for Registration:** 20/06//2018 at 1600 hrs
- **Last date and time for submission of online bid** 25/06/2018 at 1600 hrs.
- **Date & time of opening of Technical Bids** 28/06/2018 11.00am

Please Note:

1. Those firms/bidders who have already uploaded his/her bid are required to re-uploaded their bids as per the amended specification and corrigendum issued thereof.
2. Only technical/financial bids uploaded as per revised corrigendum shall be considered for evaluation.

Please Note:

All the bidders are requested to keep themselves updated and submit their e- bids through e-portal www.jktenders.gov.in as per specifications & BOQs. The amendments / modifications shall be available on www.jktenders.gov.in, and www.jkmsclbusiness.com

Inder Jeet(KAS)
Managing Director,
JKMSCL

No.: JKMSCL/PS/MD/Corg/2018/

Dated: .06.2018

Copy for information to the:-

1. Financial Advisor / Chief Accounts Officer, JKMSCL, Jammu/srinagar.
2. General Manager-J (Adm), JKMSCL
3. General Manager-K (P&S, I.T), JKMSCL.
4. Say Technology for information to uploaded on www.jkmsclbusiness.com
5. office Record.

No: JKMSCL/GM(K)/PS/2018/ 1358 - 61

Dated:// .06.2018


Dr. Mohammad Iqbal.
General Manager(P&S,I.T)
JKMSCL

