

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashrnir)
Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu
Corporate Office: Opposite J&K Motor Garage Deptt near Hajj House Bemina Srinagar Telefax: 0194-2432008
email: gmjkmscl.mi@gmail.com, <a href="mailto:gmjkmscl.mi@gm

NOTICE

Sub-Committee Report Tender No: 475 Oxygen Concentrator 5lpm

Subject: Evaluation of Technical bids for finalization of the contract for the procurement of the oxygen concentrator with reference to the NIT No. NIT/JKMSCL/MED/COVID/2020/475 Dated. 24-05-2021.

With reference to the report of the Sub-committee, the evaluation of bid documents for the finalization of the rate contract for the procurement of the oxygen concentrator 5lpm with reference to the NIT No. NIT/JKMSCL/MED/COVID/2020/475 Dated 24-05-2021, the preliminary evaluation of the bid documents have been done on the basis of uploaded documents and following infirmities have been found.

S.No	Name of firm/bidder	Minor Infirmities along with Remarks
1.	M/s Andhra Pradesh Medtech Zone Ltd. Vishakhapatnam. (Bidder)	Tender fee Submitted. No documents other than MoU in the technical bid submitted.
2.	M/s Uplay Impex Pvt. Ltd. New Delhi (Bidder) M/s GSR International. (Importer)	Kindly clarify viz-a-vis sole manufacturer with all requisite documents like Product permission, Import licence, Authorisation as per NIT format, all quality certificates etc. of the concerned manufacturer as per Annexure-A1 of NIT.
	 (Manufacturers) i) M/s Jiangsu Konsung Bio-Medical Science and Technology Company Ltd., China. ii) M/s Carer Medical Technology Co. Ltd. No. 6 Yuping Road, High Tech Zone, Suzhou city, Jiangsu, China 251000. iii) M/s Jiangsu Jumao X Care Medical Equipment Co. Ltd. 	



1

pr

	N. OCB	
	No. 36 Danyan Road 212300, Danyang, Jiangsu, China- 212300	
<u></u>	Model: Not specified	
3.	M/s Malik Sales Agencies. (Bidder)	Only 2-3 required documents submitted.
	M/s GCE India Pvt. Ltd. (Importer/Manufacturer)	
	Model: Not specified	
4.	M/s Vivek Impex. (Bidder)	A letter from principal manufacturer in favour of Bidder that he is the sole Importer of quoted product in India.
	M/s Jiangsu Konsung Bio-Medical Science and Technology Company Ltd. China.	 Bid security declaration to be submitted. Product permission by the licensing authority for the products to be quoted.
	(Manufacturer)	 Import license in case of imported product to be submitted. Latest GST return to be submitted.
	Model: KSOC-5	6. To submit the following quality certificates for the quoted product: a) IEC 60601-1,
		b) IEC 60601-1-2 c) ISO 15001-2010.
5.	M/s Surgi Mart, Chandigarh. (Bidder)	1. Average Annual Turnover Statement not less than 5.00(Five) crores of Original Manufacturer
	M/s BPL Medical Technologies Pvt. Ltd, Bangalore	for 3 years from Chartered Accountant with UDIN on letter head of CA. (Submitted is without UDIN).
	(Manufacturer)	 Copies of audited balance sheet & Profit Loss Account for last three financial years from Chartered Accountant.
	Model: OXY 5 neo	To submit the following quality certificates for the quoted product:
		a) ISO 13485. b) IEC 60601-1,
		c) IEC 60601-1-2
		d) ISO 15001-2010.
6.	M/s Bulk MRO Industrial supply Pvt. Ltd., Mumbai (Bidder)	Firm is alleging major deviations from the NIT floated. Hence cannot be considered.
	M/s Guangdong Owgels Science & Technology Co. Ltd., China (Manufacturer)	

Page 2 of 6

R

4

	Model: OZ-5-01PW0	
7.	M/s Moditech International, Punjab (Bidder/Importer) M/s Longfian Scitech Co. Ltd, China (Manufacturer)	 Product permission by the licensing authority for the products to be quoted. Import license in case of imported product to be submitted.
	Model: JAY-5BW (White/Blue)	
8.	M/s Pinnacle Industries Pvt. Ltd, Pune (Bidder)	The firm had submitted Rs. 9001/- instead of 10,000 as bid/tender charges as MSME but quoted product of other manufacturer. Hence cannot be considered.
	M/s Jiangshu Yuyue Medical Equipment & Supply Co. Ltd, China (Manufacturer)	
	Model: 8P-5A	
9.	M/S V A Technologies, Jammu (Bidder) M/S Grid Engineers Pvt. Ltd, New Delhi (Manufacturer) Model: Not specified	 Product permission by the licensing authority for the products to be quoted. Specify the model of the quoted product. To submit the following quality certificates for the quoted product: IEC 60601-1, IEC 60601-2010.
10.	M/s SPJ Solar Technology Pvt. Ltd, U.P (Bidder) M/S Kunshan Dynmed Medical Technology Co. Ltd, China (Manufacturer) M/s Zhengzhou Olive Electronic Technology Co., Ltd. China. (Manufacturer) M/s Shenzhen Tablet Electronics Ltd., China. (Manufacturer)	Kindly clarify viz-a-vis sole manufacturer with all requisite documents like Product permission, Import licence, Authorisation as per NIT format, all quality certificates etc. of the concerned manufacturer as per Annexure-A1 of NIT.

Page 3 of 6

11.	M/s Kunshan Dynmed Medical Technology Co., Ltd., China. (Manufacturer) Model: Not specified M/s Mann Electronics, Rajasthan, MSME (Bidder/Manufacturer)	Kindly clarify the authorisation whether self or through M/s Nandini Marketing Pvt. Ltd. And submit following documents.
	Model: MN-1024-5TJB	1. To submit the following quality certificates for the quoted product:a) IEC 60601-1,b) IEC 60601-1-2
12.	M/s Medi waves Inc, Ḥaryana (Bidder/Manufacturer) Model: MEDAIR	 To submit the following quality certificates for the quoted product: ISO 13485 IEC 60601-1, IEC 60601-1-2 ISO 15001-2010.
13.	M/s Nandini Marketings Pvt. Ltd, Indore (Bidder) M/S Mann Electronics, Kota Rajasthan (Manufacturer) Model: Not specified	 Bid Security Declaration. Copies of audited balance sheet & profit & loss account for last three years from chartered accountant to be submitted. Latest Non-Conviction certificate issued by the licensing authority of the respective state/ Self certified on letter Head duly sealed and signed. Product permission by the licensing authority for the product to be quoted. Import Licence in case of imported product. To submit the following quality certificates for the quoted product: CE or FDA approval. IEC 60601-1. IEC 60601-1-2
14.	M/s S S Medical System, India Pvt. Ltd. U.P (Bidder/Manufacturer) Model: SS-OC-50	 Average Annual Turnover Statement not less than 5.00(Five) crores of Original Manufacturer for 3 years from Chartered Accountant with UDIN on letter head of CA. (Submitted UDIN is not readable). Copies of audited balance sheet & profit & loss account for year 2019-20 from chartered accountant to be submitted. Product permission by the licensing authority for the products to be quoted. To submit the following quality certificates for the quoted product:

Page 4 of 6

_		
	1 2 200	a) ISO 13485
l.	1	b) IEC 60601-1,
	25	c) IEC 60601-1-2
	1	d) ISO 15001-2010.
		4, 100 1001 2010.
15.	M/s Hemodiaz Lifesciences Pvt.	1. Product permission by the licensing authority for
	Ltd. New Delhi	the products to be quoted, wherever applicable
	(Bidder/Importer)	to be submitted.
	(blader/importer)	
	M/-IBi G i l G I l	2. Import License in case of imported product to be
	M/s Long Fian Scitech Co. Ltd.	submitted.
	China	3. Authorisation from principal manufacturer to be
	(Manufacturer)	submitted as per prescribed format in favour of
		bidder as sole Importer for the quoted product in
	Model: Not specified.	India to be submitted. Clarification required
		regarding the authorisation as previous one is
		without signature and NIT format.
		Specify the model of the quoted product.
		5. To submit the following quality certificates for
	`	the quoted product:
		a) ISO 13485.
		b) IEC 60601-1.
		c) IEC 60601-1-2.
		d) ISO 15001-2010.
16.	M/s Anjana International	1. Product permission by the licensing authority for
	Industries (Bidder/Importer)	the products to be quoted, wherever applicable
		to be submitted.
	M/s Foshan MIC Medicak	2. Import License in case of importer product to be
	Technology Co. Ltd. China	submitted.
	(Manufacturer)	3. Specify the model of the quoted product.
	Model: Net enerified	4. A letter from principal manufacturer in favour
	Model: Not specified.	bidder as sole Importer in India for the quoted
		product.
		To submit the following quality certificates for
	·	the quoted product:
		a) ISO 13485.
		b) IEC 60601-1.
		c) IEC 60601-1-2.
	, •	d) ISO 15001-2010.
17.	M/s Hans Raj Nayar (Mediquip),	Average Annual Turnover Statement not less
-··.	INC, Punjab	
	(Bidder)	J. J
:	(Bluder)	Manufacturer/Importer for 3 years from
	M/- 09 11 32 10 1 5 5 5 5	Chartered Accountant with UDIN on letter head
	M/s Silverline Meditech Pvt. Ltd.	of CA with UDIN.
	Ahmdabad.	2. Product permission by the licensing authority for
	(Importer)	the products to be quoted, wherever applicable
		to be submitted.
	M/s Foshan keyhub electronics	3. Import License in case of importer product to be
	industries co. Ltd. China	submitted.
		Authorisation from principal manufacturer in
	(Manufacturer)	
	(Manufacturer)	favour of Importanta has submitted as asl-
		favour of Importer to be submitted as sole
ļ	(Manufacturer) Model:OXYCURE DURA.	favour of Importer to be submitted as sole importer in India for the quoted product. 7. To submit the following quality certificates for

Page 5 of 6

the quoted product: a) CE or FDA approval. b) ISO 13485. c) IEC 60601-1. d) IEC 60601-1-2. e) ISO 15001-2010.
--

The bidder(s) are requested to upload their infirmities 14.06.2021 upto 11:00 AM positively on website www.ikmsclbusiness.com and in case, the file(s) is/are heavy and the bidder are not in position to get the same uploaded on JKMSCL website, the same may be sent to corporation on the email of tender inviting authority gmikmscl.mi@gmail.com which shall be forwarded to technical evaluation sub-committee for scrutiny and recommendations in the matter.

In case the documents are not submitted in time, your bid shall be evaluated only on the basis of documents submitted earlier.

I/C General Manager (Adm) JKMSCL

Dated:

UNIVECE

/1 /06/2021

No.NIT/JKMSCL/GM(ADM)/2021/787~88

Copy to the:

1. Managing Director, JKMSCL.

2. I/C Website to upload the minor infirmities on the official website of JKMSCL.

3. Office Copy