

## **JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.**


*(Public Sector Undertaking of the Government of Jammu and Kashmir)*

**Corporate Head Office:** Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

**Corporate Office:** Opposite J&K Motor Garage Deptt near Hajj House Bemina Srinagar

**Telephone:** 0191-2478842; 191-3510489 (Jammu), 0194-2490662 (Srinagar)

**email:** mdjkmscl2@gmail.com; ismjkmscl2018@gmail.com **website:** [www.jkmsclbusiness.com](http://www.jkmsclbusiness.com)



### **E-BID FOR THE PROCUREMENT OF MACHINERY & EQUIPMENTS**

**(REFERENCE No: NIT/JKMSCL/M&E/2023/ 596 DATED: 02 /08/2023**

**LAST DATE OF SUBMISSION OF ONLINE BIDS: 07-09-2023 upto 1600 hrs**

**Important Note:** *Each page of e-Bid should be properly page marked and indexed. Page Number should be reflected at the bottom of each page. All documents requested in “Annexure-II”, should be reflected in the column mentioned against each (Page No. \_\_\_\_). Any deviation may result in rejection of the bid and the bidder shall be solely responsible for the same.*

## **BIDDING DOCUMENT FOR PROCUREMENT OF MACHINERY & EQUIPMENTS**

### **Table of Contents**

| <b>S.No.</b> | <b>Section</b> | <b>Description</b>                            | <b>Pages</b> |
|--------------|----------------|---|--------------|
| 1.           | NIL            | Bid Submission Letter                         |              |
| 2.           | NIL            | Notice Inviting Bid for uploading on Websites |              |
| 3.           | I              | Instructions to Bidders                       |              |
| 4.           | II             | Bid Data Sheet                                |              |
| 5.           | III            | Evaluation and Qualification Criteria         |              |
| 6.           | IV             | Bidding Forms (BF)                            |              |
| 7.           | V              | Schedule of Supply                            |              |
| 8.           | VIA            | General Conditions of Contract (GCC)          |              |
| 9.           | VIB            | Special Conditions of Contract (SCC)          |              |
| 10.          | VIC            | Contract Forms (CF)                           |              |

***(To be submitted on letter head of Firm)***

**Bid Submission Letter**  
*(Declaration Form)*

Sub: - Regarding Bid submission for **NIT/JKMSCL/M&E/2023/596**

**DATED 02-08-2023**

I/We..... *(Name, Designation and Address of Bidder)* having our office at..... *(Address of Firm)* do hereby declare that I/We have read all the terms & conditions of the bid document floated by JKMSCL and agree to abide by all the terms & conditions set forth therein.

I/We declare that we are participating in this bid in the capacity of ..... *(Manufacturer /Direct Importer/ Authorized representative of the original manufacturer)* I/We have enclosed all the requisite documents and are as per the requirement of the NIT.

I/We further declare that the rates offered by us shall remain valid for the period of 24 months extendable for a further period of three months and shall reduce the rates, if the rates are reduced by us for any other buyer during this period within Union of India. **I/We have enclosed the documents as per details given in Annexure I of the NIB and other documents asked in NIT.**

We further undertake to abide by all the terms & conditions of the NIB.

Dated

Name and signature of bidder with seal



## JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

**Corporate Head Office:** Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

**Corporate Office:** Opposite J&K Motor Garage Deptt near Hajj House Bemina Srinagar

**Telephone:** 0191-2478842; 191-3510489 (Jammu), 0194-2490662 (Srinagar)

**email:** mdjkmscl2@gmail.com; ismjkmscl2018@gmail.com **website:** www.jkmsclbusiness.com

**Tender No. NIT/JKMSCL/M&E/2023/596**

**Dated: 02 /08/2023**

### NOTICE INVITING TENDER

On Behalf of Jammu & Kashmir Medical Supplies Corporation Limited, e-bid under two cover system (Technical bid in cover 1 and Financial bid in cover-2) is invited for the finalization of Rate Contract for the procurement of “**Machinery & Equipment**” from the Original manufacturers / Direct importers/ Authorized Representatives of the manufacturers/ direct importers. Detailed tender document may be downloaded at J&K Govt. Portal [www.jktenders.gov.in](http://www.jktenders.gov.in), [www.jkmsclbusiness.com](http://www.jkmsclbusiness.com),. The cost of the tender along with tender processing charges of Rs.10,000/- (Rupees Ten thousand only/-) i.e. Rs.1,000/- (Rupees one thousand only) as cost of tender & Rs.9,000/- (Rupees Nine thousand only) as tender processing charges shall have to be paid either through **NEFT/RTGS only** in the Corporation's Bank Account No. 0373040500000032 maintained at J&K Bank Limited, Branch Medical College Jammu, IFSC Code JAKA0MEDJAM **or by depositing the amount directly into the above Account No.** along with the submission of requisite valid documentary proof.

- **IMPS mode of transfer is not verifiable and hence shall not be entertained as tender fee or tender processing charges. Bidders claiming to submit money through IMPS Mode shall be out-rightly rejected.**
- **DD as mode of payment for cost of tender/tender processing fees/Bid Security shall only be entertained if the same is deposited physically against proper receipt in the Corporate Office of JKMSCL, before the closing due date of e-bid.**
- **Bid Security** Rs. 1,00,000.00 in the form of FDR/CDR/BG/RTGS/NEFT (FDR/CDR from scheduled/Nationalised Bank / BG from Nationalised Bank) with validity of 30 months. Bids submitted without sufficient bid security & validity shall be summarily rejected. **Firms which are registered as (Micro and Small Enterprise) MSEs Unit(s) as defined in MSE Procurement Policy issued by Department of Micro, Small and Medium Enterprises (MSME) shall be considered for Exemption of bid security including tender fee of Rs. 1000/- as per provisions of MSME Policy. Tender Processing charges of Rs. 9000/- is to be paid by the MSE Unit(s) also.**
- Physical hard copy of Bid Security in form of FDR/CDR/BG may be submitted to the Corporate Head Office before closing the due date of e-bid. Scanned copy of the same shall be uploaded along with Technical Bid, failing which bid shall be out rightly rejected.
- **The bidders seeking EMD exemption must submit the valid supporting document for the relevant category. Under MSE category only manufacturers for goods and service providers for services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.**
- **Scanned copies of NEFT/RTGS/Bank Transfer/Receipt towards the cost of tender documents and tender processing charges shall have to uploaded along with Technical Bid, failing which bid shall be out rightly rejected.**

Sd/-

Managing Director

Jammu and Kashmir Medical Supplies Corporation Ltd.

**Note: The bidders who opt to bid for multiple manufacturer shall have to provide complete details of each manufacturers in a systemic way covering all documents asked in Cover-A. Separate sheet shall have to be attached for every individual item.**



## JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

**Corporate Head Office:** Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

**Corporate Office:** Opposite J&K Motor Garage Deptt near Hajj House Bemina Srinagar

**Telephone:** 0191-2478842; 191-3510489 (Jammu), 0194-2490662 (Srinagar)

**email:** mdjkmscl2@gmail.com; ismjkmscl2018@gmail.com **website:** www.jkmsclbusiness.com

### BIDDING DOCUMENT FOR Procurement of Machinery & Equipment

**Tender No.** NIT/JKMSCL/M&E/2023/596

**: 02-08-2023**

Date of publication of e-bid

: 03-08-2023

Start date and time for download of bid document

: 03-08-2023

Last date and time for download of bid document

: 07-09-2023 at 1600 hrs

Clarification start date

: 03-08-2023 at 1100 hrs

Clarification end date

: 14-08-2023 upto 1400 hrs

**Pre- bid conference**

**: 15-08- 2023 AT 11.00 A.M**

**(at Corporate Office, Jammu/Srinagar)**

Start date and time for submission of online bids

: 03-08-2023 at 1500 hrs

Last date and time for submission of online bids

: 07-09-2023 at 1600 hrs

Date and time for online opening of technical bids

: 09-09-2023 at 1100 hrs

Cost of tender document

: Rs. 1000/-

Tender Processing charges

: Rs. 9000/-

#### ADDRESS FOR COMMUNICATION:

**Managing Director or General Manager,  
J&K Medical Supplies Corporation Ltd,**

Address: Plot No. 58, Friends Colony  
Satyam Road Trikuta Nagar, Jammu  
Bemina Near Haj House- Srinagar  
(Kashmir)

#### Note: -

1. The bidder shall have to get themselves updated with the date & time fixed for Pre-bid as per the item list. After pre-bid meeting necessary changes in bid conditions shall be done with the recommendations of panel of technical experts drawn from the intending department after the approval of the competent authority. Bid should be submitted through e-portal [www.jktenders.gov.in](http://www.jktenders.gov.in) after pre-bid meeting including all the clarifications/ modifications/ amendments.
2. Corrigendum/addendum shall be the integral part of terms & conditions of bid which shall be duly signed and attached with the bid document by the bidder.
3. The JKMSCL is not bound to accept the lowest bid and may reject any/part thereof or all bids without assigning any reason thereof.
4. The bidders shall have to submit a **GST No. and valid 'GST'** clearance

certificate/returns submitted from the taxation department and the 'PAN' issued by income tax department.

5. It is clarified that the information required in bidding document should be submitted only in enclosed format bidding forms without any change or modification in its formats. Bids submitted with changed or modified annexure/ formats shall be rejected.
6. Information of award of contract shall be communicated to all participating bidders on the website [www.jktenders.gov.in](http://www.jktenders.gov.in). [www.jkmsclbuisness.com](http://www.jkmsclbuisness.com)

**Note: -**

*If any amendment/clarification is carried out in the technical specifications and bid terms & conditions following pre-bid meeting or any other information, the same shall also be uploaded on the websites mentioned above and the bidders shall keep themselves updated by regularly visiting the website/jk portal.*

**Important Note:**

- 1. No representation shall be allowed, accepted and entertained after the Pre-bid meeting (i.e upto 4.00 P.M of Pre-bid date). Bidders are requested to submit their queries/clarifications by or before the date fixed (mentioned above), so that the same can be discussed and clarified during the Pre-bid meeting.**
- 2. Google Code : <https://meet.google.com/nmf-eptv-mcv>**

TABLE-1

| S. No. | Item code | Name of the item  | Average Annual turnover for last 03 years |
|--------|-----------|---|---|
| 1      | DHJ2302   | Biological Safety Cabinet Class 2A with thimble ducting | 05 Crore                                  |
| 2.     | DHJ2303   | Deep Freezer (-86°C)                                    | 05 Crore                                  |
| 3      | DHJ2304   | Deep Freezer (-20°C)                                    | 05 Crore                                  |
| 4      | DHJ2305   | Electric micro incinerator for loop                     | 05 Crore                                  |
| 5      | DHJ2306   | Microliter Centrifuge                                   | 05 Crore                                  |
| 6      | DHJ2307   | PCR Thermocycler  | 05 Crore                                  |
| 7      | DHJ2308   | PCR Workstation   | 05 Crore                                  |
| 8      | DHJ2309   | pH Meter  | 01 Crore                                  |
| 9      | DHJ2310   | Refrigerated Centrifuge with UPS                        | 05 Crore                                  |
| 10     | DHJ2311   | Water Bath  | 01 Crore                                  |
| 11     | DHJ2312   | Water Distiller   | 01 Crore                                  |
| 12     | DHJ2313   | BD BACTECTM MAGITTM 960 System                          | 05 Crore                                  |
|        |           | MAGITTM Accessories                                     |   |
|        |           | GT Bolt 48  |   |
|        |           | Twincubator   |   |
| 13     | DHJ2314   | Thermometers  | 01 Crore                                  |
| 14.    | DHJ2315   | Electronic max-min thermo hygrometer                    | 01 Crore                                  |
| 15     | DHJ2316   | Glass Alchol thermometer                                | 01 Crore                                  |
| 16     | DHJ2317   | Digital Maxima-Minima Thermameter (-100 to 60 °C)       | 01 Crore                                  |
| 17.    | DHJ2318   | Digital Maxima-Minima Thermameter (0 to 300 °C)         | 01 Crore                                  |
| 18.    | DHJ2319   | Dial Thermometer  | 01 Crore                                  |

The Average Annual Turn Over required for the above items pertaining to Group "Procurement of Machinery & Equipment" is mentioned above. Only the bid(s) falling under the category as specified under Annual Turnover is accepted. The bid(s) not falling under the Annual Turnover clause shall be out rightly rejected. Only manufacturer(s) or Importer(s) fulfilling the turnover clause shall be eligible to participate the e.bid.

**Note:**

1. The catalogues/brochures of the item shall be submitted along with the EMD in separate envelopes, prior to submission of online bids. The catalogues/brochures pertaining to the equipment information should be signed by the authorized signatory of the manufacturer.
2. No minimum quantity is guaranteed and the bidder shall not claim any compensation from the Jammu & Kashmir Medical Supplies Corporation Ltd.

## DISCLAIMER

The information contained in this bid document for proposed procurement or subsequently provided to the Bidder(s), in documentary or any other form by or on behalf of the Jammu and Kashmir Medical Supplies Corporation Ltd. (procuring entity) or any of its employees or advisors, is provided to bidder(s) on the terms and conditions set out in this bid and such other terms and conditions subject to which such information is provided to the bidder.

Whilst the information in this bid has been prepared in good faith and contains general information in respect of the proposed procurement, the bid is not and does not purport to contain all the information which the bidder may require.

Jammu and Kashmir Medical Supplies Corporation Ltd., does not accept any liability or responsibility for the accuracy, reasonableness or completeness of, or for any errors, omissions or misstatements, negligent or otherwise, relating to the proposed procurement, or makes any representation or warranty, express or implied, with respect to the information contained in this bid or on which this bid is based or with respect to any written or oral information made or to be made available to any of the recipients or their professional advisers and liability therefore is hereby expressly disclaimed.

This document is neither an agreement and nor an offer or invitation by the Jammu and Kashmir Medical Supplies Corporation Limited, (hereinafter referred to as "procuring entity") to the prospective bidders or any other person. The purpose of the bid document is to provide interested parties with information to assist the formulation of their proposal/offer. The information contained in this bid document is selective and is subject to updating, expansion, revision, and amendment. Each recipient must conduct its own analysis of the information contained in this bid document or to correct any inaccuracies therein that may be in this bid document and is advised to carry out its own investigation into the proposed procurement, the legislative and regulatory regime which applies thereto and by and all matters pertinent to the proposed procurement and seek its own professional advice on the legal, financial, regulatory and taxation consequences of the entering into any agreement or arrangement relating to the proposed procurement.

This bid document includes certain statements, estimates and targets with respect to the procurement. Such statements, estimates and targets reflect various assumptions made by the procuring entity, (and the base information on which they are made) which may or may not prove to be correct. No representation or warranty is given as to the reasonableness of forecasts or the assumptions on which they may be based and nothing in this bid document is, or should be relied on as, a promise, representation, or warranty. Bid document and the information contained therein is meant only for those applying for this procurement, it may not be copied or distributed by the recipient to third parties, or used as information source by the bidder or any other in any context, other than applying for this proposed procurement.



The procuring entity is, its employees and advisors make no representation or warranty and shall have no liability to any person, including any bidder under any law, statute, rules or regulations or tort, principles of restitution or unjust enrichment or otherwise for any loss, damages, cost or expense which may arise from or be incurred or suffered on account of anything contained in this bid document or otherwise, including the accuracy, adequacy, correctness, completeness or reliability of the bid document and any assessment, assumption, statement or information contained therein or deemed to form part of this bid document or arising in any way for participation in this bidding process.

The procuring entity also accepts no liability of any nature whether resulting from negligence or otherwise howsoever caused arising from reliance of any bidder upon the statements contained in this bid document.

The procuring entity may in its absolute discretion, but without being under any obligation to do so, update, amend or supplement the information, assessment or assumptions contained in this bid document.

The issue of this bid document does not imply that the procuring entity is bound to select a bidder or to appoint the selected bidder or bidder, as the case may be, for the procurement and the procuring entity reserves the right to reject all or any of the bidders or bids at any point to time without assigning any reason whatsoever.

The bidder shall bear all its costs associated with or relating to the preparation and submission of its bid including but not limited to preparation, copying, postage, delivery fees, expenses associated with any demonstrations or presentations which may be required by the procuring entity or any other costs incurred in connection with or relating to its bid. All such costs and expenses shall remain with the bidder and the procuring entity shall not be liable in any manner whatsoever for the same or for any other costs or other expenses incurred by a bidder in preparation or submission of the bid, regardless of the conduct or outcome of the bidding process.

Any information/documents including information/ documents pertaining to this bid or subsequently provided to bidder and/or selected bidder and information/documents relating to the bidding process; the disclosure of which is prejudicial and/or detrimental to, or endangers, the implementation of the procurement is not subject to disclosure as public information/documents.

**Managing Director**  
**Jammu and Kashmir Medical Supplies Corporation Ltd**

### Section-I Instruction To Bidders (ITB)

Before uploading bid, kindly go through the following instructions carefully so that your bid may not be considered invalid:

| Clause No. | Description  |
|------------|--|
| 1.         | Go through the terms and conditions, annexure and other forms of the document carefully and meticulously & get your digital signatures available for uploading.  |
| 2.         | Bid form must conform the terms & conditions of the bid documents and <b>Technical Bid in Cover-'A' &amp; Financial Bid in Cover-'B' to be uploaded on <a href="http://www.jktenders.gov.in">www.jktenders.gov.in</a>. The receipt for cost of tender, tender processing fee and catalogues of the quoted items shall be submitted in the office of JKMSCL atleast one day prior to submission of online bids.</b>   |
| 3.         | It is expected from all bidders that CDR/FDR/BG in separate envelope shall be deposited with the authorised person of JKMSCL at reception against proper receipt from the concerned.   |
| 4.         | Correspondences/Complaints lodged to JKMSCL should bear signature, name, I.D proof and mobile number of the complainant. Unauthenticated correspondence/complaints may not be acted upon. If any bidder intends to lodge a complaint or make a suggestion with regards to some bid condition, it shall be done in the Pre-bid conference, in the office of JKMSCL in writing. After the stipulated period as decided by the JKMSCL, no such complaint/ suggestion would normally be considered.  |
| 5          | Certificates/Licenses/Documents which are required should be complete and updated. The bidder shall submit acceptance of terms and conditions of the tender document.  |
| 6          | If there is any query in bid document/uploading process, bidder may contact JKMSCL office at Jammu/Srinagar during working hours i.e 1000 hrs to 1600 hrs on ph. 0191-2580842, 0194-2432008 or e-mail on <a href="mailto:gmkjkmscl1@gmail.com">gmkjkmscl1@gmail.com</a> / <a href="mailto:jkmaclepm@gmail.com">jkmaclepm@gmail.com</a> / <a href="mailto:gmjjkmscl@gmail.com">gmjjkmscl@gmail.com</a>  |
| 7.         | In case a bidder is given any assurance what so ever of being provided with any advantage in JKMSCL by anybody or if a bidder is directly or indirectly threatened of being put to some deliberate disadvantage in the bidding process & in the bidder's subsequent association/ working with JKMSCL, it is requested that the concerned must immediately inform about the same to the Managing Director, JKMSCL/G.M-J(Adm), JKMSCL in writing or through e-mail on <a href="mailto:gmjjkmscl@gmail.com">gmjjkmscl@gmail.com</a> . It is advised that evidence of such unfair activity of such person, if available, is produced along with the complaint, so that action can be taken against such a person(s) and that their details can be put on the website so that other bidders can be forewarned in this regard. |
| 8          | The Bidders shall have to submit a GST No. & GST clearance certificate/return submitted from the concerned commercial taxes officer and the 'PAN' issued by income tax department.   |

|    |   |
|----|---|
| 9  | It is clarified that the information required in bidding document should be uploaded as per enclosed bidding form without any change or modification in its formats. Bids submitted with changed or modified annexure/ formats shall out rightly be rejected.   |
| 10 | The declaration of technical bid in respect of responsive/non responsive bidders shall be uploaded on website <a href="http://www.jktenders.gov.in">www.jktenders.gov.in</a> Similarly, information regarding financial bid (L-1) shall also be provided to bidders on above websites. Individual bidders shall not be informed separately.   |
| 11 | <b>No firm/bidder/manufacture/importer shall provide/supply any of the product item at the rate contract /approved by JKMSCL to any of the department/NGO/other procuring institute within or outside the Union Territory of J&amp;K. In case any supply is made in violation to the said condition (or), the supplier/firm shall be liable to be penalised to the tune of 7.5% of order placed/blacklisting for a period not less than five years (or) both as deemed fit, to the competent/Tender Inviting Authority. However, JKMSCL can procure the items for any of the departments within /outside the Union Territory of J&amp;K/after charging the administrative expenses.</b>   |
| 12 | The qualified bidders are required to submit the relevant documents and annexure uploaded with their e.bid in original along with catalogues at the time of issuance of LOI /execution of agreement before issuance of rate contract.   |
| 13 | The bidder shall not under any circumstances quote "Zero" anywhere in the BOQ.  |
| 14 | <p><b>Important Instructions to bidders</b></p> <p>The bidders shall have to abide the clauses/restrictions in terms of Rule 144 (xi) of the General Financial Rules (GFRs) issued by the Ministry of Finance, Department of Expenditure, Public Procurement Division vide No. F.No.6/18/2019-PDD dated 23.07.2020.</p> <p>The bidders are required to submit a certificate/ declaration regarding their compliance with this order. If such certificate given by a bidder whose bid is accepted and is found to be false, it will be a ground for immediate termination &amp; further legal action in accordance with law. Bidders are required to go through the said order &amp; Office Memorandum (s) for the necessary compliance</p> <p><b>Model Certificate for tenders</b></p> <p><i>"I have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India. I hereby certify that this border is not from such a country and is eligible to be considered."</i></p> <p><b>Model Certificate for Tenders</b></p> <p><i>"I have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India; I certify that this bidder is not from such country or, if from such a country, has been registered with the Competent Authority. I hereby certify that this bidder fullfills all requirements in this regard and is eligible to be considered (where applicable, evidence of valid registration by the competent Authority shall be attached"</i></p> |

**Section-II: Bid Data Sheet (BDS)**  
**Table of Contents**

| S. No. | Description  | Pages |
|--------|--|-------|
| 1.     | Introduction                                       |       |
| 2.     | Bidding Document                                   |       |
| 3.     | Preparation of Bids                                |       |
| 4.     | Submission and Opening of Bids                     |       |
| 5.     | Evaluation and Comparison of Bids                  |       |
| 6.     | Award of Contract                                  |       |
| 7.     | Redressal of Grievances during Procurement Process |       |

### Section III: Evaluation and Qualification Criteria

#### 2. Qualification Criteria

The lowest evaluated bidder shall have the necessary qualifications to successfully fulfil its obligation under the contract. Minimum acceptable levels with regards to bidder's experience in supply of goods and related services with comparable technical parameters, its financial capability and other factors are defined.

| Clause No. | Description  |
|------------|--|
| 1.         | <b>Contractual experience:-</b><br>The bidder shall be an original manufacturer; direct importer; (or) authorised representative of the original manufacturer/direct importer, who must have manufactured/ imported, supplied and installed such equipments in India satisfactorily. The list of such installations may be asked from the bidder and the bidder should submit self attested copy of purchase order, indent and invoice (inclusive of quantity & rate). |
| 2.         | <b>Technical experience:-</b><br>Client Base on Letter Head of the Bidder / Manufacture/Indian Subsidiary of Principal Manufacturer with references of the supply orders, for any of the three years in last five years along with satisfactory performance certificate of minimum one installation <b>(Copies of reference supply orders and satisfactory performance certificate need to be attached)</b>  |
| 3.         | <b>Production capacity :</b><br>The JKMSCL may fix the minimum supply and/ or production capacity required to assure that the bidder is capable of supplying the type, size and quantity of goods required. It should be dedicated quantity to JKMSCL on monthly and annual basis. Production capacity certificate be attached with uploaded document.   |
| 4.         | <b>Financial position:-</b><br>The soundness of the bidders financial position showing long term profitability demonstrated through audited annual financial statement (balance sheet, income statement etc.) for last three years.  |
| 5.         | <b>Cash Flow capacity :</b><br>The bidder should have sufficient availability of/ access to liquid assets, lines of credit and other finances to meet the possible cash flow requirement which may arise during the execution of the rate contract.  |
| 6.         | <b>Litigation history:-</b><br>The information regarding all pending claims, arbitration, or other litigation is asked by the JKMSCL   |
| 7.         | <b>Tax clearance certificates:-</b><br>The Latest GST returns and other tax clearance certificate (latest) or declaration to be submitted by the bidder. Bidders shall have to submit a valid & latest 'GST' clearance certificate/return submitted online as per GST rules along with GST No. and the 'PAN' issued by concerned department.   |
| 8.         | <b>Declaration regarding qualifications :-</b><br>Declaration regarding qualifications of the bidder shall be given in specified format provided in bidding forms.   |

## 1. Evaluation Criteria

| Claus | Description  |
|-------|--|
| 1.    | <b>Scope</b>   |
| 1.1   | <b>Local handling and inland transportation:-</b> The cost for Inland transportation, insurance, related services, installation, commissioning, demonstration and other incidental costs for delivery of goods, or port of entry, or supply point to consignee site, schedule of supply shall be quoted in price schedule.   |
| 1.2   | <b>Minor omission and missing items:-</b> Pursuant to the relevant clauses, the cost of all quantifiable non-material non-conformities or omissions from the contractual and commercial conditions shall be evaluated. The procuring entity will make its own assessment of the cost of any non-material non-conformities and omissions for the purpose of ensuring fare comparison of bids.   |
| 2.    | <b>Technical Criteria:-</b> The minimum technical level that the goods and related services shall have in order to comply with the Section V, schedule of supply are specified. These criteria are evaluated on a pass-fail system, with a minimum acceptable level for each criteria enumerated in technical specifications of item. However, a minor deficiency in technical compliance may not be cause for rejection of the bid.   |
| 3.    | <b>Economic Criteria: -</b> The economic criteria are most important when evaluating a Bid. The price, however, may not be the only criterion, as there could be technical evaluation that may be expressed in mandatory terms i.e. cost per test etc. The following may be examples: - 3.1, 3.2....   |
| 3.1   | <b>Adjustment for deviations in the delivery and completion schedule: -</b> The deviation from the delivery and completion schedule specified in Section V, schedule of supply are permitted. No credit will be given for earlier completion.  |
| 3.2   | <b>Operation and maintenance cost:</b> The operation and maintenance costs of equipments are taken into account for bid evaluation purposes. The methodology is elaborated at BOQ for determining lowest bid (L-1).  |
| 3.2   | <b>Spare parts: -</b> Only those spare parts and tools which are specified on an item wise basis in the list of goods and related services, schedule of supply shall be taken in account in bid evaluation. Supplier recommended spare parts for specified operating requirement shall not be considered in bid evaluation.<br><b>The price for evaluating L1 (as per BOQs) shall be decided on the basis of cumulative rates of Main item, Accessories (wherever asked) /Indian items (wherever asked) i.e cost of main item + cost of Accessories = Total cost of equipment.</b> |
| 3.3   | <b>Performance and productivity of goods:-</b> The performance and productivity of the equipments shall be as per the reference value or norms specified in technical specification of an item and corresponding value guaranteed by the bidder in its bid.  |
| 4.    | <b>Price preference:-</b>  |
| 4.1   | The price preference shall be given in evaluation of bids and award of contract as per MSME Policy in vogue.   |
| 4.2   | <b>Taxes as</b> applicable, should be mentioned clearly and separately.  |

## Section IV: Bidding Forms

### Table of Contents

| S.No | Name of Bidding Forms                                   | Pages |
|------|---|-------|
| 1    | Bid security  |       |
| 2    | Bid / Tender charges ( Incl. Tender processing fee)     |       |
| 3    | List of Items Quoted (Annexure I)                       |       |
| 4    | Technical bid submission sheet (Annexure II)            |       |
| 5    | Financial bid format (BOQ) (Annexure III)               |       |
| 6.   | Declaration and undertaking (Annexure IV)               |       |
| 7    | Client Base (Annexure V)                                |       |
| 8    | Authorisation from principal manufacturer (Annexure VI) |       |
| 9    | Average Annual Turnover Statement (Annexure VII)        |       |

(Annexure I)

On Firm's letter head

**LIST OF ITEMS QUOTED IN THE BID**

| S.<br>No. | Tender<br>Sr.<br>No. | Code | Name of Item | Manufactured By | Imported by | Make & Model<br>quoted/ offered | <u>Quality Certification</u> |     |    |       |           |
|-----------|----------------------|------|--------------|-----------------|-------------|---------------------------------|------------------------------|-----|----|-------|-----------|
|           |                      |      |              |                 |             |                                 | BIS License                  | ISO | CE | USFDA | Any Other |
|           |                      |      |              |                 |             |                                 |                              |     |    |       |           |
|           |                      |      |              |                 |             |                                 |                              |     |    |       |           |
|           |                      |      |              |                 |             |                                 |                              |     |    |       |           |
|           |                      |      |              |                 |             |                                 |                              |     |    |       |           |
|           |                      |      |              |                 |             |                                 |                              |     |    |       |           |

**Seal & Signature  
(Authorised Signatory)**



**(To be submitted on Firms' letter head)**  
**Technical Bid Submission Sheet (Cover 'A')**

**Managing Director**

Jammu &amp; Kashmir Medical Supplies Corporation Ltd.

J&amp;K

We, the undersigned, declare that:

1. I/We .....have examined and have no reservations to the bidding document of NIB No. .... dated.....including addenda/clarification No.:.....dated ..... We offer to supply in conformity with the bidding document and in accordance with the delivery schedule specified in Section V, schedule of supply, the following goods and related services..... *Name of the item and Guarantee period plus etc.* .....
2. Our bid shall be valid for a period of minimum 120 working days from the date of technical bid opening in accordance with the bidding document, and it shall remain bidding upon us and may be accepted at any time before the expiration of that period. However, validity may also be extended with mutual consent;
3. If our bid is accepted, we commit to submit a performance security in the amount of 3% of the contract price or as specified in bid document for the due performance of the contract;
4. Our firm, including authorised representative for any part of the contract, have nationalities from the eligible countries;
5. I/We are not participating, as bidders, in more than one bid in this bidding process, in the bidding document;
6. Our firm, its affiliates or subsidiaries, including authorised representative has not been debarred by the Union Govt/any State Government or the procuring entity.
7. I/We understand that this bid, together with your written acceptance thereof included in your notification of award, shall constitute a binding contract between us, until a formal contract is prepared and executed;
8. I/We agree to permit the JKMSCL to ask any relevant documents. I/We shall be bound to provide the said relevant document within the specified period.
9. My/our quoted items..... (*Name of item*).....fully comply with the technical specifications as per bid document Section V, schedule of supply.
10. **The bidder shall ensure that the bid document sheet shall be properly filled with particulars, page numbering and tender document should be properly numbered.**
11. I/We certify that I/We have annexed the following documents with particulars & page No. mentioned against each column :

| S. No | Item  | Particular        | Manufacturer |    |    |    |
|-------|---|-------------------|--------------|----|----|----|
|       |   |                   | M1           | M2 | M3 | M4 |
| 1.    | Bid security (as mentioned above)   |                   |              |    |    |    |
| 2.    | Cost of Tender & Tender Processing charges  |                   |              |    |    |    |
| 3.    | List of Items quoted by the Bidder mentioning name of manufacturer/ importer with make & model as per annexure. | <b>Annexure I</b> |              |    |    |    |

|    |  |               |  |  |  |  |
|----|--|---------------|--|--|--|--|
| 4  | Copy of Catalogue of the Quoted product (self attested)  |               |  |  |  |  |
| 5  | Compliance Sheet for each equipment (self attested)  |               |  |  |  |  |
| 6  | Technical bid submission sheet duly filled   | Annexure II   |  |  |  |  |
| 7  | Financial bid (To be uploaded in BOQ only)   | Annexure III  |  |  |  |  |
| 8  | <b>Declaration for Latest Non Conviction, non blacklisting on non judicial Stamp paper of Rs 100 furnished by the Principal Manufacturer / Sole Importer/ Indian Subsidiary as per proforma duly notarised.</b>  | Annexure IV A |  |  |  |  |
| 9  | <b>Declaration for Latest Non Conviction, non blacklisting on non judicial Stamp paper of Rs 100 furnished by the Bidder as per proforma duly notarised.</b>   | Annexure IV B |  |  |  |  |
| 10 | Client Base on Letter Head of the Bidder / Manufacture/Indian Subsidiary of Principal Manufacturer with references of the supply orders, for any of the three years in last five years along with satisfactory performance certificate of minimum one installation <b>(Copies of reference supply orders and satisfactory performance certificate need to be attached)</b>                                   | Annexure V    |  |  |  |  |
| 11 | <b>Authorisation from principal manufacturer / Importer</b><br><i>(On the letterhead of Principal manufacturer / Sole Importer)</i><br><i>In case authorization to the bidder is furnished by the Sole Importer/Indian Subsidiary, document confirming authorization from foreign Principal Manufacturer in favour of Indian Subsidiary / Sole Importer is to be submitted (strictly as per annexure VI)</i> | Annexure VI   |  |  |  |  |
| 12 | Average Annual Turnover Statement for Last 3 financial Years of the Indian Subsidiary of Principal Manufacturer/ Sole Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21).<br><b><i>In case of foreign manufacturer the turnover of Indian Subsidiary/Sole Importer only shall be considered and not of the original manufacturer.</i></b>                  | Annexure VII  |  |  |  |  |
| 13 | Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer/ Indian Subsidiary for 2018-19, 2019-20 & 2020-21 with UDIN.<br><b><i>In case of foreign manufacturer the balance sheets of Indian Subsidiary/Sole Importer only shall be considered</i></b>  |               |  |  |  |  |
| 14 | Nature of the Firm/Public Company / Private  | Annexure      |  |  |  |  |

|           |  |                                   |  |  |  |  |
|-----------|--|-----------------------------------|--|--|--|--|
|           | Company/ Partnership/ Proprietorship/any other with Documentary proof.   | <b>VIII</b>                       |  |  |  |  |
| <b>15</b> | Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product)           | <b>Annexure A (if applicable)</b> |  |  |  |  |
| <b>16</b> | Copy of GST Registration of the Bidder   | <b>Annexure B</b>                 |  |  |  |  |
| <b>17</b> | Latest GST Returns of the Bidder   | <b>Annexure C</b>                 |  |  |  |  |
| <b>18</b> | Copy of the PAN Card of the Bidder   | <b>Annexure D</b>                 |  |  |  |  |
| <b>19</b> | Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable.  | <b>Annexure E</b>                 |  |  |  |  |
| <b>20</b> | Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.            | <b>Annexure F</b>                 |  |  |  |  |
| <b>21</b> | Specify point of supply with full Address.<br>NB: Specifying of point of supply does not means authorization to raise, invoice and receive payments on behalf of bidder(s) | <b>Annexure G</b>                 |  |  |  |  |
| <b>22</b> | Declaration of bidder regarding acceptance Bid for terms & conditions  | <b>Annexure A1</b>                |  |  |  |  |

#### **Important Note**

- The Bidders who opt to bid for multiple manufacturers shall have to provide complete details of each manufacturer in a systemic way, sequentially, covering all documents asked in Annexure "II".**
- Please Note the Annexure A"II" should be properly filled showing the page number when the asked document has been attached. All the documents attached with the technical bid should be properly page numbered.**

I/we understand that our bid shall liable to be declared non responsive in case of any deficiency in fulfilment of above requirements on our part.

I/we accept all the terms, conditions and provisions of this bid document.

Name/Address.....  
in the capacity of.....(Designation).....  
Signed..... duly authorized to sign the bid for and on behalf of..... of Firm).....  
Dated..... Tel:.....e-mail:.....

N.B : The original manufacturer/direct importer of the bidding items/their sole authorised representative shall execute tri-partite agreement with the Corporation i.e JKMSCL, inter-alia, stating that :

i. The invoice submitted by the authorised representative for such supplies shall be endorsed by the original manufacturer/direct importer of bidding items. Original copy of the delivery challan of the manufacturer towards authorised representative for such supplies shall be endorsed along with invoice submitted by Authorised representative.

ii. JKMSCL may secure confirmation/or authenticating of such supplies from manufacturer/direct importer before releasing the payment.

iii. No original manufacturer/direct importer shall be allowed to authorize more than one representatives to bid, to negotiate/to raise invoice or to receive payments & to enter into tripartite agreement with regard to business against this specific tender.

iv. In case, original manufacturer/direct importer wish to authorise any representative to bid, to negotiate, to raise invoice to receive payments on behalf of original manufacturer/direct importer/against invoice raised by representative, Annexure All duly filled shall need to be uploaded alongwith e.bid ; otherwise no representation in this matter shall be entertained in the later stage.

## ITEM WISE FINANCIAL BID (BOQ)

## For Uploading Rates of Equipment

Please read the amended BOQ as follows:

- i) The rates shall be quoted in the BOQ as per format mentioned below.  
 ii) The rates of the accessories, if any, shall be quoted cumulative as per NIT.  
 iii) The rates of the India items, if any, shall be quoted cumulative as per NIT

| S. No.           | Item Description       | Item Code                    | Unit | Qty                          | Currency type | Basic Equipment cost for one unit | Packing & forwarding charges/freight insurance charges | Indian Agency Commission for 1 unit in foreign | Custom Duty | SGST             | CGST | IGST | Custom clearance in foreign | Total Amount including Taxes |
|------------------|------------------------|------------------------------|------|------------------------------|---------------|-----------------------------------|--|--|-------------|------------------|------|------|-----------------------------|------------------------------|
| 1                | 2                      | 3                            | 4    | 5                            | 6             | 7                                 | 8  | 9  | 10          | 11               |      |      | 12                          | 13                           |
| 1                | Main item              |                              |      |                              |               |                                   |  |  |             |                  |      |      |                             |                              |
| 2                | Accessories, if any.   |                              |      |                              |               |                                   |  |  |             |                  |      |      |                             |                              |
| 3                | Indian items, if any   |                              |      |                              |               |                                   |  |  |             |                  |      |      |                             |                              |
| 5.               | Optional Items, if any |                              |      |                              |               |                                   |  |  |             |                  |      |      |                             |                              |
| 6.               |                        |                              |      |                              |               |                                   |  |  |             |                  |      |      |                             |                              |
| CMC for 1st Year |                        | CMC for 2 <sup>nd</sup> Year |      | CMC for 3 <sup>rd</sup> Year |               | CMC for 4 <sup>th</sup> Year      |  | CMC for 5 <sup>th</sup> Year                   |             | Total amount CMC |      |      |                             |                              |
| 14               |                        | 15                           |      | 16                           |               | 17                                |  | 18   |             |                  |      |      |                             |                              |
|                  |                        |                              |      |                              |               |                                   |  |  |             |                  |      |      |                             |                              |

**Note: -**

- The rate quote should be as per BOQ.
- CGST, SGST or IGST should be separately shown in absolute amount only.
- Rate should be quoted only for packing units as mentioned in the bid
- No quantity or cash discounts should be offered.
- Read all the terms & conditions before filling the Annexure III.
- Please quote rates in absolute amount only.
- Please quote rates per unit only
- The bidder shall not under any circumstances quoted "Zero" anywhere in the BOQ.
- Finalization of the rates shall be made on the basis of price quoted in BOQ
- Custom duty, if applicable shall be indicated separately.
- The final rates quoted at Column No. 13 shall be considered as final rates and shall be considered for evaluating financial bid. L1 rate shall be finalised on the basis rate and taxes as applicable.
- The price for evaluating L1 (as per BOQs) shall be decided on the basis of cumulative rates of Main item, Accessories (wherever asked) i.e cost of main item + cost of Accessories + Indian items = Total cost of equipment.**

**13. The rates quoted for the CMC (Comprehensive Maintenance Contract) and Optional items shall not be considered for finalizing/deciding L1 rates.**

**14. Warranty of 05 years shall be applicable.**

The bidder may quote in foreign currency as per the BOQ uploaded in the e.portal on the following terms & conditions

**(For Imported equipment)**

**100% payment shall be released against 20% Bank Guarantee valid for a period of 12 months, to be submitted by the bidder. The BG shall be released on successful installation of the Machinery**

**Letter of credit would be opened subject to following additional conditions :**

1. At site LC would be opened.
2. A level vessel would be used for shipment of supplies which should not be more than 15 years old.
3. Supplies shall be insured vide comprehensive Insurance Policy including machine insurance by the OEM till the final delivery site shall also include "Force Majeure".
4. Pre-dispatch inspection shall be carried out by OEM by certified inspection agency before shipment of supply.
5. ***The CIF (cost insurance freight)/CIP (cost insurance price) upto New Delhi, should be in Foreign Currency, payable by the Principal company in that currency only as per the mode of L.C stipulations. The CIF prices shall be borne by the firm upto site.***
6. ***The custom duty shall be paid as per actual on the production of documentary proof. No Custom duty exemption certificate shall be issued by JKMSCL to facilitate custom clearance on the concessional rates.***
7. ***CIF price of optional accessories, if any, Percentage of Indian direct Importer/authorized representative's percentage (Indian agency commission), if any, on FOB (Freight on board) Price which shall be payable to the Indian direct Importer (Indian Agency) in Indian currency at the exchange rate as mentioned below. However local accessories, if quoted in Indian currency, GST shall be paid as admissible under rules.***
8. The prices quoted should be as per the price of the manufacturer applicable in within India.
9. ***The L1 shall be calculated on the basis of conversion of currency as on date of opening of financial bid.***

**Delivery Period shall be 60 days for Indian Items and 90 days for Imported items.**

**PLEASE DON'T WRITE 00 AGAINST THE ITEMS FOR WHICH YOU DIDN'T WISH TO QUOTE ; INSTEAD, LEAVE THE COLUMN BLANK" AGAINST THE SAID ITEM; AS THE SYSTEM TAKES RS. 00.00 AS L1.**

**Important Note : Besides custom duty, the firm shall also mention Health cess and Social Welfare cess amount as applicable. Demurrage charges or late fee will not be paid by JKMSCL.**

**Note: The L1 of the Instruments shall be ascertained as per the cumulative rates on complete set basis. No individual (instrument) L1 rates shall be considered. The bidders have to quote minimum 95% of the instruments for qualification in the complete set.**

**Only the Rates reflected in the comparative sheet in the BOQ (as per format uploaded) shall be considered for ascertaining L1. No Separate rates quoted by the bidders in the BOQ shall be accepted.**

**Declaration and Undertaking by the Bidder**

**(On Non Judicial Stamp Paper worth Rs. 100/- Attested by Notary Public and submitted with Cover-'A')**

1. I/We..... (Name of firm) certify that the quoted model (of quoted item) is of latest technology and is not outdated.
2. I/We certify that the rates (of quoted item) are reasonable and not sold on lower rates to anyone than charged from JKMSCL.
3. I/We do hereby accept condition of warranty/guarantee period with spare parts of each quoted equipment as per terms & conditions and/or technical specifications. (From the date of installation/ demonstration).
4. (a) I/We do hereby undertake that our company/firm has not been black listed/banned/debarred/Convicted by Union Govt. or any State Govt. or their subordinate departments from participation in bidding.
- (b) I/We do hereby declare that our company/firm has been black listed/banned/debarred/convicted by..... (Name, Address of Govt./dept./State) and detailed information is as given below:
  - (i) Cause of blacklisting/banning/debarring/conviction.
  - (ii) For which item.....
  - (iii) Period of black listing/banning/debarring/ conviction.
  - (iv) Latest Status of black listing/banning/debarring/ conviction.
3. I/We hereby confirm that we have deposited all the GST/all applicable taxes as on date of submission of tender with the concerned authority/department. No GST/other taxes is due on the firm as on date of submission of tender.
4. I/we do hereby agree to the condition that JKMSCL may, if deemed fit go for the third party maintenance under Comprehensive equipment maintenance programme of Govt. of India.

**VERIFICATION & DECLARATION**

I/we.....S/o.....age d.....years residing at ..... authorized bidder/proprietor/ partner/director of firm M/s..... verify and confirm that the contents of bidding documents , its bidding forms, Annexure and other information submitted for bid no.

..... are true and correct to the best of my knowledge and nothing has been concealed therein.

In case, any variation/discrepancy/wrong declaration is found during scrutiny at later stages, I/We shall be held personally responsible & JKMSCL may take any action including blacklisting/debarring of my/our firm for a period not less than 05 years

Place :-  
Dated:-

Signature of the Deponent  
Name :  
Designation



## Annexure IV B

### **Declaration of Manufacturer/Direct Importer/Indian Subsidiary To be submitted on non judicial stamp paper of Rs. 100 duly notarised**

Date:\_\_\_\_\_ NIB No.:\_\_\_\_\_

I/We a legally constituted firm/body..... (Name of Firm/Company with address) and represented by Mr..... (Name of Bidder/Sole proprietor/ CMD/ Chairman) declare that I am/we are ..... (manufacturers/direct importer) in the goods and related services for which I/we have bid.

I/we further declare that:-

1. The items (Name of item) is/ are ..... (manufactured/imported) at our premises at ..... (Address of factory & office).
2. I/We..... (Name of firm) certify that the quoted model (of quoted item) is of latest technology and is not outdated.
3. I/We do hereby accept condition of guarantee period with spare parts of each quoted item as per terms & conditions or technical specifications. (From the date of installation/ demonstration).
  - a. Our company/firm has not been black listed/ banned/ debarred/convicted by Union Govt. or any State Govt. or their subordinate departments from participation in bidding.
  - b. Our company/firm has been black listed/banned/debarred/ convicted by ..... (Name, Address of Govt./dept./State) and detailed information is as given below:
    - (i) Cause of black listing/banning/debarring/ conviction.
    - (ii) For which item.....
    - (iii) Period of black listing/banning/debarring/ conviction.
    - (iv) Latest Status of black listing/banning/debarring/ conviction.
4. I/We hereby confirm that we have deposited all the GST/all applicable taxes up to the date of submission of tenders with the concerned authority/department. No GST/other taxes is due on the firm as on date.
5. We undertake that in case of change of dealership, we shall be responsible for providing preventive services and maintenance of the equipment free of cost during the warranty period.
6. We fully qualify the laid down terms & conditions of the NIB including Turnover class.

#### **VERIFICATION & DECLARATION**

I/we.....S/o.....aged..... years residing at ..... authorized bidder/proprietor/partner/director of firm M/s..... verify and confirm that the contents of bidding documents , its bidding forms Annexure I to Annexure VIII and other information submitted for bid no. .... are true and correct to the best of my knowledge and nothing has been concealed therein.

In case, any variation/discrepancy/wrong declaration is found during scrutiny at later stages, I/We shall be held personally responsible & JKMSCL may take any action including blacklisting/debarring of my/our firm for a period not less than 05 years

Place :-  
Dated:-

Signature of the Deponent  
Name :  
Designation



**Client Base (Item wise)**

**On letter Head of Bidder / Manufacture/Indian Subsidiary of Principal Manufacturer**

I/We..... (Name of firm.....) do hereby certify that our client base for the offered equipments are as under **{Client Base of the Bidder/Manufacturer/Indian Subsidiary of Principal Manufacturer with reference of the supply orders, for any of three years in last five years along with satisfactory performance certificate of minimum one installation. (Copies of reference supply orders and satisfactory performance certificate need to be attached)}** :-

| Item Code | Name of the Item | Client list | Reference to supply order |
|-----------|------------------|-------------|---------------------------|
|           |                  |             |                           |
|           |                  |             |                           |
|           |                  |             |                           |
|           |                  |             |                           |
|           |                  |             |                           |
|           |                  |             |                           |
|           |                  |             |                           |
|           |                  |             |                           |
|           |                  |             |                           |

1. It shall be submitted with technical bid and the above information should be verifiable from relevant documents of the bidder.
2. Firm should have market standing of the quoted product as mentioned above.
3. The different variants of same equipment may be considered.
4. In case of supply of imported item(s), the suppliers may be asked to furnish a certificate and other information to the effect that the firm has completed all the formalities including bill of entries in custom in connection with import of the item in question.

Place:

Date :

Signature of bidder with Seal.

**AUTHORISATION from principal manufacturer/importer/Indian Subsidiary**

*(On the letterhead of Principal manufacturer / Sole Importer/Indian Subsidiary)  
In case authorization to the bidder is furnished by the Sole Importer/Indian Subsidiary,  
document confirming authorization from foreign Principal Manufacturer in favour of  
Indian Subsidiary / Sole Importer is to be submitted.*

The Managing Director  
Jammu and Kashmir Medical Supplies Corporation Limited  
J&K

Subject: Regarding authorisation for our products.

Ref.: Your NIB no. ....dated.....

Name of items.....

Dear Sirs,

I/we.....(Name) for M/S.....(Name of firm) who are proven and reputable manufacturers .....(Name of item) having factory at ..... (Address of Factory and Office) hereby authorize M/S..... (Name of Bidder firm) to submit a bid, process the same further and enter into a contract with you against your requirement as contained in the above referred bid documents/NIB for the above goods manufactured / imported by us.

I/we further confirm that no supplier or firm or individual other than M/S..... (Name of bidder firm), is authorised to submit a Bid, process the same further and enter into a contract with you against your requirement as contained in the above referred bid documents for the above goods manufactured by us.

I/we also hereby extend our full guarantee, as applicable as per bid conditions of contract, read with modifications/addendum, if any, in the general/special conditions of contract for the goods and services offered for supply by the above firm against this bid document.

I/we also hereby confirm that we shall also be responsible for the satisfactory execution of contract placed on the authorized Firm. In case of default of authorised representative (or) otherwise, we shall also be responsible for the satisfactory execution of contract placed on the authorized Firm & penalty, if any, for non-execution of contract by the authorised representative shall be borne by us.

This authorization shall be valid till the completion of rate contract period and related services i.e. guarantee and comprehensive maintenance obligations, etc., whichever is later.

Yours faithfully,

(Name & Signature)..... verification and signature by bidder  
For M/s ..... Seal and address of bidder  
AUTHORISED SIGNATORY

Accepted by the authorized Bidder Mr.....  
(Signature, Name & Address).....

((On letter head of Chartered  
Accountant))

### ANNUAL TURN OVER STATEMENT

The average annual turnover of M/S..... (Name of Firm).....  
and address

..... for the past three years are given below and certified that the  
statement is true and correct:-

It is further certified that the Annual Turnover Statement has been prepared strictly as  
per returns filed with Taxation Department for the year 1<sup>st</sup> year, 2<sup>nd</sup> year & 3<sup>rd</sup> year and  
shall be responsible, if any variation/discrepancy is found during evaluation /later  
stage.

| Sl. No. | Financial Years      | Turnover in Lakhs (Rs.) |
|---------|----------------------|-------------------------|
| 1.      | 1 <sup>st</sup> year |                         |
| 2.      | 2 <sup>nd</sup> year |                         |
| 3.      | 3 <sup>rd</sup> year |                         |
| Total   |                      | - _____ Lakhs           |

Average gross annual turnover \_\_\_\_\_ Lakhs

Note :

1. To be prepared strictly as per returns filed with Taxation Department & the statement should be supported with returns filed for the last three financial years.
2. The turnover should be supported by the balance sheets of the respective years.
3. The Certificate issued by Taxation Department shall also be considered for turnover certification.
4. **The Average Annual Turn Over required for the item(s) pertaining to the Group "Procurement of Machinery & Equipment" is as per Table 1. Only the bid(s) falling under the category as specified under Annual Turnover is accepted. The bid(s) not falling under the Annual Turnover clause shall be out rightly rejected.**

Date

Signature of the bidder

Signature of Auditor/Seal  
Chartered Accountant  
(Name & Address.)  
Tel. No.  
UDIN NO.

(On Firm's letter head)  
**Memorandum of Appeal**

Appeal no..... of..... Before the.....  
(appellate authority)

1. Particulars of appellant:  
(i) Name of the appellant:  
(ii) Official address, if any:  
(iii) Residential address:
2. Name and address of the respondent(s):  
(i)  
(ii)  
(iii)
3. Number and date of the order appealed against and name and designation of the officer/ authority that passed the order (enclose copy), or a statement of a decision, action or omission of the procuring entity in contravention to the provisions of the Act by which the appellant is aggrieved:
4. If the appellant proposes to be represented by a representative, the name and postal address of the representative:
5. Number of affidavits and documents enclosed with the appeal:
6. Ground of appeal:  
.....  
.....  
..... (supported by an affidavit)
7. Prayer:.....  
.....  
.....

Demand Draft of Rs..... bearing No. ....dated  
..... as appeal fees

Place .....  
Dated .....

Appellant's signature

## Section V: Schedule of Supply

### Table of Contents

| S. No. | Description                        | Pages |
|--------|------------------------------------|-------|
| 1.     | List of goods and related services |       |
| 2.     | Delivery and completion schedule   |       |
| 3.     | Technical specifications           |       |
| 4.     | Drawings                           |       |
| 5.     | Inspections and tests              |       |

## Section V: Schedule of Supply

| Clause No. | Description   |
|------------|---|
| <b>1</b>   | <b>List of goods and related services</b>   |
| 1.1        | Name of item.....   |
| 1.2        | Related services are delivery, local transportation, installation, commissioning, demonstration and training etc.   |
| 1.3        | Guarantee/Warranty period starts from the date of successful installation for a period of Five years.   |
| 1.4        | JKMSCL may, if deemed fit, enter into third party agreement under comprehensive equipment maintenance programme, Govt. of India.  |
| <b>2</b>   | <b>Delivery and completion schedule</b>   |
| 2.1        | <b>SUPPLY ORDERS AND SUPPLY SCHEDULE:</b>   |
| 2.1.1      | Supply order shall be placed through registered post/e-mail/any communication medium by the JKMSCL. The date of receipt of e mail/fax/other communication shall be treated as the date of order for calculating the period of execution of order. The successful bidder shall execute the orders within a delivery period of 60 days or as specified in the supply order from the date of issuing supply order and handing over of space with the availability of power & other requisite installations by the end users. |
| 2.1.2      | In case of imported items, 30 days will be given in addition to above mentioned period, as mentioned in condition No. 2.1.1 above.  |
| 2.1.3      | The successful bidder shall acknowledge the receipt of orders, if any, within 7 days from the date of dispatch of order, failing which the procuring entity may be at liberty to initiate action to purchase the items on risk & cost purchase provision. However finalization of annual Rate contract does not mean mandatory issuance of supply order. Supply order shall be as per the requirements of items at various end-users.   |
| 2.1.4      | The Site of delivery shall be Drug ware House of JKMSCL or as per the requirement of the Department. The bidders can visit the site after seeking permission from the competent authority before quoting their rates.   |
| 2.1.5      | To ensure sustained supply without any interruption, the JKMSCL reserves the right to have more than one approved supplier from amongst the qualified bidders on L1 matched rates only. In such a case, the requirement may be met by dividing be quantity among the R/C holders considering the quantity required and dedicated capacity of the successful bidders.  |
| 2.1.6      | The ready stock position of the item, if provided by the firm, may be considered by the corporation for the placement of supply orders.   |
| 2.1.7      | It may be noted that the JKMSCL does not undertake to assist in the procurement of raw material, whether imported or controlled or restricted, and as such the bidders must offer their rates to supply the specific items from own quota of raw material stock by visualizing the prospect of availability and requirement. Any of the above points if taken,  |

|       |   |
|-------|---|
|       | as argument for non-supply/delayed supply will not be entertained.  |
| 2.1.8 | The figures indicated, if any, do not constitute any commitment on the part of JKMSCL to purchase any of the articles and the quantities shown therein against each or in any quantity whatsoever and no objection against the quantity of the indent of approved item being more or less than the indicative quantity will be entertained and shall not be acceptable as a ground for non supply of the quantity indented.   |
| 2.2   | <b>PROCURING ENTITY'S RIGHT TO VARY QUANTITY:</b>   |
| 2.2.1 | If the JKMSCL procures less than the quantity indicated in the bidding documents (if asked) the bidder shall not be entitled for any claim or compensation except otherwise provided in the conditions of contract.   |
| 2.2.2 | If the bidder fails to supply, the JKMSCL shall be free to arrange / procure the item(s) from other sources and the extra cost incurred shall be recovered from the supplier.   |
| 2.3   | <b>SUBMISSION OF CONTRACT COMPLETION REPORT</b>   |
| 2.3.1 | Firms shall have to submit consolidated statement in duplicate at the end of rate contract as well as after expiry of equipment / instrument guarantee period (as provided in guarantee clause of the contract) to enable the corporation to examine the case for refund of performance security.   |
| 2.3.2 | The consignee shall intimate the contract /supplier about the defect (s) at once in such a manner, so as to reach the office of the firm immediately and before completion of guarantee period. It shall be the responsibility of the consignee to get the complaint of guarantee period. It shall be the responsibility of the consignee to get the complaint of defective equipment of defective performance registered immediately with the office of JKMSCL.  |
| 2.5   | <b>PACKING &amp; INSURANCE:</b>   |
| 2.5.1 | The good shall be delivered at the destination in perfect condition. The firm if so desires may insure valuable goods against loss by theft, destruction or damages by fire, flood, under exposure to weather of otherwise in any situation. The insurance charges will have to be borne by the supplier and the corporation shall not be required to pay any such charges, if incurred.  |
| 2.5.2 | The firm shall be responsible for the proper packing so as to avoid damages under normal conditions of transport by sea, rail, road or air and delivery of material in good condition to the procurement officer's store. In the event of any loss, damage, breakage or leakage or any shortage the firm shall be liable to make good such loss and shortage found at destination after the checking/inspection of material by the consignee. No extra cost on such account shall be admissible. The firm may keep its representative to verify any damage or loss discovered at the consignee's store, if it so likes. |
| 2.5.3 | The material received with damaged packing (or) without packing as per terms & conditions of NIT (or) in damaged state, shall be liable to the minimum penalty of 2% of the value of the damaged item (or) quantity received with damaged packing. Further packing, cases, containers and other allied material if any shall be supplied free, except where   |

|            |   |
|------------|---|
|            | otherwise specified by the firm(s) and agreed by the corporation and the same shall not be returned to him.   |
| 2.5.4.     | <p><b>Packing specifications</b></p> <p>Schedule for packing – General specifications</p> <ol style="list-style-type: none"> <li>1. All items should be packed only in first hand boxes only.</li> <li>2. Label: Every box should carry a large outer label clearly indicated that the product is for <b>“JKMSCL Supply” for the year, “Not for Sale ”</b> and it should carry the correct technical name, strength or the product, date of manufacturing, date of expiry, quantity packed and net weight of the box in bold letters.</li> </ol> <p><b>Note: The weight/size of the box for packing the item may vary for the safe delivery/installation of equipment. Any deviation in the packing, if necessary shall be made after getting permission from JKMSCL.</b></p> |
| <b>2.6</b> | <b>REJECTION OF GOODS:</b>  |
| 2.6.1      | Articles not as per specification/ or not approved shall be rejected by the corporation / consignee and will have to be replaced by the supplier firm at its own cost within 15 days or with time limit fixed by the corporation.   |
| 2.6.2      | All the stores supplied shall be of the best quality and conforming to the specification, trademark laid down in the schedule attached to agreement and in strict accordance with and equal to the approved, standard/specifications/ samples. In case of any material of which there are no standards or approved samples, the supply shall be of the best quality to be substantiated by documents/specifications. The decision of JKMSCL as to the quality of stores is final and binding upon the bidder. In case any of the articles supplied are not found as per specification or declared sub-standard, that shall be liable to be rejected and any expenses of loss caused to the supplier as a result of rejection of supplies shall be entirely at his account.    |
| 2.6.3      | The rejected item must be removed by the firm, within 15 days of the date of intimation of rejection. The officials concerned shall take reasonable care of such material upto 15 days from the date of intimation only but in no case shall be responsible for any loss, damage, shortage that may occur while it is in their premises. In case firm fails to remove the items within fifteen days, JKMSCL shall have full right to get the said item(s) removed & destroyed at the cost & risk of supplier/bidder, without any further correspondence. The destroying charges as per the actual plus 1% penalty shall be deducted from any amount payable to the firm.  |
| 2.6.4      | No payment shall be made for defective/incorrect items.   |
| 2.6.5      | In case firm wants to take back item to their service station for rectification then firm has to deposit payment received against such defective supplies. In case supplier has not received any payment then material be returned to supplier firm for rectification. In no case the defective equipment is allowed to be installed after rectification.   |
| 2.6.7      | The bidder shall be responsible for the proper packing and delivery of the material to the consignee. In the event of any loss, damage, or breakage, leakage or shortage in transit, the bidder shall be responsible. No extra cost on such account shall be admissible.  |



|            |  |
|------------|--|
| <b>2.7</b> | <b>Payment Terms (For items quoted in foreign currency)</b>  |
| 2.7.1      | <p><b><u>For Payment through Letter of Credit (for imported items only)</u></b><br/> <b>100% payment shall be released against 20% Bank Guarantee valid for a period of 12 months, to be submitted by the bidder. The BG shall be released on successful installation of the Machinery. Letter of credit would be opened subject to following additional conditions:-</b></p> <ol style="list-style-type: none"> <li>1. At site LC would be opened.</li> <li>2. In case of supply through sea, LLOYD A level vessel would be used for shipment of supplies which should not be more than 15 years old.</li> <li>3. Supplies shall be insured by the OEM till the final delivery site shall also include "Force Majeure"</li> <li>4. Pre-dispatch inspection shall be carried out by OEM by certified inspection agency before shipment of supply.</li> <li>5. <b>The product shall be comprehensively insured upto site of installation for all type of insurance.</b></li> <li>6. <b>The charges of the L.C charged by the Govt. shall be borne by the firm.</b></li> </ol> <p><b><u>For Indian items :</u></b><br/> Payment shall be made after successful installation and commissioning of the equipment duly certified by Head of the concerned department.</p> |
| 2.7.2      | Payment shall be made by RTGS. Expenses on this account, if any, shall be borne by the firm.   |
| 2.7.3      | Payment to the authorised representative shall be made as per the tripartite agreement with the Corporation i.e JKMSCL on the basis of Annexure All to e uploaded along with e.bid.  |
| 2.7.4      | No advance payments towards cost of items shall be made to the bidder.   |
| 2.7.5      | If at any time during the period of contract, the price of bid items is reduced or brought down by any law or Act of the Central or State Government or by the bidder himself, the bidder shall be bound to inform Managing Director JKMSCL immediately about it. Purchasing authority shall be empowered to unilaterally effect such reduction as is necessary in rates in case the bidder fails to notify or fails to agree for such reduction of rates. In case this reduction of rates comes to the knowledge of JKMSCL in later stage, additional payment made w.e.f of the details of rates shall be charged from the firm with 1.5% monthly interest from the date/till rates have been reduced besides action as desired fit by JKMSCL which may be debarring/any other penalty as per penalty clause.   |
| 2.7.6      | In case of any enhancement in taxes/duty due to notification of the Government after the date of submission of bids and during the bid period, the quantum of additional taxes/duty so levied shall be allowed to be charged extra as a separate item without any change in the basic price structure of the items approved under the bid. For claiming the additional cost on account of the increase in tax/ duty, the bidder should produce a letter from the concerned authorities for having paid additional tax/duty on the goods supplied to ordering authority and also must claim the same in the invoice separately. Similarly if there is any reduction in the rate of taxes/duty of items, as notified by the  |

|            |   |
|------------|---|
|            | Government, after the date of submission of bid, the quantum of the price to the extent of reduction of taxes/duty of items will be deducted without any change in the basic price structure of the items approved under the bidder.  |
| 2.7.7      | In case of successful bidder has been enjoying exemption on any criteria, such bidder will not be allowed to claim taxes/duty at later point of time during the tenure of contract, if the taxes/ duty become chargeable on goods manufactured due to any reason.   |
| 2.7.8      | If there is any hindrance by the consignee to provide the required site for installation the part payment of equipment shall be made / decided by JKMSCL. In that case, the firm has to inform JKMSCL immediately.  |
| <b>2.8</b> | <b>LIQUIDATED DAMAGES:</b>  |
| 2.8.1      | The time specified for delivery in the bid form shall be deemed to be the essence of the contract and the successful bidder shall arrange supplies within the period on receipt of order from the purchasing officers.  |
| 2.8.2      | In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at the rate of 0.25% per day for every day of delay subject to maximum of 10%. Rest of the terms and conditions of SPP with regard to penalty clause shall remained unchanged<br>Penalty shall not be imposed if claim with regard to any supply i.e. Drugs/Equipment is complete in all respects i.e. QC verification/Board verified etc. is not cleared by the JKMSCL within a period of 60 days  |
| 2.8.3      | If the supplier requires an extension of time in completion of contractual supply on account of occurrence of any hindrances, he shall apply in writing to Managing Director JKMSCL, J&K, for the same immediately on occurrence of the hindrances but not after the stipulated date of completion of supply. The firms shall ensure extension of delivery period for delayed supplies. The payment shall only be released by corporation after sanction of extension in delivery period.   |
| 2.8.4      | Delivery period may be extended with or without liquidated damages. If the delay in the supply of goods is on account of force majeure i.e., which is beyond the control of the bidder, the extension in delivery period may be granted without liquidated damage.  |
| 2.8.5      | If the bidder is unable to complete the supply within the specified or extended period, the corporation shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the bidder on his (i.e., bidders) account at his cost and risk, with the prior approval of Managing Director JKMSCL, J&K. The bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the bidder.<br><br>The bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the bidder under this or any other contract with the corporation/government. If recovery is not possible from the bill and the bidder fails to pay the loss or damage within one month of the demand, the recovery of such amount or sum due from the bidder shall be made |

|            |   |
|------------|---|
|            | from the bidder. In case supplier fails to deliver ordered goods, the risk purchases may be made at market rate from any other firm. It is mandatory for the approved supplier to acknowledge receipt of orders within seven days from the date of dispatch of order, failing which the procuring entity will be at liberty to initiate action to purchase the items on risk purchase provision at the expiry of the prescribed supply period.  |
| 2.8.6      | LD for damaged packing or loose packing equivalent to 2 % of the value of the products received with damaged packing or in loose packing or with packing not conforming to the terms and conditions, specified in the tender document.  |
| <b>2.9</b> | <b>RECOVERIES:-</b>   |
| 2.9.1      | Recoveries of liquidated damages, short supplies, breakage, rejected articles shall ordinarily be made from bills. Such amount may also be recovered from any other untied dues & security deposits available with the JKMSCL. In case recovery is not possible, action will be taken as per prevailing Acts/rules in J&K State.  |
| 2.9.2      | Any recovery on account of liquidated damage charges/risk & cost charges in respect of previous rate contracts/supply orders placed on them by the JKMSCL can also be recovered from any sum accrued against this bid after accounting for untied sum or due payment lying with JKMSCL against previous rate contracts/supply orders. Firm shall submit details of pending amount lying with JKMSCL but decision of Managing Director JKMSCL, J&K regarding authenticity of sum payable shall be final. |
| 2.9.3      | <b>Testing &amp; Handling Charges : the testing and handling charges to the tune of 1.5% of total cost shall be deducted from the invoices raised by the approved bidder against the supply orders placed from time to time.</b>  |

### 3. Technical Specifications:

Annexure: A-III (technical specifications attached for Table I)

#### General features:

- i. **Bidders are requested to send printed descriptive literature/catalogue of the quoted items duly sealed by MD/Chairman/authorised signatory of the firm/bidder in the office of Jammu and Kashmir Medical Supplies Corporation Ltd. one day prior to last day of uploading of the bid. The catalogues along with compliance sheets should also be uploaded with the technical bid.**
- ii. If bidder supplied to or have rate contract of quoted items with any other Govt. institutions within one year, he may be asked to provide copies of purchase orders, invoices and rate contract.

#### 4) .Drawings if any to be attached with the technical bid.

#### 5. Inspection and Tests

| Clause No. | Description   |
|------------|---|
| <b>5.1</b> | <b>INSPECTION OF EQUIPMENTS AND INSTRUMENTS:-</b>   |
| 5.2        | The equipments supplies shall be according to technical specifications and shall be inspected by the committee constituted by JKMSCL as mentioned in the supply order or amended thereafter by competent authority. In case of BIS Items, inspection shall be strictly as per relevant BIS specifications with latest amendments and have been made applicable by B.I.S. at the time of inspection. The inspection and testing of the material may be got done by any inspecting Agency/team of experts at site of installation/commissioning. The supplier shall provide all facilities for inspection/testing free of cost.   |
| 5.3        | Notwithstanding the fact that the authorized inspecting team had inspected and/or has approved the stores/articles, any officer(s)/team of officer nominated by the corporation may inspect the item/material as soon as it is received in the stores to ensure that the supply is in accordance with the specifications laid down in rate contract/supply order.   |
| 5.4        | In case of doubts in inspection/ test, same may be got inspected or tested in any laboratory. If the material is not found as per specifications or defective, consignee shall not accept the material and shall inform the corporation within 3 days. Consignee may also simultaneously ask the firm for removal of defect/replacement. The firm shall be bound to replace the defective equipment/item within 15 days of receipt of intimation from the consignee/corporation. However, the date of delivery, in case of defective item shall be taken as the date on which the JKMSCL accepts the item after replacement of defective material/removal of defects as the case may be. Wherever defective item is replaced, the inspection/testing charges, if any, shall be borne by the supplier. |
| 5.5        | The corporation/technical expert or team shall match the specification with available reserved sample with the corporation which is submitted by the firm/supplier at the time of technical approval before release to end user. .  |
| 5.6        | In case of imported item, the supplier shall ensure that the item shall be inspected by the third party inspection agency before dispatched to the consignee. In case any un- inspected item has been found in the item received by consignee, the firm shall be solely responsible for it and the JKMSCL shall be free to take suitable necessary action as per terms and conditions of bid documents/agreement against the firm.  |

## Section VI A: - General Conditions of Contract (GCC)

### Table of Contents

| S. NO. | DESCRIPTION   |
|--------|---|
| 1.     | DEFINITIONS   |
| 2.     | GENERAL TERMS   |
| 3.     | BID SECURITY  |
| 4.     | FORFEITURE OF BID SECURITY  |
| 5.     | GAURANNTY CLAUSE  |
| 6.     | MARKING   |
| 7.     | APPLICABILITY OF RATES  |
| 8.     | COMPARISON OF RATES   |
| 9.     | SUBMISSION OF SAMPLES AND DEMONSTRATION                           |
| 10.    | PERFORMANCE SECURITY (P.S.) AND AGREEMENT                         |
| 11.    | SUPPLY ORDERS   |
| 12.    | SUBMISSION OF CONTRACT COMPLETION REPORT                          |
| 13.    | TERMS OF PAYMENT  |
| 14.    | LIQUIDATED DAMAGES  |
| 15.    | RECOVERIES  |
| 16.    | INSPECTION  |
| 17.    | PACKING & INSURANCE   |
| 18.    | REJECTION   |
| 19.    | CORRECTION OF ARITHMETIC ERRORS                                   |
| 20.    | PROCURING ENTITY'S RIGHT TO VARY QUANTITY                         |
| 21.    | PARALLEL RATE CONTRACT  |
| 22.    | VALIDITY OF BID   |
| 23.    | PRICE ESCALATION  |
| 24.    | SUBLETTING OF CONTRACT  |
| 25.    | FALL CLAUSE   |
| 26.    | GRIEVANCE REDRESSAL DURING PROCUREMENT PROCESS                    |
| 27.    | COMPLIANCE WITH THE CODE OF INTEGRITY AND NO CONFLICT OF INTEREST |
| 28.    | DISPUTE SETTLEMENT MECHANISM                                      |
| 29.    | OTHER CLAUSES   |
| 31.    | JURISDICTION  |

## SECTION VI A: - GENERAL CONDITIONS OF CONTRACT (GCC)

Bidder should read these terms & conditions carefully and comply strictly while submitting their bids. If a bidder has any doubt regarding the terms & conditions and specifications mentioned in the bid notice/ catalogue, he should refer these to the Jammu and Kashmir Medical Supplies Corporation, J&K, before submitting bids and obtains clarifications. The decision of the Managing Director Jammu and Kashmir Medical Supplies Corporation, J&K, shall be final and binding on the bidder. The clauses of terms & conditions are as follows:-

| Clause No. | Description   |
|------------|---|
| 1.         | <b>Definitions</b>  |
|            | <p>The following words and expressions shall have the meanings hereby assigned to them:</p> <p><b>'Act/Rules'</b> means Acts &amp; rules prevailing in J&amp;K Union Territory in terms of procurement.</p> <p><b>'Completion'</b> Means the fulfilment of the supplies and Related Services by the supplier in accordance with the terms and conditions set forth in the contract.</p> <p><b>"Contract"</b> Means the Agreement entered into between the procuring entity and supplier, together with the contract documents referred to therein, including all attachments, appendices, specifications and codes and all documents incorporated by reference therein.</p> <p><b>"Contract Documents"</b> Means the documents listed in the agreement, including any amendments thereto.</p> <p><b>"Contract Price/Rate"</b> Means the price payable to the supplier as specified in the agreement, subject to such additions and adjustments thereto or deductions there from, as may be made pursuant to the contract.</p> <p><b>"Day"</b> Means calendar day.</p> <p><b>"Delivery"</b> Means the transfer of the goods from the supplier to the procuring entity in accordance with the terms and conditions set forth in the contract.</p> <p><b>"GCC"</b> Means the general conditions of rate contract.</p> <p><b>"SCC"</b> Means the special conditions of rate contract".</p> <p><b>"Goods"</b> Means all of the commodities, raw material, machinery and equipment, documents, warranties and /or other materials that the supplier is required to supply to the Procuring Entity under the Contract.</p> <p><b>"Procuring Entity"</b> Means the entity purchasing the goods and related services, Managing Director Jammu and Kashmir Medical Supplies Corporation, J&amp;K, or as specified in the special conditions of the contract (SCC).</p> <p><b>"Related Services"</b> Means the services incidental to the supply of the goods, such insurance, installation, training and initial maintenance, commissioning of equipment or machinery and other similar obligations of the supplier under the contract. <b>"Subcontractor"</b> Means any natural person, private or government entity, or a combination of the above, including its legal successors or permitted assigns, to whom any part of the goods to be supplied is subcontracted by the supplier.</p> |

|           |   |
|-----------|---|
|           | <p><b>"Supplier"</b> Means the natural person, private or government entity, or a combination of the above, whose bid to perform the contract has been accepted by the procuring entity and is named as such in the agreement, and includes the legal successors or permitted assigns of the supplier.</p> <p><b>Authorised representative :</b> Means the natural person, proprietor or Govt entity, duly authorised by the Managing Director/Prop/Chairman/Board of Director of original manufacturer/direct importer under their seal signatures duly notarized ; to bid, negotiate, raise the invoice, receive the payment against the supplies made, enter into tripartite agreement within the Corporation i.e JKMSCL, inter-alia.</p> <p><b>Authorised signatory :</b> Means the natural person authorised by the proprietor, Managing Director/Chairman/Board of Director of original manufacturer/direct importer under their seal signatures duly notarized to sign on behalf of the company.</p> <p><b>"The Site"</b> where applicable, means the place of delivery, installation, testing/ commissioning of the goods /equipment or machinery or as mentioned in the supply order.</p> <p><b>"Consignee"</b> Means the receiver of the stores as mentioned in supply order.</p> |
| <b>2.</b> | <b>General terms</b>  |
| 2.1       | Bids are invited from original manufacturers /direct importers/authorized representative of the original manufacturer/direct importer.  |
| 2.2       | Bid shall have to uploaded as per schedule, to JK e-portal : <a href="http://www.jktenders.gov.in">www.jktenders.gov.in</a> . At any time prior to the date of uploading of bid, bid inviting authority may, for any reason, whether on his own initiative or in response to a clarification requested by a prospective bidder, modify the condition in bid document by an amendment. In order to provide reasonable time to take the amendment into account in preparing their bid, bid inviting authority may at his discretion, extend the date and time for submission of bid. Interested eligible bidders may obtain further information in this regard from the office of the bid inviting authority.   |
| 2.3       | Supplies shall be made directly by the bidder to be called as "Supplier" after finalization of rate contract, and suppliers. Manufacturer bidder should have permission to manufacture the item quoted as per specification given in the bid from the competent authority.  |
| 2.4.1     | Direct importer should authenticate import/sale license for the product quoted in the bid issued by the competent authority.  |
| 2.4.2     | In case, the item/product is supplied through authorised representative, product manufacturing permission, import/sale license of the principal manufacturer (s) direct importer (s) shall have to be uploaded along with technical bid.  |
| 2.5       | Bid shall be have to be loaded on e-portal i.e <a href="http://www.jktenders.gov.in">www.jktenders.gov.in</a> submitted to Managing Director, Jammu and Kashmir Medical Supplies Corporation, J&K   |




|     |  |
|-----|--|
| 2.6 | <p>The bidder shall also submit the following documents and certificates along with the bid as per technical bid submission letter :-</p> <p>(i) A combined undertaking/declaration regarding that the quoted item :</p> <ul style="list-style-type: none"> <li>a. Model is of latest technology, the item has not become outdated, that the rate quoted is not more than the rate charged from anyone else,</li> <li>b. That the bidder is not black listed or banned or debarred by central or any state government or its append gages,</li> <li>c. Availability of spare parts and consumables for the quoted equipment for at least 10 years/life of the item, from the date of installation.</li> </ul> <p>Note : Bid should not be submitted for the quoted item(s) for which the bidder has been blacklisted/banned/debarred either by bid inviting authority or Govt. of J&amp;K or by any other State/Central Govt. and its agencies. This also applies to the bidder for its sister/ allied firm(s)/ unit(s).</p> <p>(ii) <b>The bidder, in case of representative of the manufacturer/direct importer shall submit fresh authorization of the manufacturer/direct importer duly authenticated and notarized.</b></p> <p><b>PLEASE ALSO NOTE THAT: -</b></p> <p>(A) All attested documents must be submitted in English language. If the documents are not in English, translated version of the same, in English, duly signed and attested by authorized translator must be submitted along with copy of original document.</p> <p>(B) All the above mentioned documents should be under the name and address of the premises where the quoted items are actually manufactured/ stored for supply.</p> |
| 2.7 | <p>Financial Bid duly filled in (Annexure III/BOQ) giving the rates for quoted items should be uploaded through e portal <a href="http://www.jktenders.gov.in">www.jktenders.gov.in</a>. The rate should not be disclosed/uploaded in the technical bid. <b>Rates uploaded along with technical bid shall means out rightly rejection of bid of the concerned person.</b></p>  |
| 2.8 | <p>The required amounts towards cost of bid document and tender processing charges shall be deposited as mentioned at page 5, 01 day before the last date and time of bid submission.</p> <p>All bids received will be opened in the presence of bidders, who choose to be present. Financial bid will be opened only for those bidders, who satisfy the criteria laid down by the JKMSCL on the details furnished by the bidder in technical bid in compliance of terms &amp; conditions of the bid.</p>  |
| 2.9 | <p>(i) In case of the bid being submitted by a proprietary firm, the bid must be signed by the sole proprietor. In case of a partnership firm, bid must be signed on behalf of the firm by a person authorized, holding a</p>  |



|          |  |
|----------|--|
|          | <p>power of attorney in his favour to do so; and in the case of a company, the bid must be signed by an authorized signatory, in the manner laid down in the articles of association of the bidder company.</p> <p>(ii) Any change in the constitution of the firm/ company shall be notified forthwith by the bidder/contractor in writing to the Jammu and Kashmir Medical Supplies Corporation, J&amp;K and such change shall not relieve any former member of the firm/ company from the liability under the conditions of the bid/contract. No new partner / partners shall be accepted in the firm by the bidder/contractor in respect of the bid/contract unless he/ they agree to abide by all its terms and conditions and submit a written agreement to this effect. The bidder's/contractor's receipt for acknowledgement or date of any new partner subsequently inducted, as above, shall bind all of them and will be a sufficient discharge for any of the purposes of the contract.</p>  |
| <b>3</b> | <b>BID SECURITY:</b>   |
|          | <p>(i) Bid shall have to be accompanied with a scanned copy of FDR/CDR/BG/NEFT/RTGS as bid security. The bid security shall have to be submitted before the opening of technical bid with a validity of 30 months. Bids submitted without sufficient bid security &amp; validity shall be summarily rejected.</p> <p>(ii) The bid security of bidder shall be refunded after the earliest of the following events, namely:-</p> <ol style="list-style-type: none"> <li>the expiry of validity of bid security;</li> <li>the cancellation of the procurement process; or</li> <li>the withdrawal of bid prior to the deadline for presenting bids, unless the bidding documents stipulate that no such withdrawal is permitted.</li> </ol> <p>(iii) The bid security lying with the JKMSCL in respect of other bids awaiting approval or rejection or on account of contracts being completed, shall not be adjusted towards bid security for the fresh bids. The bid security may, however, be taken into consideration in case bids are re-invited for the same item.</p> <p>(vi) In case any document submitted by the bidder or by his authorized representative is found to be forged, false or fabricated, the bid shall be rejected and bid security may be forfeited. Bidder/his representative may also be banned / debarred. Report with police station may also be filed against such bidder/his representative.</p> |
| <b>4</b> | <b>FORFEITURE OF BID SECURITY: -</b>   |
|          | <p>The bid security shall be forfeited if:</p> <ol style="list-style-type: none"> <li>The bidder withdraws or modifies the offer after opening of financial bid, but before acceptance of bid,</li> <li>The bidder does not execute the agreement, if any, prescribed within the specified time or extended time by competent authority (on the request of the bidder),</li> <li>The bidder does not deposit the 'performance security' after the supply order is placed/requested for signing the agreement,</li> <li>The bidder fails to commence the supply of the items as per supply</li> </ol>   |

|          |   |
|----------|---|
|          | <p>order within the time prescribed,</p> <p>(v) The bidder fails to submit samples/demonstration of quoted item on demand</p> <p>(vi) The bidder violates any of the terms &amp; conditions of the bid document.</p>  |
| <b>5</b> | <b>WARRANTY CLAUSE:-</b>  |
|          | <p>(i) The bidder would guarantee that the subject matter of procurement would continue to conform to the description and quality as per technical specifications and performs as per descriptions, from the date of delivery/ installation of the said subject matter of procurement. Notwithstanding the fact that the purchaser may have inspected and/or approved the said subject matter of procurement during the guarantee period, if the said subject matter of procurement is discovered not to conform to the description and quality as aforesaid or not performing, as described, the procuring entity will be entitled to reject the said subject matter of procurement or such portion thereof as may be discovered not to conform to the said description and quality or not performing as described. On such rejection, the subject matter of procurement will be at the seller's risk and all the provisions relating to rejection of goods, etc., shall apply. The successful bidder shall, if called upon to do so, replace the goods etc. or such portion thereof, as rejected by the procuring entity. Otherwise, the bidder shall pay such damages, as may arise by reason of such breach of the condition herein contained. Nothing herein contained shall prejudice any other right of the procuring entity in that behalf under this contract or otherwise.</p> <p>(ii) The bidder shall, during the Guarantee period appearing in the contract, replace the whole subject matter of procurement or part(s), if any, and remove the manufacturing defects, if found during the above period so as to make the machinery and equipment/ordered items operative.</p> <p>(iii) In case of the machinery or equipment/ordered items, the successful bidder shall be responsible for carrying out annual maintenance and repairs on the terms &amp; conditions, as agreed. The bidder shall have to ensure that consumables required for the maintenance of machine/equipment are being supplied free of cost during warranty period. The adequate regular supply of spare parts and consumables per incident for the machinery or equipment, whether under their annual maintenance and repairs contract or otherwise shall be ensured. In case of change of model the bidder shall notify the procuring entity sufficiently in advance, to facilitate procurement of sufficient quantity of consumables/ spare parts from the bidder to maintain the machinery or equipment. In case, any item supplied by the successful bidder does not conform to the required specifications, the payment thereof, if received by the supplier, shall have to be refunded to the Jammu and Kashmir Medical Supplies Corporation, J&amp;K along with interest to the tune of</p> |

|          |  |
|----------|--|
|          | 1.5% per month from the date of release of payment. The supplier will not have any rightful claim to the payment of cost for substandard supplies, which may have been consumed, either in part or whole, pending receipt of laboratory test / inspection report, wherever required. Supply of goods less in weight and volume than those mentioned on the label of the container, the same will be dealt with in the manner prescribed under rules.   |
| <b>6</b> | <b>MARKING</b>   |
|          | <p>All items and accessories supplied should bear marking "JKMSCL SUPPLY _____(engraved or non removable material) "NOT FOR SALE" or as mentioned in supply order in English, without which the supply will not be entertained. JKMSCL may ask change in art work to be printed on the item at any stage of the contract.</p> <div style="text-align: center;">  </div> <p style="text-align: center;">JKMSCL SUPPLY ( _____ ) NOT FOR SALE</p>   |
| <b>7</b> | <b>COMPARISON OF RATES:</b>  |
|          | <ul style="list-style-type: none"> <li>(i) Only net rates should be quoted. No separate free goods or cash discounts should be offered. Rates must be valid for the entire period of contract.</li> <li>(ii) Consignee may be located at a district headquarter (except equipment/ machinery requiring installation and commissioning, the place may be any other station) or as directed by Jammu and Kashmir Medical Supplies Corporation Limited, J&amp;K and the rates must be quoted accordingly. No cartage or transportation charges shall be payable.</li> <li>(iv) The net rate must be inclusive of all charges by way of packing, forwarding, incidental or transit charges, including transit insurance, and any other levies or duties etc. on the subject matter of procurement.</li> <li>(v) In the event of any subsequent variation (increase or decrease) in the rate of GST or any other taxes by the government (state /UT or central), the same will be admissible accordingly.</li> <li>(vi) If the rates of item quoted are found same from two or more bidders, then the bidders shall be asked to submit revised financial bid, containing reduced rates within given time by Managing Director, Jammu and Kashmir Medical Supplies Corporation Limited, J&amp;K.</li> <li>(vii) The bidder will exercise all due diligence at their own level regarding applicability of other taxes, duties and fees etc. for the unit of supplies as specified in the bid document and accordingly include the same in their quotes. Any additional/extra claims over</li> </ul> |

|           |  |
|-----------|--|
|           | <p>and above the rates agreed pertaining to taxes, duties and fees etc. will not be entertained later on any account.</p> <p>(viii) No part of the bid document should be detached / deleted.</p>  |
| <b>8</b>  | <b>SUBMISSION OF SAMPLES/CATALOGUES AND DEMONSTRATION</b>  |
|           | <p>(i) Catalogues/samples of the quoted item(s) must be sent free of cost to JKMSCL even though the specifications or description etc. are mentioned in the bid form are complied.</p> <p>(ii) Samples of items(s) should be collected back from the JKMSCL, J&amp;K within 15 days from the date of finalization of list of successful bidder/demonstration of product before the expert panel. The corporation shall not be responsible for any damage, wear and tear or loss during the course of testing / examination, etc. The corporation may retain the sample of approved item for one month beyond expiry of contract. The corporation shall not be responsible for any damage, wear and tear or loss in this period. The corporation shall not make any arrangement for return of samples even if the bidder agrees to pay the cost of transportation.</p> <p>(iii) The bidder may be asked to demonstrate the technique, procedure and utility of item as per specifications given in the bid document before the technical committee constituted by the Corporation for the purpose. In case of heavy equipment, the demonstration may be carried out at the nearby place where the equipment has been installed by the bidder. In that case, the decision of the technical committee shall be final. The firm shall keep ready the quoted item and arrange all logistics within the time frame as and when asked by the JKMSCL. After the due date, no request of the bidder/firm shall be entertained for demonstration.</p> <p>(iv) Sample should be strictly according to the item quoted in the bid form failing which the bid will not be considered. Sample must be submitted duly sealed and marked suitably either by writing on the sample or on a slip or durable paper securely fastened to the sample with the particulars as mentioned below:</p> <ol style="list-style-type: none"> <li>Name and full address of the firm</li> <li>Catalogue no. and name of the item</li> <li>Name of section</li> <li>Name of manufacturer</li> <li>Brand</li> </ol> <p>(v) No change in marking on sample will be allowed after the submission of the sample.</p> |
| <b>10</b> | <b>PERFORMANCE SECURITY (P.S.) AND AGREEMENT:</b>  |
|           | <p>(i) The successful bidder shall submit the original copy of Bid document signed on each page <b>at the time of agreement</b>. However, while uploading the technical bid, only the declaration regarding acceptance of terms &amp; conditions shall be uploaded.</p>  |

- (ii) The period of rate contract shall be 24 months from the date of issuance of rate contract. The Managing Director, JKMSCL can extend the original rate contract subject to original terms and conditions for a period deemed fit by them, but not exceeding three months, for which the bidder shall abide.
- (iii) Successful bidders, whose offers are accepted, shall have to deposit performance security @ 3% of the value of the supply order in favour of Chief Accounts Officer, JKMSCL within 15 days from the date of issuance of letter of intent. The performance security shall be deposited in the form of FDR/CDR/B.G (Bank Guarantee)/NEFT/RTGS. However, the bank guarantee shall be for a validity period of six months, beyond the guarantee period sought for the item.
- (iv) In case of successful bidder(s), the amount of bid security shall be adjusted for performance security for the supply order placed to the firms/bidders. The amount of performance security, if exceeds the bid security, it shall be deposited by the firm against the supply orders issued from time to time.
- (v) The firm may submit bank guarantee issued by any scheduled/nationalised bank. The minimum validity of bank guarantee should be six months after completion of guarantee period for the item.
- (vi) The Performance Security: The Performance Security (P.S.) shall be 3 % of the total value of stores ordered for supply. The payment shall not be released against supplies until the additional Performance Security due is deposited by the supplier or additional.
- (vii) The performance security shall be refunded after six months after satisfactory completion of contract and after satisfying that there are no dues outstanding against the bidder subject to guarantee provisions.
- (viii) It is to be noted that earlier year's bid security and performance security, even if lying in the JKMSCL shall not be considered towards this contract and therefore fresh bid security/performance security shall be deposited. The JKMSCL shall pay no interest on bid security or performance security amount.
- (ix) Successful bidders shall have to execute an agreement on a Non-Judicial stamp paper of an amount mentioned in the offer letter, in the prescribed form with the JKMSCL and deposit performance security within 15 days from the date of acceptance of the bid is communicated to him. However, Managing Director JKMSCL, J&K may condone the delay in execution of contract by the bidder. The expenses in this regard shall be borne by the successful bidder. The validity of contract under this agreement shall be for a period as mentioned.
- (x) The bidder shall furnish the following documents at the time of

|           |  |
|-----------|--|
|           | <p>execution of agreement:-</p> <ul style="list-style-type: none"> <li>(i) Attested copy of partnership deed in case of partnership firms.</li> <li>(ii) Registration number and year of registration, in case partnership firm is registered with registrar of firms;</li> <li>(xi) Address of residence and office, telephone numbers, in case of sole proprietorship with : <ul style="list-style-type: none"> <li>(i) Registration issued by registrar of companies under Registrar of companies Act 1956, in case of company.</li> </ul> </li> <li>(xiv) In case of breach of any terms and conditions of the contract or on unsatisfactory performance, the amount of performance security shall be liable to forfeiture by JKMSCL, J&amp;K and decision of Managing Director JKMSCL J&amp;K shall be final.</li> <li>(xv) The rate contract can be repudiate/rejected at any time by the Managing Director JKMSCL, J&amp;K if the supplies are not made to his satisfaction after giving an opportunity to the bidder of being heard and after reasons for repudiation being recorded by him in writing. However, Managing Director JKMSCL, J&amp;K may terminate the agreement of contract at any time without notice/intimation to the successful bidder.</li> </ul>  |
| <b>11</b> | <b>SUPPLY ORDERS:</b>  |
|           | <ul style="list-style-type: none"> <li>(i) Supply order shall be placed through registered post/e-mail/any communication medium by the JKMSCL. The date of receipt of letter of communication date will be treated as the date of order for calculating the period of execution of order. The successful bidder will execute the orders within a period of 60 days or as specified in the supply order.</li> <li>(ii) The successful bidder acknowledge receipt of orders within 7 days from the date of dispatch of order, failing which the procuring entity may be at liberty to initiate action to purchase the items on risk &amp; cost purchase provision.</li> <li>(iii) In case of imported items, 30 days shall be given in addition to above mentioned period,</li> <li>(iv) Except, for equipments / machinery, which requires installation / commissioning, all other supplies shall have to be to FOR district drug warehouse only. In case of non-viable size of order for supplies, the corporation shall take appropriate decision on representation from the supplier on case to case basis. The consignee for supplies shall be JKMSCL.</li> <li>(v) To ensure sustained supply without any interruption, the Managing Director, JKMSCL reserves the right to have more than one approved supplier from amongst the qualified bidders as matched L1 supplied at matched L1 rates. In such a case, the requirement may be met by dividing the quantity among the rate contract holders considering the quantity required and dedicated capacity of the successful bidders.</li> <li>(vi) The ready stock position of the item, if provided by the firm, may be considered by the corporation for the placement of supply orders.</li> </ul> |



|           |   |
|-----------|---|
|           | (vi) It may be noted that the JKMSCL does not undertake to assist in the procurement of raw material, whether imported or controlled or restricted, and as such the bidders must offer their rates to supply the specific items from own quota of raw material stock by visualizing the prospect of availability and requirement. Any of the above points if taken, as argument for non-supply/delayed supply will not be entertained.  |
| <b>12</b> | <b>SUBMISSION OF CONTRACT COMPLETION REPORT</b>   |
| 12.1      | Firms shall have to submit consolidated statement in duplicate at the end of rate contract well as after expiry of equipment / instrument guarantee period (as provided in guarantee clause of the contract) to enable JKMSCL to examine the case for refund of performance security.   |
| 12.2      | The end user shall intimate the complaint/defect arise immediately to the manufacturer/importer/representative with copy to JKMSCL for further follow up..  |
| <b>13</b> | <b>LIQUIDATED DAMAGES:</b>  |
|           | <p>I. The time specified for delivery in the tender form shall be deemed to be the essence of the contract and the successful Bidder shall arrange supplies within the period on receipt of order from the Purchasing Officers.</p> <p>II. In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at the rate of 0.25% per day for every day of delay subject to maximum of 10%. Delay beyond 120 (for Indian products) and 150 days (for imported products) shall be treated as unexecuted and attract penalty @20%.</p> <p>III. Penalty shall not be imposed if claim with regard to any supply i.e. Drugs/Equipment is complete in all respects i.e. QC verification/Board verified etc. is not cleared by the JKMSCL within a period of 60 days.</p> <p>IV. Fraction of a day in reckoning the period of delay in supplies shall be eliminated if it is less than half a day.</p> <p>V. The maximum amount of agreed liquidated damage shall be 20%.</p> <p>VI. If the supplier requires an extension of time in completion of contractual supply on account of occurrence of any hindrance, he shall apply in writing to M.D, JKMSCL, Jammu / Srinagar (J&amp;K), which has placed the supply order, for the same immediately on occurrence of the hindrance but not after the stipulated date of completion of supply. The firms shall ensure extension of delivery period for delayed supplies. The payment shall only released by purchase officer after sanction of extension in delivery period by M.D., JKMSCL.</p> <p>VII. Delivery period may be extended with or without liquidated damages. If the delay in the supply of goods is on account of hindrances beyond the control of the Bidder, the extension in delivery period may be granted</p> |

|       |   |
|-------|---|
|       | without Liquidated Damage.  |
| VIII. | <p>If the Bidder is unable to complete the supply within the specified or extended period, the purchasing officer (JKMSCL) shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the Bidder on his (i.e. Bidders) account and risk only with the prior approval from M.D., JKMSCL, Jammu / Srinagar (J&amp;K). The Bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the Bidder. The Bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the Bidder under this or any other contract with the government. If recovery is not possible from the bill and the Bidder fails to pay the loss or damage, within one month of the demand, the recovery of such amount or sum due from the Bidder shall be made under the law for the time being in force. In case more than one supplier has been approved for any item under the approved list circulated to the purchasing officers, the risk purchases may be made at a higher rate from any other firm whose rate is duly approved. It is mandatory for the approved supplier to acknowledge receipt of orders with in fifteen days from the date of dispatch of order, failing which the purchasing officer will be at liberty to initiate action to purchase the items on risk purchase system at the expiry of the prescribed supply period, after taking required approval from M.D., JKMSCL (J&amp;K).</p> |
| IX.   | <p>If the bidder is unable to complete the supply within the specified or extended period, the purchasing officer shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the bidder on his (i.e., bidders) account at his cost and risk, with the prior approval from JKMSCL. The bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the bidder. The bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the bidder under this or any other contract with the corporation/government. If recovery is not possible from the bill and the bidder fails to pay the loss or damage within one month of the demand, the recovery of such amount or sum due from the bidder shall be made or any other law for the time being in force. In case supplier fails to deliver ordered goods, the risk purchases may be made at market rate from any other firm. It is mandatory for the approved supplier to acknowledge receipt of orders within seven days from the date of dispatch of order, failing which the procuring entity will be at liberty to initiate action to purchase the items on risk purchase provision at the expiry of the prescribed supply period.</p>   |
| X.    | <p>In case of wrong quoting, (or) if successful bidder refuses (or) fails to execute the supplies on the basis of wrong quoting of rates, the bidder shall be penalized with forfeiting of amount equivalent to the Performance security for the said product (or) debaring/ blacklisting of firm for that particular product(s) for a period not less than 02 years (or) both as deemed fit by TIA i.e. MD, JKMSCL.</p>  |
| 14    | (i) JKMSCL shall procure the machinery & equipment for the Health & Medical Education Institutes of UT of J&K inter-alia.   |



|           |  |
|-----------|--|
|           | (ii) The funds shall be transferred to JKMSCL with indent form and supply orders shall be placed by JKMSCL to suppliers.   |
| <b>15</b> | <b>RECOVERIES</b>  |
|           | <p>(i) Recoveries of liquidated damages, short supplies, breakage, rejected articles shall ordinary be made from bills. Such amount may also be recovered from any other untied dues &amp; security deposits available with Corporation. In case recovery is not possible, recourse will be taken under law in force.</p> <p>(ii) <b>Any recovery on account of L.D. charges/risk &amp; cost charges in respect of previous rate contracts/ supply orders placed on them by the corporation can also be recovered from any sum accrued against this tender after accounting for untied sum or due payment sum lying with corporation against previous rate contracts/supply orders. Firm shall submit details of pending amount lying with corporation but decision of M.D., JKMSCL, J&amp;K regarding authenticity of sum payable shall be final.</b></p>   |
| <b>16</b> | <b>INSPECTION:-</b>  |
|           | <p>(i) The equipments supplied shall be according to specifications provided at Section IV (3) schedule of supply and may be inspected by the technical panel/team constituted for the purpose by JKMSCL deemed fit on the site of manufacturer (in case of Indian manufacturer)/ importer (importer site). The manufacturer/importer shall facilitate the demonstration of the said machine/equipment/on the site only. After the receipt of "Certificate of satisfaction" from the technical panel, the supply order shall placed. In case of BIS Items, inspection shall be strictly as per relevant BIS specifications with latest amendments and have been made applicable by B.I.S. at the time of inspection. The machine/equipment shall be further inspected at the time of installation/commissioning at site i.e the end user site. The supplier shall provide all facilities for inspection/testing free of cost.</p> <p>(ii) Notwithstanding the fact that the authorized inspecting agency had inspected and/or has approved the stores/articles, the procurement officer or his representative may inspect the item/material as soon as it is received in the stores to ensure that the supply is in accordance with the specifications laid down in rate contract.</p> <p>(iii) In case of doubts in inspection/ test, same may be got inspected or tested in any laboratory. If the material is not found as per specifications or defective, consignee will not accept the material and shall inform the JKMSCL, J&amp;K within 3 days. Consignee may also simultaneously ask the firm for removal of defect/replacement. The firm shall be bound to remove the defect or replace the defective equipment/item within 15 days of receipt of intimation from the consignee. However, the date of delivery, in case of defective item shall be taken as the date on which the JKMSCL accepts the item after replacement of defective material/removal of defects as the case may be. Wherever defective item is replaced, the inspection /</p> |

|    |  |
|----|--|
|    | <p>testing charges, if any, shall be borne by the supplier.</p> <p>(iv) If required, the consignee may refer inspection committee to match the specification with available reserved sample with the corporation which is submitted by the firm/supplier at the time of technical approval.</p> <p>(v) In case of imported item, the supplier shall ensure that the item shall be inspected by the third party inspection agency before dispatched to the consignee. In case any un-inspected item has been found in the item received by consignee, the firm shall be solely responsible for it and the JKMSCL shall be free to take suitable necessary action as per terms and conditions of bid documents/agreement against the firm.</p>   |
| 17 | <b>PACKING AND INSURANCE</b>   |
|    | <p>(i) The goods will be delivered at the destination in perfect condition. The firm if so desires may insure valuable goods against loss by theft, destruction or damages by fire, flood, under exposure to weather or otherwise in any situation. The insurance charges will have to be borne by the supplier and the corporation shall not be required to pay any such charges, if incurred.</p> <p>(ii) The firm shall be responsible for the proper packing so as to avoid damages under normal conditions of transport by sea, rail, road or air and delivery of material in good condition to the procurement officer's store. In the event of any loss, damage, breakage or leakage or any shortage the firm shall be liable to make good such loss and shortage found at destination after the checking/inspection of material by the consignee. No extra cost on such account shall be admissible. The firm may keep its representative to verify any damage or loss discovered at the consignee's store, if it so likes.</p> <p>(iii) Packing, cases, containers and other allied material if any shall be supplied free, except where otherwise specified by the firm(s) and agreed by the JKMSCL and the same shall not be returned to him.</p> |
| 18 | <b>REJECTION</b>   |
|    | <p>(i) Articles not as per specifications/or not approved shall be rejected by the JKMSCL and will have to be replaced by the supplier firm at his own cost within 15 days or as time limit fixed by the JKMSCL.</p> <p>(ii) All the stores supplied shall be of the best quality and conforming to the specification, trademark laid down in the schedule attached to agreement and in strict accordance with and equal to the approved, standard, samples. In case of any material of which there are no standards or approved samples, the supply shall be of the best quality to be substantiated by documents. The decision of Managing Director JKMSCL as to the quality of stores be final and binding upon the bidder. In case any of the articles supplied are not found as per specification or declared sub-standard/spurious, that shall be liable to be rejected and any expenses of loss caused to the supplier as a result of rejection of supplies shall be entirely at his account.</p> <p>(iii) The rejected item must be removed by the firm, within 15 days of the date of intimation of rejection. The officials concerned will take reasonable care</p>  |

|            |   |
|------------|---|
|            | <p>of such material but in no case shall be responsible for any loss, damage, shortage that may occur while it is in their premises.</p> <p>(iv) No payment shall be made for defective/incorrect items. However, if payment has been made, then defective items shall be allowed to be removed only after the firm replaces material as per specifications, duly inspected. If the payment has not been made, the firm may be allowed to remove the material without prior replacement (provided firm has performance security as per condition No. 18). Joint inspection of defective material may be carried out as required by the JKMSCL. However sample of ISI marked material found defective shall be kept by consignee for reference to BIS.</p> <p>(v) In case firm wants to take back item to their works for rectification then firm has to deposit payment received against such defective supplies. In case supplier has not received any payment then material be returned to supplier firm for rectification.</p> <p>The Bidder shall be responsible for the proper packing and delivery of the material to the consignee. In the event of any loss, damage, or breakage, leakage or shortage in transit, the Bidder shall be responsible. No extra cost on such account shall be admissible.</p> |
| <b>19.</b> | <b>CORRECTION OF ARITHMETIC ERRORS</b>  |
|            | <p>Provided that a financial bid is substantially responsive, the procuring entity will correct arithmetical errors during evaluation of financial bids on the following basis:</p> <p>(i) If there is a discrepancy between the unit price and the total price that is obtained by multiplying the unit price and quantity, the unit price shall prevail and the total price shall be corrected, unless in the opinion of the procuring entity there is an obvious misplacement of the decimal point in the unit price, in which case the total price as quoted shall govern and the unit price shall be corrected;</p> <p>(ii) If there is an error in a total corresponding to the addition or subtraction of subtotals, the subtotals shall prevail and the total shall be corrected.</p> <p>(iii) If there is a discrepancy between words and figures, the amount in words shall prevail, unless the amount expressed in words is related to an arithmetic error, in which case the amount in figures shall prevail subject to clause (a) and (b) above.</p> <p>If the bidder that submitted the lowest evaluated bid does not accept the correction of errors, its bid shall be disqualified and its bid security shall be forfeited or its bid securing declaration shall be executed.</p>                 |
| <b>20</b>  | <b>PROCURING ENTITY'S RIGHT TO VARY QUANTITY:</b>   |
|            | <p>(i) The quantity of equipments and instruments originally indicated in the bidding document may vary without any change in the unit prices and other terms and conditions of the bid and the conditions of contract.</p> <p>(ii) If the Managing Director JKMSCL J&amp;K procures less than the</p>  |

|            |   |
|------------|---|
|            | <p>quantity indicated in the bidding documents the bidder shall not be entitled for any claim or compensation except otherwise provided in the conditions of contract.</p> <p>(i) If the Bidder fails to supply the Managing Director JKMSCL J&amp;K shall be free to arrange/procure the items and the extra cost incurred shall be recovered from the Supplier.</p>   |
| <b>21.</b> | <b>PARALLEL RATE CONTRACT</b>   |
|            | <p>The JKMSCL may also execute parallel rate contract to with more than one firm for each item on the lowest approved rates on the same terms and conditions, if the original lowest one each not in a position to supply material as per JKMSCL requirement.</p> <p>(i) To ensure sustained supply without any interruption, the bid inviting authority reserves the right to approve more than one supplier to supply the requirement among the qualified bidders.</p> <p>(ii) Orders will be placed with Lowest I (L-1) firm. However in case of any exigency at the discretion of the bid inviting authority, the orders may also be placed with the other firms, in the ascending order, L-2, L-3 and so on who have matched with the L-1 rates and executed agreement with corporation on same rates (L1), terms and conditions.</p> <p>(iii) After the conclusion of financial bid opening (Cover B) the lowest offer of the bidder is considered for negotiation and rate arrived after negotiations is declared as L-1 rate and L-1 supplier for an item for which the bid has been invited.</p> <p>(iv) The bid who has been declared as L-1 supplier for certain item shall execute necessary agreement for the supply of the required quantity of such item on depositing the required amount performance security and on execution of the agreement such bidder is eligible for the placement of supply orders.</p> <p>(v) JKMSCL will inform the L-1 rate to the bidders who had qualified for financial bid (Cover B) opening, inviting their consent to match with the L-1 rates for the item/items quoted by them and the bidders who agree to match L-1 rate, will be considered as matched L-1</p> <p>(vi) The bidder who agrees to match L-1 rate shall furnish the breakup detail (Rate, GST, CUSTOM DUTY etc.) of rates (L-1 rates).</p> <p>(vii) The supplier, on receipt of the supply orders deems that the purchase orders exceeds the production capacity declared in the bid documents and the delay would occur in executing the order, shall inform the JKMSCL immediately without loss of time and in executing the order, shall be returned within 7 days from the date of issuing order, failing which the supplier would be deprived from disputing the imposition of liquidated damages, and penalty for the delayed supplies.</p> <p>(viii) If the L-1 supplier has failed to supply / intimated JKMSCL about his inability / delay in supply as per the supply order, the required items within the stipulated time or as the case may be, JKMSCL may also</p> |

|           |  |
|-----------|--|
|           | <p>place purchase orders with the matched L-1 Bidders for purchase of the items provided such matched L-1. Bidders shall execute necessary agreement indicating the production capacity as specified in the bid document on depositing the required amount. Such bidder is eligible for the placement of purchase orders for the item quoted by them.</p> <p>(ix) Subject to para (vii) above, while JKMSCL has chosen to place purchase orders with matched L-1 supplier and there are more than one such matched L-1 supplier, then the purchase orders for the requirement of items will be place with L-2 first on matched rates of L-1 and in case L-2 does not have the required capacity than L-3 would be considered on matched L-1 rates and the same order would be flowed in case of L-3, L-4, etc.</p> <p>(x) The matched L-1 supplier, on placement of purchase orders, will be deemed as L-1 rate supplier for the purpose of the bid and all provisions of the bid document applicable to L-1 rate bidder will apply to the matched L-1 supplier.</p> <p>(xi) If the supplier fails to supply the item for the purchase orders, at any point of time, either fully or partly, within the stipulated time, JKMSCL is at liberty to place purchase orders with other bidders (in ascending order, viz, L-2, L-3 and so on) at the price offered by then and in such cases the supplier is liable to indemnify JKMSCL, without any protest or demur, for the difference in cost incurred by JKMSCL and the JKMSCL is entitled to recover the difference in cost from the amount due / payable to the supplier.</p> <p>(xii) Parallel rate contract may be concluded as described above during any time / currency of rate contract subject to matching of L-1 rates, price fall clause and on same terms and conditions.</p> |
| <b>22</b> | <b>VALIDITY OF BID:</b>  |
|           | Bids shall be valid for a minimum period of 120 days from the date of opening of technical bid. Prior to the expiry of the period of validity of bid, the procuring entity, may request the bidders to extend the bill validity period for an additional specified period of time. A bidder may refuse the request and such refusal shall be treated as withdrawal of the bid but in such circumstances bid security shall not be forfeited.   |
| <b>23</b> | <b>PRICE ESCALATION:</b>   |
|           | Price escalation or price variation shall not be applicable or considered under any circumstances for the purchases made under this bid or agreement. However, the provisions provided for tax variations are exclusive to this clause.  |
| <b>24</b> | <b>SUBLETTING OF CONTRACT:</b>   |
|           | Subletting or assigning contract to third party is prohibited. In the event of bidder violating this condition, the Jammu and Kashmir Medical Supplies Corporation Limited shall be at liberty to place the contract elsewhere on the Bidder's account and at his risk. The bidder shall be liable for any loss or   |

|           |   |
|-----------|---|
|           | damage, which the Government may sustain in consequence or arising out of such replacement of the contract.   |
| <b>25</b> | <b>FALL CLAUSE:-</b>  |
|           | <p>(i) The prices under contract shall be subject to price fall clause. The prices charged for the store supplies under the contract by successful bidder shall in no event exceed the lowest price at which the successful bidder sells the stores of identical description to any other persons during the period of the contract anywhere in India. If any time, during the period of the contract, the bidder reduces the sales price chargeable under the contract, he shall forth with notify such reduction to the JKMSCL, J&amp;K and the price payable under the contract for the stores supplied after the date of coming into force of such reduction or sale shall stand reduced correspondingly. It imply that if the contract holder quotes/ reduces its price to render similar goods at a price lower than the contract price to anyone in the State /UT of India at any time during the currency of contract including extension period, the contract price shall be automatically reduced with effect from the date of reducing or quoting lower price for all delivery of subject matter of procurement under contract and the contract shall be amended accordingly.</p> <p>(ii) The firms holding parallel rate contract shall also reduce their price. Firms shall notify their reduced price and intimate their acceptance to the revised price within 15 days to JKMSCL. Similarly, if parallel rate contract holding firm reduced its price during currency of the rate contract, its reduced price shall be conveyed to other parallel rate contract holding firms and the original rate contract holding firms for corresponding reduction in their prices. If any rate contract holding firm does not agree to reduce price, further transaction with it, shall not be conducted.</p> |
| <b>26</b> | <b>COMPLIANCE WITH THE CODE OF INTEGRITY AND NO CONFLICT OF INTEREST:</b>   |
|           | <p>Any person participating in a procurement process shall-</p> <ol style="list-style-type: none"> <li>Not offer any bribe, reward or gift or any material benefit either directly or indirectly in exchange for an unfair advantage in procurement process or to otherwise influence the procurement process;</li> <li>Not misrepresent or omit misleads or attempts to mislead so as to obtain a financial or other benefit or avoid an obligation;</li> <li>Not indulge in any collusion, bid rigging or any-competitive behaviour to impair the transparency, fairness and progress of the procurement process;</li> <li>Not misuse any information shared between the procuring entity and the bidders with an intent to gain unfair advantage in the procurement process;</li> <li>Not indulge in any coercion including impairing or harming or threatening to do the same, directly or indirectly, to any part or to its property to influence the procurement process;</li> <li>Not obstruct any investigation or audit of a procurement process;</li> </ol>   |



|           |   |
|-----------|---|
|           | <p>g) Disclose conflict of interest, if any; and</p> <p>h) Disclose any previous transgressions with any entity in India or any other country during the last three years or any debarment by any other procuring entity.</p> <p><b>Conflict of Interest :</b></p> <p>The bidder participating in a bidding process must not have a conflict of interest. A conflict of interest is considered to be a situation in which a party has interests that could improperly influence that party's performance of official duties or responsibilities, contractual obligations, or compliance with applicable laws and regulations.</p> <p>A bidder may be considered to be in conflict of interest with one or more parties in bidding process if, including but not limited to :</p> <ol style="list-style-type: none"> <li>Have controlling partners/shareholders in common; or</li> <li>Receive or have received any direct or indirect subsidy from any of them; or</li> <li>Have the same legal representative for purposes of the bid; or</li> <li>Have a relationship with each other, directly or through common third parties, that puts them in a position to have access to information about or influence on the bid of another bidder, or influence the decisions of the procuring entity regarding the bidding process; or</li> <li>The bidder participates in more than one bid in a bidding process. Participation by a bidder in more than one bid will result in the disqualification of all bids in which the bidder is involved. However, this does not limit the inclusion of the same subcontractor, not otherwise participating as a bidder, in more than one bid; or</li> <li>The bidder or any of its affiliates participated as a consultant in the preparation of the design or technical specification of the goods, works or services that are the subject of the bid; or bidder or any of its affiliates has been hired (or is proposed to be hired) by the procuring entity as engineer-in charge/consultant for the contract.</li> </ol> <p>Bidder or any of its affiliates has been hired (or is proposed to be hired) by the Procuring Entity as engineer-in-charge / consultant for the contract.</p> |
| <b>27</b> | <p>All correspondence in this connection should be addressed to the Managing Director JKMSCL, J&amp;K. Technical questions should be referred to the Managing Director JKMSCL, J&amp;K direct by correspondence or by personal contact.</p>   |
| <b>28</b> | <ol style="list-style-type: none"> <li>Direct or indirect canvassing on the part of bidders or their representative shall disqualify their bids.</li> <li>Supplier may be disqualified, banned or suspended from business during the rate contract if : <ol style="list-style-type: none"> <li>fails to execute a contract or fails to execute it satisfactorily ;</li> <li>no longer has the technical staff or equipment considered necessary ;</li> <li>is declared bankrupt or insolvent or its financial position has become unsound, and in the case of a limited company, it is</li> </ol> </li> </ol>   |

|           |   |
|-----------|---|
|           | <p>wound-up or taken into liquidation ;</p> <p>(d) The firm is suspected to be doubtful loyalty to state.</p> <p>(e) The State Bureau of Investigation (SBI) or any other Investigating agency recommends such a course in respect of a case under investigation.</p> <p>(f) Managing Director JKMSCL, J&amp;K is prima- facie of the view that the firm is guilty of an offence involving moral turpitude in relation to business dealings, which if established would result in business dealing with it banned.</p>  |
| <b>29</b> | No action on the letter head of the bidder /firm regarding any complaints against the JKMSCL will be considered unless the letter head bears the signature of the bidder or the authority higher than the bid signatory of the firm.  |
| <b>30</b> | <p>(i) If any certificate/documents/information submitted by the bidder found to be false/ forged/ fabricated/vexatious or frivolous or malicious appeals or complaints etc. than bidder shall be liable for the appropriate legal action along with disqualification, banning, suspension etc. for limited or unlimited period.</p> <p>(ii) Bidders are required to submit wanted information (if any) based on the facts. If the furnished information by the firm found to misleading or not based on facts disciplinary action against the firm may be taken as to banning concerned item/items for certain or uncertain period.</p>  |
| <b>31</b> | The JKMSCL reserves the right to accept any bid not necessarily the lowest. The JKMSCL may reject any bid without assigning any reasons and accept bid for all or anyone or more of the articles for which bidder has been given or distribute items of stores to more than one firm/supplier.  |
| <b>32</b> | <b>GRIEVANCE</b>  |
|           | <b>Grievance</b> regarding interpretation of any clause of the contract/agreement executed between the parties shall be referred to Managing Director, JKMSCL for its clarification.  |
| <b>33</b> | <b>ARBITRATION</b>  |
|           | <p>33.1 Governing Law: This NIT shall be governed by and construed in accordance with the laws of the Union Territory of Jammu and Kashmir and the laws of India as applicable to the Union Territory of Jammu and Kashmir.</p> <p>33.1.1 Amicable Settlement: Either party is entitled to raise any claim, dispute or difference of whatever nature arising under out of or in connection with the NIT including its existence or validity or termination (collectively "dispute") by giving a written notice to the other party, which shall contain</p> <ol style="list-style-type: none"> <li>a description of the dispute</li> <li>the ground for such dispute</li> <li>all written material in support of its claim</li> </ol> <p>33.1.2 The other party shall, within thirty days of issuance of dispute notice issued, furnish:</p> |



|           |   |
|-----------|---|
|           | <p>I. Counter claim and defences, if any, regarding the dispute; and</p> <p>II. All written material in support of its defences and counter claim</p> <p>34.1.3 Within thirty days of issuance of notice by any party pursuant to para 29.1.2 both the parties to the dispute shall meet to settle such dispute amicably. If the parties fail to resolve the dispute amicably within thirty days of the receipt of the notice referred to in the above para the dispute shall be referred to Managing Director, JKMSCL, J&amp;K for its reference to arbitration.</p> <p><b>Dispute Resolution: Besides, as referred above may also include any dispute arising out of contract with regard to the interpretation, meaning and breach of the terms of the contract, the matter shall be referred to the Administrative Department H&amp;ME, who will, through Law Department, appoint a senior most officer as sole Arbitrator, of the dispute, who will not be related to this contract and whose decision shall be final and binding on both the parties. The Arbitrator proceedings shall be governed by the J&amp;K Arbitration and Conciliation Act, 1997. The venue of the Arbitration shall be in the UT of Jammu and Kashmir.</b></p> <p><b>Note:</b> - Small grievances regarding interpretation of any clause of the Contract / Agreement executed between the parties shall be referred to Managing Director, JKMSCL for its clarification and such interpretation(s) given shall not become subject matter for reference to Arbitration</p> |
| <b>34</b> | The JKMSCL will have the right of rejection of all or any of the bids without assigning any reason for the same. The right to conclude parallel rate contracts with another firm for the stores detailed in Table I is also reserved by the Managing Director JKMSCL, J&K   |
| <b>35</b> | Extra stipulation or any other condition contrary to the above bid conditions are not acceptable and may render the bid liable to rejection.  |
| <b>36</b> | The bidder must sign all the pages of bid document at the below of terms & conditions agreeing to abide by all conditions of the bid and accept them in toto. The Signing of Annexure A1 shall be treated as acceptance of all the terms and conditions of the bid document.  |
| <b>37</b> | The Managing Director JKMSCL, J&K may relax or change/ modification in terms and conditions in the exigency excluding fundamental changes. In case of such urgency the terms & conditions shall be got approved from Purchase committee of Managing Director JKMSCL, J&K as the case may be.  |
| <b>38</b> | <b>JURISDICTION:-</b> All actions, legal proceedings and suits arising from or connected to this bid shall be subject to the exclusive jurisdiction of courts in J&K only.  |

## Section VI B: - Special Conditions of Contract (SCC)

The following Special Conditions of Contract (SCC) shall supplement the General Conditions of Contract (GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC. The clauses of special conditions of contract are as follows:-

| Clause No. | Particulars   |
|------------|---|
| 1.         | Technical details, bid security, tender cost, tender processing fee and all other required documents should be uploaded under <b>Cover "A" Technical Bid</b> and financial details (BOQ) should be uploaded under <b>Cover "B"</b> . No document except financial instrument (DD/FDR) & catalogues of the bid items shall be entertained physically by the Corporation.   |
| 2.         | Pre-requisite, if any, for installation, including UPS, computer, printer, and other items should be provided by the firm in technical bid and financial bid respectively.  |
| 3.         | Firm shall provide comprehensive maintenance with spare parts for item(s), as mentioned in Technical specification (from the date of installation / demonstration).   |
| 4.         | Conditional bids shall not be considered.   |
| 5.         | Normally, payment shall be released after installation, demonstration and successful commissioning of equipment/ITEM and satisfactory operational training.   |
| 6.         | All certificates should be valid on the date of submission of bids and issue of supply order.   |
| 7.         | The bidder should have well equipped local service centre in India preferably in J&K.   |
| 8.         | <ol style="list-style-type: none"> <li>i. The bidder shall be a manufacturer/direct importer/authorised representative of the original manufacturer/importer who must have manufactured/ imported and supplied and installed this equipment(s) in India satisfactorily.</li> <li>ii. The merger / amalgamation / transfer of business / transfer of assets etc. of a firm affects the bid condition relating to 'past performance' in preceding years. In cases where bidder acquired an ongoing business or assets of another entity, eligibility in respect of the past performance and condition relating to minimum turn over in preceding years shall be decided based on specific mention in purchase and transfer of ownership agreement / agreement of sale of business and / or its assets / board of directors (B.O.D) resolution chartered accountant certification or any other document (s) in this regard, which the bidder shall have to submit preferably with the bid. The eligibility of a bidder in this regard shall be ascertained by the purchase committee on the basis of the above stated agreement or any other document(s) and the decision of purchase committee shall be final.</li> </ol> |

|     |   |
|-----|---|
| 9.  | The name, make, model and brand of equipments, which are offered, should be mentioned in against each item. Mere indication of English/USA/Indian will not serve the purpose.   |
| 10. | In the case of supply of imported item the suppliers may be asked to furnish a certificate to the effect that the firm has completed all the formalities in connection with import of the item in question.   |
| 11. | In case the item approved by the JKMSCL is procured by any other department on the rate contract of JKMSCL, the administrative charges to the extent of 5% of the invoice value shall be deposited by the approved firm or else, the firm/supplier shall be liable to be penalised which may lead to blacklisting/debarring from entering into the tender process for not less than 05 years by JKMSCL besides forfeiture of earnest money or any other action as deemed fit by the Managing Director, JKMSCL.  |
| 13  | The Supplier/service providing firm shall be liable to pay a <b>penalty of Rupees five thousand per day</b> , if the firm didn't respond after 48 hours from the time of receiving first complaint. The complaint may be sent to firm by way of telephone /fax/letter or e-mail. The amount of liquidation damage shall be directly deducted from the security deposit of the firm at the time of refund or before by way of any adjustment order. All breakdown calls to be attended within 24 hrs at (within city limits) and 48 hrs for other districts/peripheral areas |

**APPLICABILITY OF CLAUSES:** - All the clauses from 1 to 38 of general terms and conditions and from 1 to 13 of special terms and conditions and their annexure, formats & enclosures are applicable for the bid items.

Managing Director  
Jammu and Kashmir Medical Supplies Corporation Limited

I/We have read the above terms and conditions and I/We agree to abide myself/ourselves by the above terms & conditions of the bid document

Signature of bid with seal

## Section VI C: Contract Forms (CF)

### Table of contents

| S.No. | Description  | Pages                             |
|-------|--|-----------------------------------|
| 1.    | Declaration of bidder regarding acceptance bid for terms & conditions (Annexure A1)              |                                   |
| 2.    | Agreement Form (Annexure IV)   | To be downloaded from the website |
| 3.    | Form for bank guarantee (on bank letter head)  | To be downloaded from the website |
| 4.    | Format-Authorized Representatives/Agents of Original Manufacturer/Direct Importer (Annexure AII) |                                   |
| 5.    | Technical Specifications (Annexure AIII)   |                                   |

(On Letter Head of the Bidder)

**DECLARATION**

I/We M/s. .... represented by its Proprietor/managing Partner/Managing Director having its Registered Office at ..... and its Factory Premises at ..... do declare that I/we have carefully read all the conditions of bid no. .... Dated.....including all the amendments in ..... Ref. ....for supply cum rate contract of ..... **Item name** ..... for Jammu and Kashmir Medical Supplies Corporation Ltd. for the year 2022-23 and accepts all conditions of bid including amendments, if any.

I/We agree that the M.D. JKMSCL, Jammu / Srinagar (J&K) may forfeit bid security and or performance security and debar me/us for a period specifying in orders, if any information/document furnished by us is proved to be false/fabricated at the time of inspection and not complying with the terms and conditions of the bid document as presented in bid, Annexure-B and other relevant documents.

Signature & Seal of bidder  
Name & Address:

Format-Authorized Representative of Original Manufacturer/Direct Importer

***In case, original manufacturer/direct importer wish to authorise any representative to bid, to negotiate, to raise invoice to receive payments on behalf of original manufacturer/direct importer/against invoice raised by representative.***

The Managing Director,  
Jammu & Kashmir Medical Supplies Corporation Ltd.  
J&K  
Dear Sir,

We \_\_\_\_\_ who are established and reputed manufacturers of \_\_\_\_\_ having factories at \_\_\_\_\_ Registered office at \_\_\_\_\_ possessing manufacturing license No. \_\_\_\_\_ and do hereby authorize M/S \_\_\_\_\_ (Name and Address of Representative) to submit a bid and subsequently negotiate with you against the above mentioned tender, subject to the condition that I/we, the original manufacturer/direct Importer of the bidding items and our authorized representative \_\_\_\_\_ are ready to execute Tripartite agreement with the Corporation i.e JKMSCL stating inter-alia that:-

1. The invoice submitted by the authorised representative for such supplies shall be endorsed by me /us i.e. the original Manufacturer/Direct Importer of bidding items and original copy of the delivery challan of Manufacturer's towards authorised representative for such supplies shall also be endorsed along with invoice submitted by our Authorized Representative.
2. JKMSCL may secure an e-mail /alternative confirmation for authenticity of such supplies from Manufacturer/Direct Importer, before releasing the payment, which we are committed to provide.
3. The payment shall however be released on the terms and conditions of tripartite agreement to be signed between JKMSCL, Original Manufacturer / Direct Importer and the authorized representative of Original Manufacturer / Direct Importer of the bidding items for such supplies made by the authorized representative, on behalf of me / us.
4. In case of change of Dealership we shall be responsible for providing after sales services and maintenance of the equipment free of cost during the warranty period..

No company or firm or individual other than M/S \_\_\_\_\_ are authorized to bid, negotiate and conclude the contract in regard to this business against this specific tender.

I / we, further agree to comply with the conditions specified under Clause 2(a) –Eligibility Conditions, of the tender document. We hereby extend our full guarantee as per the tender conditions for the goods offered for supply against this invitation for bid by the above Firm.

Yours faithfully

Name

For and on behalf of M/S

(Name of the manufacturer/Direct Importer)

**Note: This letter of authority should be on the letter head of original Manufacturer/Direct Importer of bidding items and should be signed and sealed by the Proprietor/ Managing Director of the firm / authorized signatory and shall have to be duly notarised.**

## **Machinery/Equipment (Bipartite Agreement)**

Annexure IV

[on Rs. 100/- Non-Judicial Stamp Paper- "Affidavit"]

### **Agreement: 1**

(For Manufacturers/ Direct Importers only)

This deed of agreement is made on this ..... day of ..... 2022 between Jammu & Kashmir Medical Supplies Corporation Limited represented by its General Manager(P&S) having its registered office at Plot No. 58 Friends Colony, Trikuta Nagar, Jammu, 180020 /Opp. State Motor Garage office, Bemina, Srinagar (herein after referred to as "First Party" (Purchaser) which term shall include its successor, representatives, executors assigns, administrator and beneficiaries unless excluded by the contract) and M/s ..... (Original Manufacturer/ Direct Importer) represented by its Proprietor/ Managing Director/ Managing Partner/ Authorized Signatory of the company/ firm having its registered office at ..... and its factory premises at ..... (herein after referred to as "Second Party" (Suppliers) which term shall include its successors representatives, heirs, executors and administrators unless excluded by the contract).

Whereas the (Original Manufacturer/ Direct Importer) (Second Party)) have agreed to supply to First Party ( Purchaser),Machinery & Equipment/Instruments with specifications mentioned in the scheduled attached here to at the prices noted herein and in the manner and under the terms and conditions herein after mentioned and whereas the second party has agreed to deposit performances security to first party, equivalent to 3% of the tentative cost/ contract value (rounded to the nearest round number) in the scheduled attached as per clause 11 of the tender document in the form bank of guarantee for the due and faithful performance of this agreement, to be forfeited in the event of Second Party failing duly and faithfully to perform it. Now these presents witness that for carrying out the said agreement in this behalf into execution the Second Part and the First Party (Purchaser) do hereby mutually covenant, declare, contract and agree each of them in the manner following, that is to say,

1. The term "Agreement", wherever used in these connection shall mean and includes the terms and conditions contained in the invitation to bid floated for the rate contract cum supply for Machinery & Equipment/Instruments for Jammu & Kashmir Medical Supplies Corporation Limited (Rate Contract for two years (24) months period, extendable for another three (03) months with mutual consent) (NIT/JKMSCL/M&E/2022/..... dated ..... and technical bid opened on ..... , the instructions to bidders, the condition of bid, acceptance of bid, particulars herein after defined and those eligibility criteria, general conditions and other conditions that may be added from time to time.

2.1. The agreement is for the supply, by the Second Party (Suppliers) to the First Party (Purchaser), of the Machinery & Equipment/Instruments on terms and conditions set forth in the agreement.

2.2. This agreement shall be deemed to have come into force with effect from the date of signing of the agreement and it shall remain in force up to a period of two years (24) months which can further be extended for another three (03) months with mutual consent of First Party and Second Party.

2.3. The bid quantity noted against each item in the scheduled attached here to indicates only the probable/ tentative total requirement of the First Party in respect of each item for the agreement period indicated in clause “2.2” above. This quantity may increase or decrease at the discretion of the First Party. The Second Party (Supplier) shall make supplies of the Machinery & Equipment/Instruments on the basis of Purchase order only placed on him/ her from time to time by the ordering authority of First Party (Purchaser-JKMSCL) specifying the quantity required to be supplied at a specific location/ locations within the UT of Jammu and Kashmir.

2.4. The Second Party shall have no right/ query regarding placing of orders against the tentative requirement mentioned in the schedule enclosed which may increase or decrease or First Party may not issue any order for certain item/ items mentioned therein the schedule enclosed/ tentative/ Indicative quantity.

2.5. The release of payment shall be as per terms and conditions/ payment clause 17 of the tender document and after successful installation of the Machinery/Equipments at the end user site and after due verification of bills by the end user department and deduction and penalties as per the clause 18 & 19 of the tender document.

### **3. SUPPLIES ON THE RATE CONTRACT OF JKMSCL:**

The Second Party shall in no case, use the rate contract of JKMSCL for making supplies and / or comparing of rates to/ with any of other department(s)/ agency(ies)/ NGO etc. In case Second Party supplies any of the item(s) at the rate contract or provides the document for comparison of rates or otherwise, to any other department(s)/ agency(ies)/ NGO(s) etc, the defaulted Second Party shall have to pay 7.5% of the total invoice value of the product(s) supplied to other department(s)/ agency(ies) etc at the rate contract of JKMSCL as penalty to the first party (JKMSCL-purchaser) and further the Second Party shall be liable to be considered for Debarring/ Blacklisting for a period not less than five years.

### **4. TERMINATION OF CONTRACT ON BREACH OF CONDITION.**

4.1. In case the supplier fails or neglects or refuse to faithfully perform any of the covenants on his part herein contained, it shall be lawful for the First Party to forfeit the amount deposited by the supplier (second party) as performance security and cancel the contract.

4.2. In case the Second Party neglects or refuse to observe, performs, fulfil and keep, or any one or more or any part of any one of covenants, stipulation and provisions herein contained, it shall be lawful for the First Party on any such failure, neglect or refusal, to put an end to this agreement and there upon on every article, cause and thing herein contained on the part of First Party shall cease and be void and in case of any damage, loss, expenses, differences in cost or other from out of deposit/ due for the time being payable to the Second Party under this and/ or any other contract and in case such last mentioned deposit/ dues are insufficient to cover all such damages, loses, expenses, difference in cost and other deposit as aforesaid, it shall be lawful for the First Party to appropriate the performance security made by the supplier as herein before mentioned to reimburse all such damages, losses, expenses and difference in cost and other money as the purchaser shall be sustained, incurred or been put to by reason of the Second Party (Supplier) having



been guilty of any such failure negligence or refusal as aforesaid or other breach in the performance of this contract.

4.3. If any time during the course of contract it is found that the information furnished by the Second Party (Supplier) to the First Party (Purchaser) either in his bid or otherwise, is false, the purchaser may put an end to the contract/ agreement wholly or in part and thereupon the provision of clause “4.1” above shall apply or any other action deemed fit by the First Party may also apply.

4.4. The First party (Purchaser-JKMSCL) reserves the right to terminate, without assigning any reasons the contract/ agreement either wholly or in part, without any notice to the Second Party. The Second Party shall not be entitled for any compensation what so ever in respect of such termination of the contract/ agreement by the First Party.

5. All certificates or notices or orders for time or for extra, varied or altered suppliers which are to be the subject of extra or varied charges whether so described in the Agreement or not, shall be in writing and unless in writing shall not be valid, binding or be of any effect what so ever.

6. The Second Party (Supplier) shall not be in any way interested in or concerned directly or indirectly with any of the officer, subordinate or servants of the First Party. In any trade, business or transaction nor shall the Second Party give or pay or promise to give or pay any such officer, subordinate, servant directly or indirectly any money or fee or other consideration under designation of “Custom” or otherwise; nor shall the Second Party permit any person or persons whomsoever to interfere in the management or performance hereof under the Power of Attorney or otherwise without the consent in writing of the First Party obtained in first hand.

7. In case the Second Party (Suppliers) at any time during the continuance of the contract becomes bankrupt or insolvent or commits any act of bankruptcy or insolvency under the provisions of any law in that behalf for the time being in force or should compound with his creditors, it shall be lawful for the First Party to put an end to the agreement and there upon on every article, clauses and thing herein contained to be operative on the part of the purchaser, shall cease and be void and the First Party shall have all the rights and remedies given to him under the preceding clauses.

## **8. SERVING OF NOTICE TO SUPPLIER**

8.1. All notice or communication relating to or arising out of this agreement or any of the terms thereof shall be considered duly served on or given to the Second Party (Suppliers) if delivered to him or left at his/ her premises, place of business or abode.

9. And it is hereby agreed and declared between the parties hereto that in case any question of dispute arises touching the construction or wording of any of clause herein contained the rights, duties, liabilities of the parties hereto or any other way, touching or arising out of the presents the decision of the Managing Director, JKMSCL in the matter shall be final and binding.

10. All disputes arising out of this agreement and all questions relating to the interpretation of this agreement shall be decided by first and the final appellant authority and decision of said authority shall be final.

11. The following documents shall be deemed to form & be read & construed as part of this agreement:

- a) LOI
- b) NIT & Corrigendum issued thereof, if any.

11.1 Secondparty shall indemnify and keep the First party free from any harm, against all losses, expenditures, damages, costs & claims incurred or suffered by or made against the Second/Third party by reason of any breach by the same of any of its obligations, covenants, representations & warranties.

11.2 NIT Ref. No: NIT/JKMSCL/M&E/2022/.....dated.....

11.3 List of Equipment(s), accessories, optional items (if any) under this agreement.

12. All terms and conditions of the NIT shall be the part of this agreement.

Original Manufacturer/ Direct Importer  
(Supplier)

Jammu & Kashmir Medical Supplies Corporation  
Ltd  
(First Party)

(Second Party)  
(Signature, Name & full Address with stamp)

Represented by  
General Manager (P&S)/ JKMSCL  
(Signature, Name & full Address with Stamp)

Witness (Signature, Name & Address)

Witness (Signature, Name & Address)

1.

1.

2.

2.

## **Machinery/Equipment (Tripartite Agreement)**

[on Rs. 100/- Non-Judicial Stamp Paper- “Affidavit”]

### **Agreement: 2**

(Tripartite Agreement for Authorized Agents/Dealers/Facilitators)

This deed of agreement is made on this ..... day of ..... 2022 between Jammu & Kashmir Medical Supplies Corporation Limited represented by its General Manager (P&S) having its registered office at Plot No. 58 Friends Colony, Trikuta Nagar, Jammu, 180020 /Opp. State Motor Garage office, Bemina, Srinagar (herein after referred to as “First Party” (Purchaser) which term shall include its successor, representatives, executors assigns, administrator and beneficiaries unless excluded by the contract), M/s ..... (Original Manufacturer/ Direct Importer) represented by its Proprietor/ Managing Director/ Managing Partner/ Authorized Signatory of the company/ firm having its registered office at ..... and its factory premises at ..... (herein after referred to as “Second Party” (Suppliers) which term shall include its successors representatives, heirs, executors and administrators unless excluded by the contract) and M/s ..... (Authorized agent/ dealer/ facilitator) represented by its Proprietor/ Managing Partner/ Managing Director having its registered office at ..... (herein after referred to as “Third Party”- (Authorized Agent/ Suppliers/ Dealers) of Second Party, which term shall include its successors representative, heirs, executors and administrators unless excluded by the contract).

Whereas the (Original Manufacturer/ Direct Importer) (Second Party/ Third Party (Authorized agents/ dealer)) have agreed to supply to First Party ( Purchaser), Machinery & Equipment/Instruments with specifications mentioned in the scheduled attached here to at the prices noted herein and in the manner and under the terms and conditions herein after mentioned and whereas the second party/ third party have agreed to deposit performances security to first party, equivalent to 3% of the tentative cost/ contract value (rounded to the nearest round number) in the scheduled attached as per clause 11 of the tender document in the form of bank guarantee for the due and faithful performance of this agreement, to be forfeited in the event of Second Party/ Third Party failing duly and faithfully to perform it. Now these presents witness that for carrying out the said agreement in this behalf into execution the Second Part/ Third Party, and the First Party (Purchaser) do hereby mutually covenant, declare, contract and agree each of them in the manner following, that is to say,

1. The term “Agreement”, wherever used in these connection shall mean and includes the terms and conditions contained in the invitation to bid floated for the rate contract cum supply for Machinery & Equipment/Instruments for Jammu & Kashmir Medical Supplies Corporation Limited (Rate Contract for twenty-four (24) months period, extendable for another three (03) months with mutual consent) (NIT/JKMSCL/M&E/2022/.....dated.....and technical bid opened on....., the instructions to bidders, the condition of bid, acceptance of bid, particulars herein after defined and those eligibility criteria, general conditions and other conditions that may be added from time to time.

2.1. The agreement is for the supply, by the Second Party/ Third Party (Suppliers) to the First Party (Purchaser), of the Machinery & Equipment/Instruments on terms and conditions set forth in the agreement.

2.2. This agreement shall be deemed to have come into force with effect from the date of signing of the agreement and it shall remain in force upto a period of twenty-four (24) months period which can further be extended for another three (03) months with mutual consent of First Party and Second Party/ Third Party.

2.3. The bid quantity, if mentioned against each item in the schedule indicates only the probable/ tentative total requirement of the First Party in respect of each item for the agreement period indicated in clause “2.2” above. This quantity may increase or decrease at the discretion of the First Party. The Second Party/ Third Party (Supplier) shall make supplies of the Machinery & Equipment/Instruments on the basis of Purchase order only placed on him/ her from time to time by the ordering authority of First Party (Purchaser-JKMSCL) specifying the quantity required to be supplied at a specific location/ location within the UT of Jammu and Kashmir.

2.4. The Second Party/ Third Party shall have no right/ query regarding placing of orders against the tentative requirement mentioned in the schedule enclosed which may increase or decrease or First Party may not issue any order for certain item/ items mentioned therein the schedule enclosed/ tentative/ Indicative quantity.

### **3. AUTHORIZED AGENTS/ DEALERS OF SECOND PARTY:**

3.1. In this agreement, the Second Party (Original Manufacturer/ Direct Importers) have authorised M/s ..... ; (Third Party) as Agent/ Distributors/ Dealers to submit bid, to negotiate with First Party, to raise invoice and receive payment on behalf of Second Party; and as such, supplies shall be endorsed by the Second Party M/s ..... (Original Manufacturer/ Direct Importers) and original copy of delivery challan of Second Party towards the Third Party for such supplies shall be endorsed along with invoice submitted by Third Party to First Party.

3.2. The Corporation under such arrangements shall have a right to secure confirmation to authority of suppliers from Second Party before releasing the payments.

3.3. The release of payment shall be as per terms and conditions/ payment clause 17 of the tender document and after successful installation at end user site and proper verification of bills from the end user department and deduction & penalties as per the clause 18 & 19 of the tender document.

### **4. SUPPLIES ON THE RATE CONTRACT OF JKMSCL:**

The Second Party or Third Party shall in no case, use the rate contract of JKMSCL for making supplies and / or comparing of rates to/ with any of other department(s)/ agency(ies)/ NGO etc. In case Second Party/ Third Party supplies any of the item(s) at the rate contract or provides the document for comparison of rates or otherwise, to any other department(s)/ agency(ies)/ NGO(s) etc, the defaulted Second Party or Third Party, wherever applicable, shall have to pay 7.5% of the total invoice value of the product(s) supplied to other department(s)/ agency(ies) etc at the rate contract of JKMSCL as penalty to the first party (JKMSCL-purchaser) and further the Second Party/ Third Party shall be liable to be considered for Debarring/ Blacklisting for a period not less than five years.

### **5. TERMINATION OF CONTRACT ON BREACH OF CONDITION.**

5.1. In case the supplier fails or neglects or refuse to faithfully perform any of the covenants on his part herein contained, it shall be lawful for the First Party to forfeit the amount deposited by the supplier (second party/ third party) as performance security and cancel the contract.

5.2. In case the Second Party/ Third Party fails, neglects or refuse to observe, performs, fulfil and keep, or any one or more or any part of any one of covenants, stipulation and provisions herein contained, it shall be lawful for the First Party on any such failure, neglect or refusal, to put an end to this agreement and there upon on every article, clause and thing herein contained on the part of First Party shall cease and be void and incase of any damage, loss, expenses, differences in cost or other from out of deposit/ due for the time being payable to the Second Party/ Third Party under this and/ or any other contract and in case such last mentioned deposit/ dues are insufficient to cover all such damages, loses, expenses, difference in cost and other deposit as aforesaid, it shall be lawful for the First Party to appropriate the performance security made by the supplier as herein before mentioned to reimburse all such damages, losses, expenses and difference in cost and other money as the purchaser shall be sustained, incurred or been put to by reason of the Second Part/ Third Party (Supplier) having been guilty of any such failure negligence or refusal as aforesaid or other breach in the performance of this contract.

5.3. If any time during the course of contract it is found that the information furnished by the Second Party/ Third Party (Supplier) to the First Party (Purchaser) either in his bid or otherwise, is false, the purchaser may put an end to the contract/ agreement wholly or in part and thereupon the provision of clause “5.1” above shall apply or any other action are deemed fit by the First Party may also apply.

5.4. The First party (Purchaser-JKMSCL) reserves the right to terminate, without assigning any reasons the contract/ agreement either wholly or in part, without any notice to the Second Party/ Third Party. The Second Party/ Third Party shall not be entitled for any compensation what so ever in respect of such termination of the contract/ agreement by the First Party.

6. All certificates or notices or orders for time or for extra, varied or altered suppliers which are to be the subject of extra or varied charges whether so described in the Agreement or not, shall be in writing and unless in writing shall not be valid, binding or be of any effect what so ever.

7. The Second Party/ Third Party (Supplier) shall not be in any way interested in or concerned directly or indirectly with any of the officer, subordinate or servants of the First Party. In any trade, business or transaction nor shall the Second Party/ Third Party give or pay or promise to give or pay any such officer, subordinate, servant directly or indirectly any money or fee or other consideration under designation of “Custom” or otherwise; nor shall the Second Party/ Third Party permit any person or persons whomsoever to interfere in the management or performance hereof under the Power of Attorney or otherwise without the consent in writing of the First Party obtained in first hand.

8. In case the Second Party/ Third Party (Suppliers) at any time during the continuance of the contract becomes bankrupt of or in solvent or commits any act of bankrupt or insolvency under the provisions of any law in that behalf for the time being inforce or should compound with his creditors, it shall be lawful for the First Party to put an end to the agreement and there upon on every article , clauses and thing herein contained to be operative on the part of the purchaser, shall cease and be void and the First Party shall have all the rights and remedies given to him under the preceding clauses.

8.1. In case Third Party, (Authorized Agent/ Dealer/ facilitator- clause 3) at any time during the continuance of the contract become bankrupt of or insolvent or commits any act of bankrupt or insolvency either provisions of any law in that behalf for the time being in force, or should compound with his creditors, the Second Party, (Original Manufacturer/ Direct Importers) shall be bound to continue with the supplies directly for the First Party till the completion of contract otherwise it shall be lawful for the purchaser to put an end to the agreement and thereupon every article, clause and thing herein contained to

be operative as part of First Party, shall cease and be void and the First Party shall have all the rights and remedies given to him under the preceding clauses.

## 8.2 SERVING OF NOTICE TO SUPPLIER

**All notice or communication relating to or arising out of this agreement or any of the terms thereof shall be considered duly served on or given to the Second Party/ Third Party (Suppliers) if delivered to him or left at his/ her premises, place of business or abode.**

9. And it is hereby agreed and declared between the parties hereto that in case any question of dispute arises touching the construction or wording of any of clause herein contained the rights, duties, liabilities of the parties hereto or any other way, touching or arising out of the presents, the decision of the **Managing Director, JKMSCL** in the matter shall be final and binding.

10. All disputes arising out of this agreement and all questions relating to the interpretation of this agreement shall be decided by first and the final appellant authority as reflected in NIT and decision of said authority shall be final.

11. The following documents shall be deemed to form & be read & construed as part of this agreement:

- c) LOI
- d) NIT & Corrigendum issued thereof, if any.

11.1 Second/Third party shall indemnify and keep the First party free from any harm, against all losses, expenditures, damages, costs & claims incurred or suffered by or made against the Second/Third party by reason of any breach by the same of any of its obligations, covenants, representations & warranties.

11.2 NIT Ref. No: NIT/JKMSCL/M&E/2022/.....dated.....

11.3 List of Equipment(s), accessories, optional items (if any) under this agreement.

12. All terms and conditions of the NIT/relevant clause of SPP of JKMSCL shall be the part of this agreement.

Authorized Agent/ Dealer  
(Third Party)  
(Signature, Name & full Address with stamp)

Original Manufacturer/ Direct Importer (Supplier)  
(Second Party)  
(Signature, Name & full Address with stamp)

Witness (Signature, Name & Address)

1.

2.

Witness (Signature, Name & Address)

1.

2.

# Technical Specifications of Equipment's for TB Containment Lab

## **1. Biological Safety Cabinet Class 2A with thimble ducting and with UPS**

**Description of function:** The class II BSC is used in a TB laboratory for processing specimens consisting of liquefaction of sputa and handling cultures of tubercle bacilli. This type of BSC is not adequate for handling volatile or toxic chemicals or radionuclides. Before ordering a BSC, facility and engineering requirements to be checked. Organization of a periodical recertification of the BSC by an authorized agency to be in place.

### **Main specifications**

- The BSC to meet the requirements of class IIA2 NSF 49 or class II EN 12469; specifically, with regard to inward airflow ( $\geq 0.40$  m/s according to EN 12469:2000 or  $\geq 0.50$  m/s according to NSF 49:2004).
- External height  $\leq 2200$  mm including support stand, allowing an available space of at least 400 mm from the top of the BSC to the ceiling. Higher versions may be accepted, provided the 400 mm over the BSC is available to measure air velocity above the exhaust filter, and to have enough space for changing the filter and for ducting and/ or a thimble connection to outlet.

### **Internal working area (approximate):**

- For a BS Cof 120 cm (4 Ft.): width 1150 mm x depth 630 mm x height 650-750 mm.
- A BSC of 120 cm (4 ft.) provides the minimal space needed for safe work.
- Inside finish: stainless steel, high quality (e.g. grade 304)
- External housing, including screws, made of stainless steel or equivalent resistant galvanized (zinc-coated) sheet steel, subsequently powder coated and thermally hardened; minimum 80  $\mu$ m thick, or other material that is hard-wearing, resistant to disinfectants and chemicals used in a TB laboratory, and abrasion resistant.
- Vertically adjustable sliding window: aerosol-tight, sliding, safety glass (laminated multilayer safety glass only), thickness  $\geq 6.7$  mm, Counterbalanced.
- High optical transmission, but absorption of UV light; minimal reflection.
- Working aperture:  $\geq 170$  mm measured from work surface to the bottom of the sash window.
- Maximal lifting height of front window: 500 mm.
- Ability to lock the window hermetically for gaseous disinfection for filter decontamination.
- Single-piece working surface with integrated (v-shaped) front air grill.
- Alternative: Working surface as segments
- Noise pressure level:  $\leq 60$  dB A.

### **Internal fittings**

Two plugs, 230 $\pm$ 10V, AC, 50Hz, protected with separate T5A (slow blow) fuse. Voltage and plugs adapted to those used in the country. The line cord/power cord supplied with the equipment



shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

**Warning:** Plugs inside the BSc may differ from the main connection to the electricity network.

Flicker-free, Low-glare, warm-coloured light,  $\geq 1000\text{lux}$ .

Control display on the front of the BSC.

- Electrical control or indicators.
- Electronic fan control.
- Flow meter for air inflow velocity.
- Flow indicator or meter for air down flow velocity.
- Operating hours' indicator (counter).
- Optional: UV light timer.
- Filter and flow conditions

**Optional:** ultraviolet C (UVC) light (253.7 nm wavelength); 30 W with hour counter; with interlock with white light so that the UVC light can be switched on only when the white light source is switched off.

**Optional:**(if a safety gas burner will be used): Gas tap with solenoid valve, optional right or left side. For a laboratory located in a seismic area, gas pipes are not recommended; small gas containers (approximately 200-400 ml) with butane gas directly fixed to the burner to be used instead.

Not necessary when a micro-incinerator is used.

Pre filter construction preferred; easily accessible, filter change without tools preferred.

High-efficiency particulate air (HEPA) filter (exhaust air filter); classification at least H14; conforming with EN 1822; metal framed.

Air down flow velocity:

- NSF 49-2002: Requires compliance with the manufacturer's set points, or down flow velocity with a deviation of  $<0.025\text{ m/s}$  from a nominal set point.
- EN 12469: Airflow velocity should be between  $0.25$  and  $0.50\text{ m/s}$  and is defined by the manufacturer according to the construction. Additionally, no individual measurement should differ by more than 20% of the value requested by the manufacturer within the limits given. Air circulation volume flow (Modify according to the BSC dimensions):
- For a BSC of  $120\text{ cm}$  ( $4\text{ ft.}$ ):  $700\text{-}1200\text{ m/h}$ .
- Influx air velocity:
- According to NSF 49, the average airflow velocity at front aperture should be  $0.51\text{ m/s}$  for class A2.
- EN 12469 does not sub classify within class II BSC. The average airflow velocity at front aperture should be at least  $0.4\text{ m/s}$ , according to the manufacturer's specifications.

Exhaust volume airflow/fresh airflow inward:



For a BSC of 120 cm (4 Ft.): 300-600 m<sup>3</sup>/h.

Blower system to be able to maintain the airflow within a minimum window (narrow limits) on voltage fluctuations. Data to be available on request. Alarms, visible and/or audible, for any unsafe condition of the BSC (e.g. airflow, window position, hardware or software errors). Possibility to shut down alarm for cleaning and maintenance.

### **Electricity requirements**

**Supply voltage:** 230 ±10 V, AC, 50/60 Hz.

Voltage and plugs to be adapted to meet the country requirements. The line cord /Power cord Supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

Lead fuse T16A (slow blow) or circuit breaker B16. The electrical regulations valid in the country of use as well as the relevant connection conditions are required.

**Power consumption** (approximate): Modify according to the BSC dimensions

- For a BSC of 120 cm (4 ft.): 600 W.  
Power consumption for plugs inside: Approximately 1000 W. Note: In areas with frequent breakdown of electricity supply, BSCs with low energy consumption can be an advantage; a UPS with lower capacity can be used. Conform to electrical safety standards IEC 60601-1, UL 61010-1, EN 61010-1, Protection class (in accordance with EN 60529).  
Designed not to interfere with circuit radio (in accordance with EN 55014).

### **Documentation**

#### **Manufacturer's certificate**

The manufacturer must have a management system certified to ISO 9001. The manufacturer to individually test each BSC before shipment. The test report to be provided to the customer, with a duplicate fixed to the BSC. The tests to be performed with research-grade instruments for valid calibration according to test methods outlined in EN 12469 or NSF 49.

The test report to contain at least data on:

- inflow air velocity
- down flow air velocity
- filter leak scan for both filters to document filters' efficiency and integrity.

**Quality and safety standards met by the product to be listed.**

### **Accessories**

**Table or support frame** (support stand) for a working height of 78 ± 2 cm, adjustable at least at three points (feet) to level.

**A telescopic support stand** is advisable for a flexible use.

**Ergonomic laboratory chair**, designed for infectious laboratory areas: -

- adjustable height to suit different users, seat range approximately 400-490 mm
- adjustable-angle back rest (no arm rest)
- caster wheels (five)
- all metal parts chrome plated
- disinfect able with alcohol-containing disinfectants.

**Thimble Ducting:** Air duct construction with thimble to exhaust air from the BSC. The air duct to be made for the BSC offered and fit precisely. A thimble connection (see WHO TB Bio-safety guidelines 2012 page 33) is used with Class II type A2 BSC that is ducted to the outside. The thimble fits over the cabinet's exhaust housing, sucking the air expelled from the cabinet into ducts that lead outside.

A small Opening (usually 5 cm wide) is maintained between the thimble and the cabinet's exhaust housing. This opening enables room air to be drawn into the exhaust ducting system. The thimble must be removable or be designed to allow for operational testing of the cabinet. The power of the external extraction fan installed at the end of the ducting should exceed the volumetric flow rate of each BSC by 30-50%, and should be controlled and connected to an uninterrupted power supply. The air from the BSC should be ducted with ventilation pipes that have a diameter exceed 20 cm.

The extractor fan assembly must be easily accessible and preferably kept at the end of ducting with stable fitting. Ducting design should be straight and number of bands should be minimal, bend should be round shaped (sharp/square bends should not be used). Ducting should have adjustable balancing dampers with easily accessibility so that flow can be controlled as and when required.

## **Operation, maintenance and installation**

### **Operation and maintenance manual**

At least one set of operation, maintenance and service manuals, written in United Nations languages (or at least in English) and preferably also in the official national language of the country requesting the BSC.

### **Installation and maintenance**

-The bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.

- The bidder to also provide user training (including how to use and maintain the equipment) and a comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed over the period of warranty.

-The supplier to provide an after-sale service that covers the whole country. The service to have competent staff, adequate infrastructure and sufficient spare parts to be able to respond to any complaints and to repair or replace the BSC within 14 days.

- Initial on-site testing (aerosol leak test, recirculating air filter, exhaust air filter, airflow measurements inside the BSC and inward/exhaust airflow) to be carried out by a certified expert and certified compliant for satisfactory installation and safe operation. Measurement results to be printed out for documentation in the maintenance record.

### **Warranty starts with certification on site.**

#### **Standard maintenance tools**

All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning equipment, to be included in the offer. Bidders to specify the quantity of every item included in their offer (including items not specified above). If special tools are needed (e.g. to change filters), they must be provided.

Spare parts

- Each assembled BSC to be accompanied by an authorized list of accessories and spare parts.
  - At least one, and preferably two, additional sets of HEPA filters as specified above.
- Warning:** Special clamps may be needed to fix HEPA filter.

**Warranty:** Five years, except for filters and (UV) lamp.

### Remarks

The equipment offered, including its power supply, to be designed and constructed to operate properly and continuously in the conditions of the purchaser's country; the equipment may need to tolerate high humidity (as high as 90% at 35 °C), ambient temperatures of 5-40° C, fungi, and spikes in the electricity supply.

Bidders may propose products additional to the requirements listed above.

Uninterrupted power supply with battery pack for BSC

Description of function and use: The UPS must be used in any settings that have frequent problems in the electricity network (e.g. surges, sags, spikes and blackouts) to assure and back up the function of the BSC, so that any current work can be finalized and all potentially infectious sources closed. If the BSC is connected to a generator, the UPS will maintain the function of the BSC during the time needed for the generator to start and to provide full power.

### Main specifications

- UPS: microprocessor controlled, online continuous transducer, 20 minutes.
- Booster function to regulate up voltage breakdown to 170 V.
- Buck function to regulate down voltage increase up to 280 V.
- Filter to protect against voltage spikes.
- Protection against overload and short circuit.
- Advanced battery check for automated periodic battery inspection. - -
- Indicators for status (e.g. normal function, net down, working on battery, loading battery, battery capacity).
- Sleep mode if item consuming power is shut off.
- Power: 230 V $\pm$  25%, 50 Hz or 60 Hz ( $\pm$  10%) with automatic recognition. -
- Battery: maintenance-free, automatic shut-off before reaching the level of discharge from which recharging to the original capacity will no longer be possible.
- Time for recharging: approximately 4 hours to reach at least 90% of total capacity.
- Outlet voltage: 230 V  $\pm$  3%, 50 or 60 Hz  $\pm$ 0.5% (if the country's standard voltage is 110 V AC, adjustment will be needed).
- Efficiency coefficient: approximately 98%, on battery >85%.
- Noise at 1 m distance <48 dBA.

**Electricity requirements-** Supply voltage: 230  $\pm$  10 V, AC, 50/60 Hz.

Voltage and plugs to be adapted to meet the country requirements. The line cord / Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

**Power consumption:** Approximately 1500 W (may change depending on requirement for the model chosen as well as the extractor fan connected with thimble connection at the outer end).

Protection class (in accordance with EN 60529)

Designed not to interfere with circuit radio (in accordance with EN 55014).

## **Documentation**

### **Manufacturer's certificate**

The manufacturer must have a management system certified to ISO 9001.

**Quality and safety standards** met by the product to be listed.

### **Accessories**

- Battery pack.
- Connection (cable and fittings) for battery pack.
- Battery rack according to the number of batteries.

### **Operation, maintenance and installation**

#### **Operation and maintenance manual**

At least one set of operation, maintenance and service manuals, written in United Nations languages (or at least in English) and preferably also in the official national language of the country requesting the UPS.

#### **Installation and maintenance**

the bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.

The bidder to provide user training (Including how to use and maintain the equipment) and comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed Over the period of warranty.

#### **Standard maintenance tools**

All standard accessories, consumables and parts required to operate the equipment, including all Standard tools and cleaning material, to be included in the offer. Bidders to specify the quantity of every item included in their offer (including items not specified above).

#### **Spare parts**

Each UPS to be accompanied by an authorized list of accessories and spare parts.

**Warranty:** At least two years for UPS; at least five years for battery pack.

## **2. Deep Freezer (-86°C):**

### **Technical Specifications:**

**Upright Model:** CFC free high efficiency double refrigeration system for cooling and freezing filled in the bottom.

**Temperature:** 0-degree C-86-degreeC + - 0.5-degree C (to work in -70 to -86-degree C range)

**Temperature Control:**

- Digital temperature controller (including display at suitable eye level) –
- Microprocessor Control/Microcontroller-for temperature setting
- Alarms for-Voltage, Over heat, Over cool as well as for under temperature, power fail, Door ajar conditions.

**Size:** 300 litres or more (up to 450 L) with minimum 4 Stainless Steel, rust free shelves. Fixed in casters for easy manoeuvrability, Polystyrene Insulated inner doors for the compartments. Pre- coated metal body to prevent environmental damage.

**Electricity:** 230 volts AC, 50Hz single phase. The line cord / Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

**Refrigeration system:** Heavy-duty maintenance free refrigeration system with hermetically sealed refrigeration compressors and reliable cascaded refrigeration to minimize noise and vibration. Air Cooled with security lock to prevent unintentional switch off. Short cooling time of 4 to 5 hours at 40°C ambient temperature. The equipment should be of continuous duty and frost-free. Convenient Air Filter Grill allows easy access for cleaning and changing. Access port for CO2 back up. It shall be fitted with 24x7 temperature recorders/ data loggers which allows for a minimum of 3 GB data storage and the data must be downloadable via a USB port. Deep freezer shall not have an automated defrosting system without a manual override. Audiovisual Electronic Alarm System independent of power supply and Remote alarm contact in case of equipment failure/ power failure Electrically heated doors for quick opening of frozen doors.

Accessories to include suitable boxes and racks for storage of specimen deposits/ DNA extracts/culture isolates in cryo-vials (16 in No. of suitable dimension for 2 ml cryo-vials). Cyro gloves (4 sets wrist length 12") to be provided as part of accessories.

#### **Remarks:**

The apparatus should conform to Indian Standard Institution Guidelines with latest amendments in Indian Standard Specification for Laboratory Deep Freezers or equivalent National or international Standards covering Markings, tests and Safety requirements.

Voltage regulator of appropriate rating to be included to cope with 160-260 V

### **3. Deep Freezer (-20 °C):**

**Description of function and use:** The freezer is mainly used to store strains, enzymes and other temperature-sensitive reagents.

#### **Main specifications**

- 100% CFC- free.
- One-door freezer, to be used as free-standing freezer.
- Capacity (gross): ≥140 litres (up to 200L).
- Dimensions (W xDx H): approximately 60 cm x 60 cm x 85 cm.
- Net (interior compartment) W x D x H: approximately 47 cm x 43 cm 69 cm.
- Cooling system, static.
- Defrosting of freezing compartment, initiated manually.
- Temperature range of freezer compartment: -9 °C to- 25° C which can be set by user tor specific temperature

- Housing material and door: steel, coated, white.
- Polyurethane foam "PUF" insulation
- Door hinges right or left as desired, reversible.
- Fungus-resistant door gasket.
- Door with key lock.
- Adjustable feet for levelling.
- Interior container made of white plastic. At least three shelves with wire baskets.
- Optional- Four shelves in the freezing compartment, at least three closed with a freezing flap.
- External digital temperature display for freezer compartment with high accuracy sensor mounted inside the chamber suspended in air.
- Control panel at the top of cabinet with thermometer Alarms for-Voltage, Over heat, Over cool as well as for under temperature, power fail, Door ajar conditions
- Refrigerant: CFC Free.

### **Electricity requirements**

**Supply voltage:** 230 ± 10 V, AC, 50/60 Hz. Voltage and plugs to be adapted to meet the country requirements. The line cord/ Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

**Power consumption:** Approximately 150 W.

Conform to Indian Electrical Safety standards or International standards like IEC 60601-1, UL 610101, EN 61010-1.

Protection class (in accordance with EN 60529).

### **Documentation**

#### **Manufacturers certificate**

- The manufacturer must have a management system certified to ISO 9001.
- Declaration of conformity to the requirements of standards and regulations of the directives that apply to the product, including energy classification, gas used as refrigerant, climate class.
- One certificate to state that the freezer has been calibrated at the factory.

**Quality and safety standards** met by the product must be listed. **Operation, maintenance and installation**

#### **Operation and maintenance manual**

At least one set of operation, maintenance and service manuals for the freezer, written in United Nations language (or at least in English) and preferably also in the official national language of the country requesting the freezer.

#### **Installation and maintenance**

All prerequisites for installation to be communicated to the purchaser in advance, in detail.

The Supplier to provide an after-sale service that covers the whole country. The service to have competent staff,adequate infrastructure and sufficient spare parts to be able to respond to any complaints and to repair or replace the freezer within 14 days.

#### **Standard maintenance tools**

All Standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning material, to be included in the offer. Bidders to specify the quantity of every item included in their offer (including items not specified above).

**Spare parts**

Each freezer to be accompanied by an authorized list of accessories and spare parts.

**Warranty:** Five years.

**Remarks**

The equipment offered, including its power supply, to be designed and constructed to operate properly and continuously in the conditions of the purchaser's country; the equipment may need to tolerate high humidity (as high as 90% at 35 °C), ambient temperatures of 5-40 °C, fungi, and spikes in the electricity supply.

Bidders may propose products additional to the requirements listed above

## **4. Electric micro-incinerator for loops:**

**Description of function and use:** The micro-incinerator allows sterilization of metal inoculating loops without a flame and is suited for work in a BSC.

**Main specifications**

- Heating element of ceramic surrounded by isolating cover.
- Quick infrared heating to temperatures  $\geq 800$  °C for fast sterilization.
- Stand with suction-cup feet (or equivalent) for stable, safe operation.
- Possibility to fix the incinerator to a stand at different angles.
- Electricity requirements: Supply voltage:  $230 \pm 10$  V, AC, 50/60 Hz. Voltage and plugs to be adapted to meet the country requirements.
- The line cord/ Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.
- Power consumption: 2000 W,
- Protection class (in accordance with EN 60529).
- Designed not to interfere with circuit radio (in accordance with EN 55014).

**Documentation**

Manufacturer's certificate

- The manufacturer must have a management system ISO 9001.
- Certificates to be provided for each item supplied.
- Quality and safety standards met by the product must be listed.

**Accessories**

Attached loop holder.

**Operation, installation and maintenance**

Operation and maintenance manual: At least one set of operation, maintenance and service manuals, written in United Nations languages (or at least In English) and preferably also in the official national language of the country requesting the micro-incinerator.

Installation and maintenance



The bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail..

The bidder to also provide user training (including how to use and maintain the equipment comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed over the period of warranty.

The supplier to provide an after-sale service that covers the whole country The service to have competent staff, adequate Infrastructure and sufficient spare parts to be able to respond to any complaints and to repair or replace the micro-incinerator within 14 days  
Standard maintenance tools

All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning material, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).

### **Spare parts**

The equipment offered, including its power supply, to be designed and constructed to operate properly and continuously in the conditions of the purchaser's country; the equipment may need to tolerate high humidity (as high as 90% at 35 °C), ambient temperatures of 5-40 °C, Fungi and spikes in the electricity supply  
Bidders may propose products additional to the requirements listed above.

## **5. Microliter Centrifuge:**

**Description of function and use:** The centrifuge is designed for routine. applications in molecular biology.

### **Main specifications**

- Robust metal housing; compact design with chemical-resistant (coated) housing.
- Easy-to-clean, smooth rotor chamber that is resistant to acids, alkalines, disinfectants used in the laboratory.
- Low access height (≤23 cm) and space-saving design (≤24 cm x 32 cm; W x D).
- Microprocessor-controlled centrifuge.
- Standard rotor with a capacity of at least 18 positions; aerosol-tight (chemical-resistant coated); exchangeable.
- Maintenance-free motor.
- Maximal relative centrifuge force: 15,000 G.

- Automatic lid lock, starting with and during run of rotor.
- Option: Automatic opening at the end of the run.
- Emergency unlock for electricity blackout.
- LCD display; protected; showing time and relative centrifugal force or speed in rcf or rpm.
- Speed adjustable in 100 rpm steps.
- If a keypad is used, it should be foil protected.
- Timer for runs between 30 seconds and 30 minutes and an option for continuous operation for longer runs.
- Short time operation by pressing a time button for short spin.
- Adjustment of running time in steps of 30 seconds.
- Short acceleration time to maximum rcf in  $\leq 20$  seconds.
- Short breaking time from maximum rcf in  $\leq 20$  seconds.
- Noise level:  $\leq 58$  dbA.

### **Electricity requirements**

**Supply voltage:** 230  $\pm$  10 V, AC, 50/60 Hz.

Voltage and plugs to be adapted to meet the country requirements. The line cord / Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

Power consumption: Approximately 250w.

Conform to electrical safety standards IEC 60601-1, UL 61010-1, EN 61010-1.

Protection class (in accordance with EN 60529).

Designed not to interfere with circuit radio (in accordance with EN 55014).

Documentation

Manufacturer's certificate

The manufacturer must have a management system certified to ISO 9001. One certificate to state that the centrifuge has been calibrated at the factory.

**Quality and safety standards** met by the product must be listed.

### **Accessories**

Optional: Adapter set; reduction device for smaller tubes to be centrifuged in the standard rotor to maximal rotor capacity, for 0.5 ml and 0.2 ml tubes Operation, maintenance and installation

### **Operation and maintenance manual**

At least one set of operation, maintenance and service manuals written in UN languages (or at least in English) and preferably also in the official national language of the country requesting the microlitre centrifuge.

### **Installation and maintenance**

The bidder must arrange for the equipment to be installed by certified or qualified personnel, any prerequisites for installation to be communicated to the purchaser in advance, in detail. The bidder to also provide user training (including how to use and maintain the equipment) and a comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed Over the period of warranty. The supplier to provide an after-sale service that covers the whole country. The service to have competent staff, adequate infrastructure and sufficient spare parts to be able to

respond to any complaints and to repair or replace the microlitre centrifuge within 14 days.

#### **Standard maintenance tools**

All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication material, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above). **Spare parts**

Each centrifuge to be accompanied by an authorized list of accessories and spare parts.

#### **Warranty: Five years**

#### **Remarks**

The equipment offered, including its power supply, to be designed and constructed to operate properly and continuously in the conditions of the purchaser's country; the equipment may need to tolerate high humidity, ambient temperatures of 5-40 °C, fungi, and spikes in the electricity supply. Bidders may propose additional products to the requirements listed above

### **06. PCR Thermocycler:**

**Description of function:** The thermos cycler is used in a TB laboratory to amplify fragments mycobacterial DNA.

#### **Main specifications**

Thermocycler, Peltier elements.

Block for 96 x 0.2 ml tubes; possibility to use block with 48 x 0.5 ml tubes and 96-well PCR plates.

Blocks must be resistant to oxidation.

Heating rate: 4 °C/s.

Cooling rate: 2 °C/s.

Temperature range (block): 4-100 °C.

Regulating accuracy for block temperature:  $\pm 0.1$  °C.

Temperature uniformity at 70 °C (block):  $\pm 0.4$  °C.

Internal memory for at least 50 programs with up to 99 steps/program, freely editable.

Heatable lid with automatic height adaptation.

Electromechanical lid blocking to prevent accidental opening during a run. Temperature range for lid: 80 °C to at least 103 °C .

Optional: Interface for remote control via PC; activated RS 232 serial port.

#### **Electricity requirements**

**Supply voltage:** 230 $\pm$  10 V, AC, 50/60 Hz.

Voltage and plugs to be adapted to meet the country requirements. The line cord / Power cord Supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

**Power consumption:** Approximately 500 W.

Conform to electrical safety standards IEC 60601-1, UL 61010-1, EN 61010-1. Protection class (in accordance with EN 60529). Designed not to interfere with circuit radio (in accordance with EN 55014).

#### **Documentation**

##### **Manufacturer's certificate**

The manufacturer must have a management system certified to ISO 9001. One certificate to state that the thermocycler has been calibrated at the factory and certified according to ISO 13485 quality regulations.

**Quality and safety standards** met by the product must be listed.

##### **Operation, maintenance and installation**

##### **Operation and maintenance manual**

At least one set of operation, maintenance and service manuals written in UN languages (or at least in English) and preferably also in the official national language of the country requesting the thermocycler.

#### **Installation and maintenance**

The bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.

The bidder to also provide user training (including how to use and maintain the equipment) and a comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed over the period of warranty. The supplier to provide an after-sale service that covers the whole country.

The service to have competent staff, adequate infrastructure and sufficient spare parts to be able to respond to any complaints and to repair or replace the thermocycler within 14 days.

##### **Standard maintenance tools**

All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders to specify the quantity of every item included in the offer (including items not specified above).

##### **Spare parts**

Each thermocycler to be accompanied by an authorized list of accessories and spare parts. Set of fuses, if used separately in the instrument. **Warranty:** Five years.

##### **Remarks**

The equipment offered, including its powersupply, to be designed and constructed to operate Properly and continuously in the conditions of the purchasers country; the equipment may need to tolerate high humidity (as high as 90% at 35 °C), ambient temperatures of 5-40° C, fungi, and spikes in the electricity supply.

Bidders may propose products additional to the requirements listed above.

## **07. PCR Workstation:**

**Description of function and use:** The workstation is used to prepare reagents under clean conditions to avoid contamination with DNA

**Main specifications**

Exterior dimensions (Hx Wx D): approximately 700 mm x 750 mm x 600 mm.

interior working area (W x D): approximately 700 mm x 500 mm.

Exterior: stainless steel or powder-coated metal

interior: stable formed stainless steel.

Side panels transparent, able to absorb wavelengths below 400 nm. Overhead Uv light for DNA decontamination; two lamps, 25 W each Separate, switchable, UV air-sterilizing circulation unit; UV lamp (25 W). Timer and key lock for Uv lamp; timer operates only when key lock is on. Overhead white light; 15 W; at least 800 lux.

At least two plug outlets built into the chamber; AC 230 ±10V; 50 Hz; 5A fuse.

**Electricity requirements**

**Supply voltage:** 230±10 V, AC, 50/60 Hz. Voltage and plugs to be adapted to meet the country requirements. The line cord/ Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

**Power consumption:** Depends on the electrical equipment used inside the workstation; maximum 1200 W.

Conform to electrical safety standards IEC 60601-1, UL 61010-1, EN 61010-1.

Protection class (in accordance with EN 60529).

Designed not to interfere with circuit radio (in accordance with EN 55014).

**Documentation****Manufacturer's certificate**

The manufacturer must have a management system certified to ISO 9001 and a type-test certificate of relevant optical and mechanical tests.

**Quality and safety standards** met by the product must be listed. **Operation, maintenance and installation**

**Operation and maintenance manual**

At least one set of operation, maintenance and service manuals written in UN languages (or at least in English) and preferably also in the official national language of the country requesting the workstation.

**Installation and maintenance**

The bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.

The bidder to also provide user training (including how to use and maintain the equipment) and a comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed over-the period of warranty. The supplier to provide an after-sale service that covers the whole country. The service to have competent staff, adequate infrastructure and sufficient spare parts to be able to respond to any complaints and to repair or replace the workstation within 14 days.

**Standard maintenance tools**

All Standard accessories, consumables and parts required to operate the equipment, including all Standard tools and cleaning and lubrication materials, to be Included in

the offer, Bidders to specify the quantity of every item included in their offer (including items not specified above)

**Spare parts**

Each workstation to be accompanied by an authorized list of accessories and spare parts. Set of fuses for the workstation. Two UV lamps

**Warranty:-** Five years.

**Remarks**

The equipment offered, including its power supply, to be designed and constructed to operate properly and continuously in the conditions of the purchaser's country; the equipment may need to tolerate high humidity (as high as 90% at 35 °C), ambient temperatures of 5-40 °C, fungi, and spikes in the electricity supply.

Bidders may propose products additional to the requirements listed above.

**08. pH Meter:**

**Technical Specifications:**

- pH range 0-14 with digital display and stand by and calibration mode;
- Bench top with shielded electrode bulb and waterproof housing.
- Temperature compensation should be provided.
- calibration with at least three standard calibration buffers (pH 4.0,7.0, 10.0)
- Resolution: 0.01 pH units.
- Accuracy:  $\pm 0.01$  pH units.
- No. of Display Digits: Three
- Supply voltage:  $230 \pm 10$  V, AC, 50/60 Hz, Voltage and plugs to be adapted to meet the country requirements. The line cord/Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.
- The manufacturer must have a management system certified to ISO 9001 - Provision of spare electrode
- A certificate to state that the pH meter has been calibrated at the factory.

**09. Refrigerated Centrifuge with UPS**

**Description of function and use:** The centrifuge will be used in a TB laboratory to spin down mycobacteria in liquefied, decontaminated materials. The compartments containing the centrifuge tubes (buckets) need to be absolutely aerosol tight and closed by transparent lids. At least 3000 G will be required for efficient sedimentation within 20 minutes of centrifugation. The centrifuge should preferably be a desk/bench top one.

**Main specifications**

- Metal housing, chamber stainless steel.
- Cooling capacity at maximum speed at  $+4^{\circ}\text{C}$ .
- Standby cooling.
- Induction motor, brushless drive.

- Rotor with swing-out buckets, at least four positions.
- Rotor buckets with aerosol-tight, transparent, clipping lids.
- Inserts for buckets adapted to 50 ml centrifuge tubes, conical.
- Capacity: about 16 or more x 50 ml,
- Maximum revolutions per minute: approximately 4500 rpm, corresponding to a radius of approximately 15 cm.
- Relative centrifugal force (rcf): approximately 3400
- Noise level at maximum speed: not more than 60 dbA
- programmable for all parameters (switchable between rpm and rcf) of arun large display.
- Soft start and different acceleration levels (1-9); different braking levels (1-9) and brake force cut-off.
- Input and recall of programs; at least 20 storage positions.
- LCD display (protected against splash of liquids) for indication of run time, speed (rpm) or rcf (after entering centrifugation radius)-switchable, actual temperature, time left to finish run.
- Imbalance switch-off.
- Motor overheating protection.
- Chamber overheating protection.
- Rotor recognition for appropriate over-speed protection.
- Safety lock of lid during run and as long as the rotor is moving.
- Possibility of mechanical opening of lid if there is current blackout.

### **Electricity requirements**

Supply voltage: 230 ± 10 V, AC, 50/60 Hz. Voltage and plugs to be adapted to meet the country requirements. The line cord /Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

Power consumption: Approximately 1800 W.

Conform to electrical safety standards IEC 60601-1, UL 61010-1, EN 61010-1.

Excess-voltage category II.

Protection class (in accordance with EN 60529).

Designed not to interfere with circuit radio (in accordance with EN 55014).

### **Documentation**

Manufacturers certificate: The manufacturer must have a management system certified to ISO 9001. One certificate to state that the centrifuge has been calibrated at the factory. Quality and safety standards met by the product to be listed. Operation, maintenance and installation

### **Operation, maintenance and installation**

**Operation and maintenance manual:** At least one set of operation, maintenance and service manuals, written in United Nations languages (or at least in English) and preferably also in the official national language of the country requesting the centrifuge.

**Installation and maintenance** The bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail. The bidder to also provide user



training (including how to use and maintain the equipment) and a comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed over the period of warranty. The supplier to provide an after-sale service that covers the whole country. The service to have competent staff, adequate infrastructure and sufficient spare parts to be able to respond to any complaints and to repair or replace the centrifuge within 14 days.

#### **Standard maintenance tools**

All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning equipment, to be included in the offer. Bidders to specify the quantity of every item included in their offer (including items not specified above).

#### **Spare parts**

Each centrifuge to be accompanied by an authorized list of accessories and spare parts. Device for mechanical opening of the centrifuge after automatic blocking as consequence of currency blackout. Lubricants for movable parts and gaskets.

Spare fuses.

#### **Uninterrupted power supply with battery pack for Refrigerated Centrifuge**

**Description of function and use:** The UPS must be used in any settings that have frequent problems in the electricity network (e.g. surges, sags, spikes and blackouts) to assure and back up the function of the Refrigerated Centrifuge, so that any current work can be finalized and all potentially infectious sources closed. If the Refrigerated Centrifuge is connected to a generator, the UPS will maintain the function of the Refrigerated Centrifuge during the time needed for the generator to start and to provide full power.

#### **Main specifications**

UPS :microprocessor controlled, online continuous transducer, 20/30 minutes.

Booster function to regulate up voltage breakdown to 170 V.

Buck function to regulate down voltage increase up to 280 V.

Filter to protect against voltage spikes.

Protection against overload and short circuit.

Advanced battery check for automated periodic battery inspection. Indicators for status (e.g. normal function, net down, working on battery, loading battery, battery capacity).

Sleep mode if item consuming power is shut off.

Power: 230 V  $\pm$  25%, 50 Hz or 60 Hz ( $\pm$  10%) with automatic recognition. Battery maintenance-free, automatic shut-off before reaching the level of discharge from which recharging to the original capacity will no longer be possible.

Time for recharging: approximately 4 hours to reach at least 90% of total capacity .

Outlet voltage: 230 V  $\pm$  3%, 50 or 60 Hz  $\pm$  0.5% (if the country's standard voltage is 110 v AC, adjustment will be needed).

Efficiency coefficient: approximately 98%, on battery >85%.

Noise at 1m distance <48 dBA.

Permissible ambient temperature and relative humidity: 0-40° C and 95% (not condensing).

#### **Electricity requirements**

**Supply voltage:** 230  $\pm$  10 V, AC, 50/60 Hz.

Voltage and plugs to be adapted to meet the country requirements. The line cord / Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

**Power consumption:** Approximately 3000 W (depending on the model chosen).

Protection class (in accordance with EN 60529)

Designed not to interfere with circuit radio (in accordance with EN 55014).

### **Documentation**

#### **Manufacturer's certificate**

The manufacturer must have a management system certified to ISO 9001. **Quality and safety standards** met by the product to be listed.

### **Accessories**

Battery pack.

Connection (cable and fittings) for battery pack.

### **Operation, maintenance and installation**

#### **Operation and maintenance manual**

At least one set of operation, maintenance and service manuals, written in United Nations languages (or at least in English) and preferably also in the official national language of the country requesting the UPS.

### **Installation and maintenance**

The bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.

The bidder 'to provide user training (including how to use and maintain the equipment) and a comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed over the period of warranty.

### **Standard maintenance tools**

All Standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning material, to be

included in the offer. Bidders to specify the quantity of every item included in their offer (including items not specified above).

### **Spare parts**

Each UPS to be accompanied by an authorized list of accessories and spare parts.

Warranty: Five years for UPS; Five years for battery pack

## **10. Water Bath:**

### **Technical Specifications:**

- Stainless Steel, insulated double walled,
- Inner wall of stainless steel,
- Thermostatic temp. control from ambient to 85- 90° C (A0.50) complete with immersion heater,
- Aluminium /SS cover,
- Brass drain cock,

-220-240 Volts AC, 50Hz. The line cord/ Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

-Dimensions outside:- (approx.)36x41x25 cms; inside:- (approx.)27x30x15 cms; Capacity not given because the dimensions specified are required for the proper functioning of the water bath in the Laboratory for optimum utilization of the space.

- Power: (approx.) 480W;

- Digital microprocessor display to set temperature point preventing thermal runaway.

-Seamless reservoir with no welds to leak or rust, see- through cover is hinged and removable, and steeply gabled to accept taller samples.

## **11. Water Distiller**

### **Technical Specifications:**

-Complete unit made in stainless steel.

-Constant level arrangement for inlet water

-Wall hanging model, provided with MS Bracket for wall mounting

-Boiling Chamber, Lid, Cooling Condenser and any other part coming into contact with water/steam should be made of 304 quality stainless steel.

- 'Heating is achieved with kettle type immersion heaters (two)'.

-Electricals: should work on 220/440 volts, AC Mains, the line cord/ Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

-Capacity of the plant should be **5 litres (approx.) per hour**, should be provided with necessary safety features such as automatic disconnection of heater connections.

## **12. BD BACTECTM MGITTM 960 System**

Non-radiometric, fully automated system (annual capacity of 2,700 tubes) for the rapid detection of mycobacteria in clinical specimens, other than blood, as well as antimicrobial susceptibility testing of anti-tuberculosis drugs known as S.I.R.E, I.R., and PZA Includes Instrument, Starter Kit, AST Starter Kit, Printer and UPS, Installation, Training and Technical support, with comprehensive instrument service including replacement parts for three years.

### **13. MGIT Accessories:**

A)

a. AST Carrier Set (5-tube)-3 per unit, BD's designated catalog number 445943

b. AST Carrier Set (4-tube)-3 per unit, BD's designated catalog number 445944

C. AST Carrier Set (3-tube)-3 per unit, BD's designated catalog number 445945

d. AST Carrier Set (2-tube)-3 per unit, BD's designated catalog number 445946

e. AST Carrier Set (8-tube)-3 per unit, BD's designated catalog number 445993

### **B) GT Blot 48 List Nr. 1003/1 (Hain Life Sciences, GMBH)**

with maintenance service contract including technical support, calibration and replacement parts for three years

**C) Twincubator® List Nr. 7025009 (Hain Life Sciences, GMBH)**

with maintenance service contract including technical support, calibration and replacement parts for three years

**14 .Thermometers:**

**Description of function and use:** For TB C&DST Labs, we need five different types of thermometers for various devices as shown below:

- a) Electronic Thermo Hygrometer (Two) for at least two rooms (Expert room and for Media Preparation room+ TB Containment if not available) as these rooms require monitoring of temperature and humidity.
- b) Alcohol Thermometer (Eight) for refrigerator, incubator, WII and WIC. A total of 8 will be required assuming there will be 3 refrigerators, 2 incubators, 1 WII and WIC each and 1 spare.
- c) Digital Maxima- Minima Thermometer- two ranges (one each) for calibration of
  - a.-100 to +60 deg C (for -80 deep freezer, -20 deep freezer, refrigerator, WIC)
  - b. 0 to +300 (for incubator, oven, WII, autoclave, GT Blot, Twincubator, Thermocycler)
- d.)Dial Spring Thermometer (one) For Hot Air Oven used in DNA extraction, we need dial spring thermometer

The thermometers are mercury free which helps avoid unnecessary emergency management in case of breakage/spill of mercury

**15. Electronic max-min thermo-hygrometer**

Use :Electronic Thermo Hygrometer is used for at least two rooms (Xpert room and for Media preparation room+ TB Containment if not available) as these rooms require monitoring of temperature and humidity.

Features:

- Recording of MAX/MIN Humidity & Temperature value automatically
- Alarm function (optional)
- Calendar function with months and years by pushing ADJ button(optional) -Back stand & hanging hole to be provided with each thermometer

**Technical Specification:**

- Temperature Range: -10~60°C (-14F ~122F)
  - Humidity Range: 20% ~99% RH
- Resolution: Temperature: 0.1 °C (0.1 °F), Humidity: 1%RH
- Accuracy: Temperature: ±1 °C (1.8 °F), Humidity: ±3%RH (40%-80%)
- LCD size: Approx. 80 x 60 mm
- Memory of MAX/MIN value of temperature & humidity
- Optional:2 modes for time showing: 12 or 24 (changeable by using MODE & ADJ button to set)
- Accessories: Operational Manual, one set of battery
- Power Supply: battery supply

- Warranty Period: Five year, Warranty starts with installation at site.
- Factory Calibrated Certificate: Equipment must be supplied with valid calibration certificate stating that the equipment has been calibrated at factory.

## **16. Glass Alcohol Thermometer:**

Use: Glass Alcohol Thermometer is used for monitoring temperature of refrigerator, incubator, WII and WIC.

### **Technical Specification:**

- Range: 0°C to +100°C
- Accuracy:  $\pm 1.0^\circ\text{C}$  or better within the range 0°C to +70°C
- Resolution:  $\pm 0.5^\circ\text{C}$
- Sensor: Coloured alcohol in glass column.
- Unit of measurement: Temperatures must be displayed in degrees Centigrade only.
- Casing Non-corrodible plastics or metal case. The glass column must be protected against breakage and strongly supported so that the column cannot be displaced more than 0.5mm vertically with respect to the scale.
- Overall dimensions: Maximum 200 x 25 x 25mm.
- Weight: The product must be fully portable.
- Easily readable centigrade scale
  - with a minimum space of 1 mm between each line
  - Long lines with numbers: for each 10 degrees.
  - Short lines: for even numbered degrees
  - Shorter lines: for odd numbered degrees.
  - Font: high-legibility font.
  - Reading angle: between 80 and 100° to the plane of the support plate.
  - Colour of markings: dark blue or black on a white background
- **Warranty Period:** 05 year, Warranty starts with installation at site

## **17. Digital Maxima- Minima Thermometer (-100 to 60 °C)**

Use: -100 to +60 deg C Digital Maxima- Minima Thermometer is used for calibration for -80 deep freezer, -20 deep freezer, refrigerator, WIC

### **Technical Specifications:**

- Measuring range:-100 °C to +60 °C
- Resolution: 0.1°C
- Accuracy:  $\pm 0.1^\circ\text{C}$
- Precision: $\pm 0.2\%$  full scale
- Flexible probe (Preferably K type) Thermocouple, waterproof cable, cold resistance.
- Measuring frequency: 3 measuring per second
- Display: Digital LCD display
- Min-/Max-value memory(optional)
- Operating key: ON/OFF key, Min-max value display, Hold key
- Power supply: 9V battery type
- Low battery warning
- Auto off function

- Calibration certificate: Multi point Calibration Certificate must be submitted with traceability to ISO/IEC 17025 standards
- Warranty Period: 1 year, Warranty starts with installation at site.
- Accessories: Spare one set of battery

### **18. Digital Maxima- Minima Thermometer (0 to 300 °C)**

**Use:** 0 to +300 deg C Digital Maxima- Minima Thermometer is used for calibration for incubator, oven, WII, autoclave, GT Blot, Twincubator, Thermocycler

#### **Technical Specifications:**

- Measuring range: 0 °C to +300 °C
- Resolution: 0.1°C
- Accuracy: ±0.1°C
- Precision: ± 0.2 % full scale
- Flexible probe (Preferably K type) Thermocouple, waterproof & Heat resistance cable/probe.
- Measuring frequency: 3 measuring per second
- Display: Digital LCD display
- Min-/Max-value memory optional.
- Operating key: ON/OFF key, Min-max value display, Hold key
- Power supply: 9V battery type
- Low battery warning
- Auto off function
- Calibration certificate: Multi point Calibration Certificate must be submitted with traceability to ISO/IEC 17025 standards
- Warranty Period: Five year, Warranty starts with installation at site.
- Accessories: One set of battery

### **19. Dial Thermometer:**

**Use:** Dial Spring Thermometer is used for monitoring temperature of for Hot Air Oven used in DNA extraction

#### **Technical Specifications:**

- Temperature Range: +50 °C to +300° C
- Accuracy: ±1 % Full Scale. Reading stabilization within 1 min
- Casing Material: Optically clear strong glass cover with Stainless steel surface mount (highly resistance to corrosion)
- Dial: 2 inch.
- Type: Bimetallic dial thermometer
- Weight: Not critical, provided the product is fully portable
- Easily readable centigrade scale with
  - Long lines with numbers: for each 10 degrees.
  - Short lines/dots: for each 5 degrees.
- Font: high-legibility font

- Colour Marking: dark black on a white background with Red pointer/dark red on a white background with black pointer
- Factory Calibrated Certificate: Equipment must be supplied with valid calibration certificate stating that the equipment has been calibrated at factory.
- Warranty period: Five year, warranty starts with Installation at site.

Provide Performa Invoice of sterilization Machine to avoid chemical fumigation for Operation Theatre.

|                       |   |
|-----------------------|---|
| <b>Specifications</b> |   |
| Power (Overall)       | 500 Watts & 50 Watts for Sterilization Machine  |
| Individual Lamp Power | 30 Watts & 15 Watts for Sterilization Machine   |
| Number of Lamps       | 16 Lamps for Sterilization Machine  |
| Choke Type            | 2 x 36 Watts choke(electronic) for Sterilization Machine.<br>3x 18 Watts choke (electronic) for Sterilization Machine.        |
| Life                  | 8000 hours  |
| Replacement           | Yes (applicable with service fee)   |
| Light Wavelength      | UVC Sterilization at 254 nm( Germicidal)  |
| Filter                | HEPA Filter (0.3uM)   |
| Lamp duration         | Normal operation of about 10 minutes max per use  |
| Room coverage         | 1500 sq. ft. for Sterilization Machine  |
| Safety                | Remote /Local emergency stop. Automatic closure for lamp safety motion deletion sensors for human safety                      |
| Protection            | Motion sensor ( 360 degree) integrated technology and lamp close automatically when the human intervenes in the contract area |
| Remote Operation      | Android phone/IOs/Tablet interface. Intelligent app to set all the equipment operating parameters                             |
| Usage Information     | Total hours of utilization and the information available to the use   |
| Ionizer               | 5x 10 <sup>6</sup> ions/cm  |
| <b>Electronics</b>    |   |
| Serviceability        | Yes (applicable with service fee)   |
| Controller            | Microcontroller 16-bit Higher Performance   |
| <b>Mechanical</b>     |   |
| Mirrors               | Stainless steel polished surface for maximum reflectivity   |
| Wheels                | Medical grade wheel with brake  |
| Size                  | 530 x530 x1500 mm and 220 x 620 mm for Sterilization Machine  |
| Weight                | 40 kg gross weight for Sterilization Machine  |
| Manoeuvrability       | 4 independent wheel to access tight comers  |
| <b>Electrical</b>     |   |
| Input voltage         | 230 Volt AC& 50Hz for other voltages, specify at the time of purchase   |
| Maximum current       | 2.5 Amps  |
| Power Chord Length    | 2 meters  |