



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House, Bemina (190018), Srinagar

email: mdjkscl2@gmail.com; website: www.jkmsclbusiness.com

M/s Krishna Medi Equipment,
New Delhi. (Bidder)

No: JKMSCL/GM(Adm.)/2025/5178-92

Dated: 23-08-2025.

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" - issuance of Rate Contract thereof for the item "(Anaesthesia Work Station (Basic))".

Ref. No.

1 NIT/JKMSCL/M&E/2023/605 dated 19-10-2023

2. Minutes of Purchase Committee vide No. JKMSCL/PS/MD/2025-26/395-407 dated 12.08.2025.

3. LOI No. JKMSCL/GM(Adm)/2025/4662-65 dated 18.08.2025.

Dear Sir,

In accordance with the decision of the Purchase Committee vide its minutes mentioned in reference cited above, the Jammu and Kashmir Medical Supplies Corporation Limited is pleased to inform that your firm has been declared as L1 for purchase of following item(s)/equipments which are given below:

S.No.	Name of item	Currency	Basic Rate without Tax with conversion (1)	Taxes (2)	Total Amount (in Rs) with taxes (1+2)
1.	Anaesthesia Work Station (Basic) M/S Cardiolabs Healthcare India Pvt Ltd., Punjab. (Manufacturer) Model No. C-FLOW 10C	INR	995000.00 Consumables 9500.00	119400.00 1140.00	11,14,400.00 (Rupees Eleven Lakhs Fourteen Thousand Four Hundred Only) Consumables 10640.00

The approved rates for Comprehensive Maintenance Contract after the expiry of warranty period of 05 (Five) years are as follows:


S.No	Name of Item	CMC for five years
1	Anaesthesia Work Station (Basic)	6 th yr : 58500.00 7 th yr : 64500.00 8 th yr : 70800.00 9 th yr : 78400.00 10 th yr : 86500.00 Total : 358700.00

Terms and Conditions:

- Delivery Site:** Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
- Validity:** The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
- Taxes:** As mentioned above and as applicable.
- Warranty:** Two Years from the date of installation.
- Performance Security**
You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
- Supplies**
Within 60 days from the date of issuance of supply order.
- Penalty**
In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at the rate of 0.25% per day for every day of delay subject to maximum of 10%.


8. **Installation:** Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.
 9. **Availability of Spares:** Firm shall certify that the parts of the equipment shall remain available at least for a period of **five years** after the year of expiry of warranty period of two years.
 10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.
 11. The supplier/manufacturer/importer shall ensure to supply the Brand new and unused item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.
 12. JKMSCL is at liberty to get the feedback from the end-user department/intending department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.
 13. **Fall Clause:** The prices under contract shall be subject to price fall clause. The prices charged for the goods supplies under the contract by successful bidder shall in no event exceed the lowest price at which the successful bidder sells the goods or offers to sell the goods of identical description to any other person(s)/Organization(s) including procurement agencies or any other department of Central Govt. or State Govt. as the case may be during the period till the performance of the contract. If at any time, during the period of the contract, the bidder reduces the sales price chargeable under the contract, he shall forth with notify such reduction to the JKMSCL, J&K and the price payable under the contract for the stores supplied after the date of coming into force of such reduction or sale shall stand reduced correspondingly. It imply that if the contract holder quotes/ reduces its price to render similar goods at a price lower than the contract price to anyone in the Central/State Govt. at any time during the currency of contract including extension period, the contract price shall be automatically be reduced with effect from the date of reducing or quoting lower price for all delivery of subject matter of procurement under contract and the contract shall be amended accordingly.
 14. **Risk Purchase:** If the bidder is unable to complete the supply within the specified or extended period, the corporation shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the bidder on his (i.e., bidders) account at his cost and risk, with the prior approval of Managing Director JKMSCL, J&K. The bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the bidder.
The bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the bidder under this or any other contract with the corporation/government. If recovery is not possible from the bill and the bidder fails to pay the loss or damage within one month of the demand, the recovery of such amount or sum due from the bidder shall be made from the bidder. In case supplier fails to deliver ordered goods, the risk purchases may be made at market rate from any other firm. It is mandatory for the approved supplier to acknowledge receipt of orders within seven days from the date of dispatch of order, failing which the procuring entity will be at liberty to initiate action to purchase the items on risk purchase provision at the expiry of the prescribed supply period.
 15. **Orders shall only be placed on the basis of actual requirement and funds as available under the relevant Head of Accounts.**
 16. **Payment shall be released after successful installation and commissioning of the equipment duly verified by the intending department.**
 17. **Approved technical specifications shall remain same as per NIT (Annexure I)**
- All the terms and conditions of the NIT, Standard Procurement of Procedure (SPP) of JKMSCL & LOI shall be part of this Rate Contract.**

Issued with the Decision and Authorization of Purchase Committee of JKMSCL vide its minutes No. JKMSCL/PS/MD/2025-26/395-407 dated 12.08.2025.


Yours sincerely,
23.8.25
(Mohd. Ashraf Choudhary) JKAS
General Manager (Adm.), JKMSCL.

Copy for information to the:

1. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
2. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).

3. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
 4. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
 5. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
 6. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
 7. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
 8. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
 9. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
 10. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
 11. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
 12. General Manager -K (P&S, IT), J&K Medical Supplies Corporation Limited.
 13. Asstt. Programmer to upload on official website.
 14. I.T cell JKMSCL.
 15. File
- 

3. Anaesthesia Ventilator should be suitable for New Born, paediatric and adult and should have colored touch screen with 10" screen size or more.
4. Display should have at least any 2 waveforms of Paw vs. time/flow vs time/ Vol vs, time and 2 spirometry Loops (PV & FV) simultaneously.
5. Unit should have tidal volume compensation /fresh gas compensation / fresh gas decoupling valve.
6. Should monitor and display the measure value of Minute volume. Tidal volume, flo2 concentration. Peak pressure, Mean pressures, Plateau and PEEP.
7. Should have timer to calculate total anaesthesia time with Case Start and Stop functional key
8. Should have adjustable high/ low limits setting for FIO2, Expired tidal volume, Minute volume, frequency and airway pressure.
9. Should have a battery backup at least 180 minutes or more.
10. Ventilator should have upgradable facility for monitoring of Multi Gas Analysis of all Anaesthetic Agent, MAC Value and Etco2 and should be quoted extra as optional.
Vaporizer
1. Vaporizer must be isolated from Gas Flow in off position and prevent simultaneous activation of more than one vaporizer.
2. Temperature pressure should be compensated and flow should be compensated by independent vaporizer.
3. The Unit should have dual position selectate bar for vaporizer and should be supply with isoflurane and sevoflurane vaporizers.
The Anesthesia Work Station should be supplied Monitor
1. It should have Monitor of at least 15" touch screen colour display minimum of 6 or more waveform at a time
2. It should be touch screen and knob controlled.
3. Monitor should be able to measure following parameter: 3.1.3/5 lead ECG SPo2, NIBP, Respiration, Temperature, End tidal Co2(ET CO2) and Bispectral index (BIS) Monitoring System .
3.2.It should also have electro cautery and defibrillator filter.
3.3.Should have audio and visual alarms
3.4. It should have Trends up to 48 hours
3.5. Should have the Battery backup of more than 2 hr.
Monitor be supplied with following standard accessories.
1. Reusable Spo2 Sensors Scach for adult - 01 Nos
2. ECG cable 3/5 Leads - 1 no's
3. NIBP cuff for adult and paediatric age group - 1 No each.
4. Temperature probe - 1 Nos
5. ECG Disposable Electrodes - pack of 30 pc
6. Power Cable.
The Anesthesia Work Station should be supplied with the following accessories.
1. Main unit with two vaporizer, Anesthesia Ventilator and Monitor.
2. Disposable breathing circuit for adult - 5 units
3. Reusable / autoclave breathing circuit for adult - 1 no's
4. Face Mask 01 each (small, medium, large size)
5. Air, N2O & O2 supply hose
6. Breathing bag - 1 each (1L and 2L).

7. Power Cable

8. Scavenging System should be available with work station.

1. Unit should be BIS/ISO13485/CDSCO (Registered) and should confirm to EN 60601-2-13 (Requirement for safety and essential performance of anaesthesia system).

